

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 2/6/2026 5:00:44 PM

## JOB DESCRIPTION

RED-HILL  
PFAS: Ka'amilo Wells P2

## JOB NUMBER

380-196020-1

# Eurofins Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Authorized for release by  
Maria Lopez, Project Manager  
[Maria.Lopez@et.eurofinsus.com](mailto:Maria.Lopez@et.eurofinsus.com)  
(626)386-1100

Generated  
2/6/2026 5:00:44 PM



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	11
Surrogate Summary . . . . .	12
Isotope Dilution Summary . . . . .	13
QC Sample Results . . . . .	14
QC Association Summary . . . . .	25
Lab Chronicle . . . . .	26
Certification Summary . . . . .	27
Method Summary . . . . .	28
Sample Summary . . . . .	29
Chain of Custody . . . . .	30
Receipt Checklists . . . . .	31

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Qualifiers

### LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-196020-1

**Job ID: 380-196020-1**

**Eurofins Pomona**

## Job Narrative 380-196020-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

### Receipt

The samples were received on 2/4/2026 9:22 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 3 coolers at receipt time were 1.7°C, 2.1°C and 2.5°C.

### PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Ka'amiilo Wells P2 (380-196020-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Client Sample ID: Ka'amilo Wells P2

## Lab Sample ID: 380-196020-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	3.0		2.0	ng/L	1		533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.5		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	3.7		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.5		2.0	ng/L	1		533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	3.9		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.1		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanoic acid (PFHxA)	4.5		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	4.6		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.9		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorobutanesulfonic acid (PFBS)	3.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluoroheptanoic acid (PFHpA)	2.3		2.0	ng/L	1		EPA 537.1 V2	Total/NA

## Client Sample ID: FB: Ka'amilo Wells P2

## Lab Sample ID: 380-196020-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-1**

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluorobutanesulfonic acid (PFBS)</b>	<b>3.0</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.5</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluorohexanoic acid (PFHxA)</b>	<b>3.7</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>4.5</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluorooctanoic acid (PFOA)</b>	<b>3.7</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Perfluoropentanoic acid (PFPeA)</b>	<b>3.9</b>		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:37	1
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
13C3 HFPO-DA	103		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C6 PFDA	102		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C5 PFHxA	107		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C4 PFHpA	112		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C8 PFOA	109		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C9 PFNA	108		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C7 PFUnA	105		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C2 PFDoA	104		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C4 PFBA	109		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C5 PFPeA	109		50 - 200			02/05/26 06:46	02/05/26 18:37	1
13C3 PFBS	117		50 - 200			02/05/26 06:46	02/05/26 18:37	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-1**

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	115		50 - 200	02/05/26 06:46	02/05/26 18:37	1
13C8 PFOS	112		50 - 200	02/05/26 06:46	02/05/26 18:37	1
13C2-4:2-FTS	126		50 - 200	02/05/26 06:46	02/05/26 18:37	1
13C2-6:2-FTS	118		50 - 200	02/05/26 06:46	02/05/26 18:37	1
13C2-8:2-FTS	111		50 - 200	02/05/26 06:46	02/05/26 18:37	1

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>5.1</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
N-methylperfluorooctanesulfonamide cetic acid (NMeFOSAA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
N-ethylperfluorooctanesulfonamide cetic acid (NEtFOSAA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluorohexanoic acid (PFHxA)</b>	<b>4.5</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluorooctanoic acid (PFOA)</b>	<b>4.6</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluorohexanesulfonic acid (PFHxS)</b>	<b>3.9</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluorobutanesulfonic acid (PFBS)</b>	<b>3.4</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
<b>Perfluoroheptanoic acid (PFHpA)</b>	<b>2.3</b>		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:32	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	114		70 - 130	02/05/26 01:18	02/05/26 16:32	1
13C2 PFHxA	117		70 - 130	02/05/26 01:18	02/05/26 16:32	1
13C2 PFDA	123		70 - 130	02/05/26 01:18	02/05/26 16:32	1
13C3-GenX	121		70 - 130	02/05/26 01:18	02/05/26 16:32	1

**Client Sample ID: FB: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-2**

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: FB: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-2**

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

**Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		02/05/26 06:46	02/05/26 21:36	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	100		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C6 PFDA	102		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C5 PFHxA	105		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C4 PFHpA	109		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C8 PFOA	107		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C9 PFNA	105		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C7 PFUnA	101		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C2 PFDoA	101		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C4 PFBA	110		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C5 PFPeA	110		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C3 PFBS	112		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C3 PFHxS	113		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C8 PFOS	111		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C2-4:2-FTS	111		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C2-6:2-FTS	115		50 - 200	02/05/26 06:46	02/05/26 21:36	1
13C2-8:2-FTS	105		50 - 200	02/05/26 06:46	02/05/26 21:36	1

Eurofins Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: FB: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-2**

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/05/26 01:18	02/05/26 16:41	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>			<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>
d5-NEtFOSAA	112		70 - 130			02/05/26 01:18	02/05/26 16:41	1
13C2 PFHxA	121		70 - 130			02/05/26 01:18	02/05/26 16:41	1
13C2 PFDA	117		70 - 130			02/05/26 01:18	02/05/26 16:41	1
13C3-GenX	116		70 - 130			02/05/26 01:18	02/05/26 16:41	1

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

**Client Sample ID: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.5		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>4.5</b>		ng/L	<b>4</b>	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
<b>Perfluorooctanesulfonic acid (PFOS)</b>	<b>5.1</b>		ng/L	<b>4</b>	2.0	EPA 537.1 V2	Total/NA
<b>Perfluorooctanoic acid (PFOA)</b>	<b>4.6</b>		ng/L	<b>4</b>	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.9		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

**Client Sample ID: FB: Ka'amilo Wells P2**

**Lab Sample ID: 380-196020-2**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-196020-1  
 SDG: PFAS: Ka'amilo Wells P2

**Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020**

**Matrix: Water**

**Prep Type: Total/NA**

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-196006-A-1-B MS	Matrix Spike	110	116	120	114
380-196006-A-1-C MSD	Matrix Spike Duplicate	107	118	119	118
380-196020-1	Ka'amilo Wells P2	114	117	123	121
380-196020-2	FB: Ka'amilo Wells P2	112	121	117	116
LCS 380-203609/21-A	Lab Control Sample	111	119	118	116
MBL 380-203609/19-A	Method Blank	112	120	118	110
MRL 380-203609/20-A	Lab Control Sample	115	119	123	115

**Surrogate Legend**

d5NEFOS = d5-NEtFOSAA  
 PFHxA = 13C2 PFHxA  
 PFDA = 13C2 PFDA  
 GenX = 13C3-GenX



# Isotope Dilution Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-196020-1  
 SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDoA (50-200)
380-196020-1	Ka'amilo Wells P2	103	102	107	112	109	108	105	104
380-196020-1 MS	Ka'amilo Wells P2	103	103	104	105	103	105	103	104
380-196020-1 MSD	Ka'amilo Wells P2	106	105	107	115	107	109	106	104
380-196020-2	FB: Ka'amilo Wells P2	100	102	105	109	107	105	101	101
LCS 380-203617/22-A	Lab Control Sample	106	108	108	112	110	110	107	106
MBL 380-203617/20-A	Method Blank	104	109	109	113	114	110	107	106
MRL 380-203617/21-A	Lab Control Sample	103	107	107	110	108	108	106	101

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-196020-1	Ka'amilo Wells P2	109	109	117	115	112	126	118	111
380-196020-1 MS	Ka'amilo Wells P2	105	106	115	113	113	115	115	108
380-196020-1 MSD	Ka'amilo Wells P2	111	111	112	113	112	122	116	107
380-196020-2	FB: Ka'amilo Wells P2	110	110	112	113	111	111	115	105
LCS 380-203617/22-A	Lab Control Sample	109	111	115	112	114	113	116	109
MBL 380-203617/20-A	Method Blank	111	109	116	116	112	121	116	112
MRL 380-203617/21-A	Lab Control Sample	112	110	112	116	111	119	120	109

### Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

**Lab Sample ID: MBL 380-203617/20-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		02/05/26 06:46	02/05/26 18:08	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	104		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C6 PFDA	109		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C5 PFHxA	109		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C4 PFHpA	113		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C8 PFOA	114		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C9 PFNA	110		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C7 PFUnA	107		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C2 PFDoA	106		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C4 PFBA	111		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C5 PFPeA	109		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C3 PFBS	116		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C3 PFHxS	116		50 - 200	02/05/26 06:46	02/05/26 18:08	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MBL 380-203617/20-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	112		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C2-4:2-FTS	121		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C2-6:2-FTS	116		50 - 200	02/05/26 06:46	02/05/26 18:08	1
13C2-8:2-FTS	112		50 - 200	02/05/26 06:46	02/05/26 18:08	1

**Lab Sample ID: LCS 380-203617/22-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.1	54.3		ng/L		90	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.1	57.1		ng/L		95	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.1	59.5		ng/L		99	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.1	56.3		ng/L		94	70 - 130
Perfluorodecanoic acid (PFDA)	60.1	57.2		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	60.1	57.7		ng/L		96	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.1	55.6		ng/L		93	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.1	56.9		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	60.1	56.1		ng/L		93	70 - 130
Perfluorononanoic acid (PFNA)	60.1	56.6		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.1	56.2		ng/L		93	70 - 130
Perfluorooctanoic acid (PFOA)	60.1	56.9		ng/L		95	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.1	57.9		ng/L		96	70 - 130
Perfluorobutanoic acid (PFBA)	60.1	57.1		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.1	58.3		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.1	61.0		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.1	58.6		ng/L		97	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.1	57.4		ng/L		96	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	60.1	55.7		ng/L		93	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.1	57.2		ng/L		95	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.1	57.3		ng/L		95	70 - 130
Perfluoropentanoic acid (PFPeA)	60.1	57.1		ng/L		95	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.1	57.5		ng/L		96	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: LCS 380-203617/22-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.1	57.5		ng/L		96	70 - 130
<b>LCS LCS</b>							
<b>Isotope Dilution</b>	<b>%Recovery</b>	<b>Qualifier</b>	<b>Limits</b>				
13C3 HFPO-DA	106		50 - 200				
13C6 PFDA	108		50 - 200				
13C5 PFHxA	108		50 - 200				
13C4 PFHpA	112		50 - 200				
13C8 PFOA	110		50 - 200				
13C9 PFNA	110		50 - 200				
13C7 PFUnA	107		50 - 200				
13C2 PFDoA	106		50 - 200				
13C4 PFBA	109		50 - 200				
13C5 PFPeA	111		50 - 200				
13C3 PFBS	115		50 - 200				
13C3 PFHxS	112		50 - 200				
13C8 PFOS	114		50 - 200				
13C2-4:2-FTS	113		50 - 200				
13C2-6:2-FTS	116		50 - 200				
13C2-8:2-FTS	109		50 - 200				

**Lab Sample ID: MRL 380-203617/21-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.12	J	ng/L		106	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.07	J	ng/L		103	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.95	J	ng/L		97	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.90	J	ng/L		95	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.97	J	ng/L		98	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.14	J	ng/L		107	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.90	J	ng/L		95	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.09	J	ng/L		104	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.14	J	ng/L		107	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.04	J	ng/L		102	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	1.89	J	ng/L		94	50 - 150

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: MRL 380-203617/21-A**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.09	J	ng/L		104	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.15	J	ng/L		107	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.40	J	ng/L		120	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	1.95	J	ng/L		97	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	1.99	J	ng/L		99	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.06	J	ng/L		103	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	2.04	J	ng/L		102	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	103		50 - 200
13C6 PFDA	107		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	110		50 - 200
13C8 PFOA	108		50 - 200
13C9 PFNA	108		50 - 200
13C7 PFUnA	106		50 - 200
13C2 PFDoA	101		50 - 200
13C4 PFBA	112		50 - 200
13C5 PFPeA	110		50 - 200
13C3 PFBS	112		50 - 200
13C3 PFHxS	116		50 - 200
13C8 PFOS	111		50 - 200
13C2-4:2-FTS	119		50 - 200
13C2-6:2-FTS	120		50 - 200
13C2-8:2-FTS	109		50 - 200

**Lab Sample ID: 380-196020-1 MS**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Ka'amilo Wells P2**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.4	58.3		ng/L		97	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.4	58.0		ng/L		96	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.4	58.9		ng/L		98	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-196020-1 MS**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Ka'amilo Wells P2**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.4	59.0		ng/L		98	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	3.0		60.4	61.3		ng/L		97	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.4	59.9		ng/L		99	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.4	59.6		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.4	61.4		ng/L		99	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.5		60.4	62.1		ng/L		97	70 - 130
Perfluorohexanoic acid (PFHxA)	3.7		60.4	60.9		ng/L		95	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.4	58.3		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	4.5		60.4	62.8		ng/L		97	70 - 130
Perfluorooctanoic acid (PFOA)	3.7		60.4	65.5		ng/L		102	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.4	58.5		ng/L		97	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.4	59.3		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.4	61.1		ng/L		101	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.4	62.6		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.4	63.0		ng/L		104	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.4	59.3		ng/L		98	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.4	57.6		ng/L		95	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.4	60.3		ng/L		100	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.4	60.3		ng/L		100	70 - 130
Perfluoropentanoic acid (PFPeA)	3.9		60.4	62.5		ng/L		97	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.4	61.9		ng/L		102	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.4	58.7		ng/L		96	70 - 130

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>
13C3 HFPO-DA	103		50 - 200
13C6 PFDA	103		50 - 200
13C5 PFHxA	104		50 - 200
13C4 PFHpA	105		50 - 200
13C8 PFOA	103		50 - 200
13C9 PFNA	105		50 - 200
13C7 PFUnA	103		50 - 200
13C2 PFDoA	104		50 - 200
13C4 PFBA	105		50 - 200
13C5 PFPeA	106		50 - 200
13C3 PFBS	115		50 - 200
13C3 PFHxS	113		50 - 200
13C8 PFOS	113		50 - 200

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

**Lab Sample ID: 380-196020-1 MS**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Ka'amilo Wells P2**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	115		50 - 200
13C2-6:2-FTS	115		50 - 200
13C2-8:2-FTS	108		50 - 200

**Lab Sample ID: 380-196020-1 MSD**  
**Matrix: Water**  
**Analysis Batch: 203798**

**Client Sample ID: Ka'amilo Wells P2**  
**Prep Type: Total/NA**  
**Prep Batch: 203617**

<b>Analyte</b>	<b>Sample Result</b>	<b>Sample Qualifier</b>	<b>Spike Added</b>	<b>MSD Result</b>	<b>MSD Qualifier</b>	<b>Unit</b>	<b>D</b>	<b>%Rec</b>	<b>%Rec Limits</b>	<b>RPD</b>	<b>RPD Limit</b>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.5	55.9		ng/L		92	70 - 130	4	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.5	55.7		ng/L		92	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.5	54.8		ng/L		91	70 - 130	7	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.5	57.9		ng/L		96	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	3.0		60.5	58.5		ng/L		92	70 - 130	5	30
Perfluorodecanoic acid (PFDA)	<2.0		60.5	57.8		ng/L		96	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		60.5	59.8		ng/L		99	70 - 130	0	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.5	55.6		ng/L		89	70 - 130	10	30
Perfluorohexanesulfonic acid (PFHxS)	3.5		60.5	58.8		ng/L		92	70 - 130	5	30
Perfluorohexanoic acid (PFHxA)	3.7		60.5	60.6		ng/L		94	70 - 130	0	30
Perfluorononanoic acid (PFNA)	<2.0		60.5	56.7		ng/L		94	70 - 130	3	30
Perfluorooctanesulfonic acid (PFOS)	4.5		60.5	60.9		ng/L		93	70 - 130	3	30
Perfluorooctanoic acid (PFOA)	3.7		60.5	59.5		ng/L		92	70 - 130	10	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.5	59.4		ng/L		98	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		60.5	57.2		ng/L		92	70 - 130	4	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.5	58.3		ng/L		96	70 - 130	5	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.5	57.0		ng/L		94	70 - 130	9	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.5	59.2		ng/L		98	70 - 130	6	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.5	54.2		ng/L		90	70 - 130	9	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		60.5	58.0		ng/L		96	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.5	57.2		ng/L		94	70 - 130	5	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.5	59.7		ng/L		99	70 - 130	1	30
Perfluoropentanoic acid (PFPeA)	3.9		60.5	63.1		ng/L		98	70 - 130	1	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.5	57.8		ng/L		96	70 - 130	7	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.5	57.3		ng/L		94	70 - 130	2	30

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	106		50 - 200
13C6 PFDA	105		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	115		50 - 200
13C8 PFOA	107		50 - 200
13C9 PFNA	109		50 - 200
13C7 PFUnA	106		50 - 200
13C2 PFDoA	104		50 - 200
13C4 PFBA	111		50 - 200
13C5 PFPeA	111		50 - 200
13C3 PFBS	112		50 - 200
13C3 PFHxS	113		50 - 200
13C8 PFOS	112		50 - 200
13C2-4:2-FTS	122		50 - 200
13C2-6:2-FTS	116		50 - 200
13C2-8:2-FTS	107		50 - 200

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

**Lab Sample ID: MBL 380-203609/19-A**  
**Matrix: Water**  
**Analysis Batch: 203712**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 203609**

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		02/05/26 01:18	02/05/26 15:35	1
Surrogate	MBL MBL		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
d5-NEtFOSAA	112		70 - 130			02/05/26 01:18	02/05/26 15:35	1
13C2 PFHxA	120		70 - 130			02/05/26 01:18	02/05/26 15:35	1
13C2 PFDA	118		70 - 130			02/05/26 01:18	02/05/26 15:35	1

Eurofins Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MBL 380-203609/19-A**  
**Matrix: Water**  
**Analysis Batch: 203712**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 203609**

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	110	Qualifier	70 - 130	02/05/26 01:18	02/05/26 15:35	1

**Lab Sample ID: LCS 380-203609/21-A**  
**Matrix: Water**  
**Analysis Batch: 203712**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203609**

<i>Analyte</i>	<i>Spike</i>	<i>LCS</i>	<i>LCS</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i>	<i>Limits</i>
	<i>Added</i>	<i>Result</i>	<i>Qualifier</i>					
Hexafluoropropylene Oxide	25.0	26.4		ng/L		106		70 - 130
Dimer Acid (HFPO-DA/GenX)								
Perfluorooctanesulfonic acid (PFOS)	25.0	26.2		ng/L		105		70 - 130
Perfluoroundecanoic acid (PFUnA)	25.0	27.9		ng/L		112		70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.0	26.5		ng/L		106		70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.0	25.8		ng/L		103		70 - 130
Perfluorohexanoic acid (PFHxA)	25.0	28.6		ng/L		114		70 - 130
Perfluorododecanoic acid (PFDoA)	25.0	27.1		ng/L		109		70 - 130
Perfluorooctanoic acid (PFOA)	25.0	26.9		ng/L		107		70 - 130
Perfluorodecanoic acid (PFDA)	25.0	27.2		ng/L		109		70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.0	26.6		ng/L		106		70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.0	26.5		ng/L		106		70 - 130
Perfluoroheptanoic acid (PFHpA)	25.0	28.0		ng/L		112		70 - 130
Perfluorononanoic acid (PFNA)	25.0	27.0		ng/L		108		70 - 130
Perfluorotetradecanoic acid (PFTA)	25.0	26.5		ng/L		106		70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.0	28.2		ng/L		113		70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	25.0	25.7		ng/L		103		70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.0	25.0		ng/L		100		70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.0	27.5		ng/L		110		70 - 130

  

<i>Surrogate</i>	<i>LCS</i>	<i>LCS</i>	<i>Limits</i>
	<i>%Recovery</i>	<i>Qualifier</i>	
d5-NEtFOSAA	111		70 - 130
13C2 PFHxA	119		70 - 130
13C2 PFDA	118		70 - 130
13C3-GenX	116		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

**Lab Sample ID: MRL 380-203609/20-A**  
**Matrix: Water**  
**Analysis Batch: 203712**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 203609**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.13	J	ng/L		107	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.21	J	ng/L		111	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	2.01	J	ng/L		101	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.14	J	ng/L		107	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.21	J	ng/L		111	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.14	J	ng/L		107	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.23	J	ng/L		112	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.28	J	ng/L		114	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.27	J	ng/L		114	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	2.15	J	ng/L		108	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.28	J	ng/L		114	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.19	J	ng/L		110	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.09	J	ng/L		104	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.03	J	ng/L		101	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	115		70 - 130
13C2 PFHxA	119		70 - 130
13C2 PFDA	123		70 - 130
13C3-GenX	115		70 - 130

**Lab Sample ID: 380-196006-A-1-B MS**  
**Matrix: Water**  
**Analysis Batch: 203712**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 203609**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	26.1		ng/L		104	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	27.4		ng/L		109	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	28.5		ng/L		114	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	26.7		ng/L		106	70 - 130

Eurofins Pomona





# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## LCMS

### Prep Batch: 203609

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-196020-1	Ka'amilo Wells P2	Total/NA	Water	537.1 DW	
380-196020-2	FB: Ka'amilo Wells P2	Total/NA	Water	537.1 DW	
MBL 380-203609/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-203609/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-203609/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-196006-A-1-B MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-196006-A-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

### Prep Batch: 203617

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-196020-1	Ka'amilo Wells P2	Total/NA	Water	533	
380-196020-2	FB: Ka'amilo Wells P2	Total/NA	Water	533	
MBL 380-203617/20-A	Method Blank	Total/NA	Water	533	
LCS 380-203617/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-203617/21-A	Lab Control Sample	Total/NA	Water	533	
380-196020-1 MS	Ka'amilo Wells P2	Total/NA	Water	533	
380-196020-1 MSD	Ka'amilo Wells P2	Total/NA	Water	533	

### Analysis Batch: 203712

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-196020-1	Ka'amilo Wells P2	Total/NA	Water	EPA 537.1 V2	203609
380-196020-2	FB: Ka'amilo Wells P2	Total/NA	Water	EPA 537.1 V2	203609
MBL 380-203609/19-A	Method Blank	Total/NA	Water	EPA 537.1 V2	203609
LCS 380-203609/21-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	203609
MRL 380-203609/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	203609
380-196006-A-1-B MS	Matrix Spike	Total/NA	Water	EPA 537.1 V2	203609
380-196006-A-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	EPA 537.1 V2	203609

### Analysis Batch: 203798

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-196020-1	Ka'amilo Wells P2	Total/NA	Water	533	203617
380-196020-2	FB: Ka'amilo Wells P2	Total/NA	Water	533	203617
MBL 380-203617/20-A	Method Blank	Total/NA	Water	533	203617
LCS 380-203617/22-A	Lab Control Sample	Total/NA	Water	533	203617
MRL 380-203617/21-A	Lab Control Sample	Total/NA	Water	533	203617
380-196020-1 MS	Ka'amilo Wells P2	Total/NA	Water	533	203617
380-196020-1 MSD	Ka'amilo Wells P2	Total/NA	Water	533	203617

# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Client Sample ID: Ka'amilo Wells P2

## Lab Sample ID: 380-196020-1

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			203617	XTD8	EA POM	02/05/26 06:46
Total/NA	Analysis	533		1	203798	SZ9R	EA POM	02/05/26 18:37
Total/NA	Prep	537.1 DW			203609	G9MN	EA POM	02/05/26 01:18
Total/NA	Analysis	EPA 537.1 V2		1	203712	SZ9R	EA POM	02/05/26 16:32

## Client Sample ID: FB: Ka'amilo Wells P2

## Lab Sample ID: 380-196020-2

Date Collected: 02/02/26 12:31

Matrix: Water

Date Received: 02/04/26 09:22

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			203617	XTD8	EA POM	02/05/26 06:46
Total/NA	Analysis	533		1	203798	SZ9R	EA POM	02/05/26 21:36
Total/NA	Prep	537.1 DW			203609	G9MN	EA POM	02/05/26 01:18
Total/NA	Analysis	EPA 537.1 V2		1	203712	SZ9R	EA POM	02/05/26 16:41

### Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

## Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

\* Accreditation/Certification renewal pending - accreditation/certification considered valid.

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

**Protocol References:**

EPA = US Environmental Protection Agency

**Laboratory References:**

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-196020-1  
SDG: PFAS: Ka'amilo Wells P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-196020-1	Ka'amilo Wells P2	Water	02/02/26 12:31	02/04/26 09:22	Hawaii
380-196020-2	FB: Ka'amilo Wells P2	Water	02/02/26 12:31	02/04/26 09:22	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

**Monrovia, CA (Suite 100)**  
 750 Royal Oaks Drive Suite 100  
 Monrovia, CA 91016  
 Phone (626) 386-1100

# Chain of Custody Record



<b>Client Information</b>		Lab PM: Lopez, Maria	Carrier Tracking No(s):	COC No:
Client Contact: kirik Iwamoto		E-Mail: Maria.Lopez@et.euronisus.com	State of Origin:	Page:
Company: City & County of Honolulu		PWSID:		Job #:
Address: 630 South Beretania Street, Chemistry Lab				
City: Honolulu				
State, Zip: HI, 96843				
Phone: 808-748-5840 (tel)				
Email: kiwamoto@hbws.org				
Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill				
Site:				
Due Date Requested:		<b>Analysis Requested</b>		
TAT Requested (days): RUSH		633 - All Analytes		
Compliance Project: Δ No		637.1_DW_PREC - 637.1 Full List		
PO #: C20525101 exp 05312023		625.2_PREC - (MOD) 625plus PLUS TICs		
WO #: 38001111		80158_DRO_LL_G8 - HNL Ranges: C10-C24/C24-C36/C38-C18		
Project #: 38001111		80158_GRO_LL - (MOD) GRO		
SSOW#:		SUBCONTRACT - 625 PAH Physic LL (EAL) + TICs		
Sample Identification		Perform MS/MSD (Yes or No) <input checked="" type="checkbox"/>		
Ka'amilo Wells P2		Field Filtered Sample (Yes or No) <input checked="" type="checkbox"/>		
Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Preservation Code	Matrix (Preservative, Stabilizer, Other)
2-Feb-2026	1231	G	G	Water
FB: Ka'amilo Wells P2	2-Feb-2026	1231	11	Water
Special Instructions/Note: 380-196020 COC				
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological Deliverable Requested: I, II, III, IV, Other (specify)				
Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months				
Special Instructions/QC Requirements:				
Empty Kit Relinquished by		Time:		
Date/Time: 2/26/2026 1400		Date/Time: 2/26/2026 9:22		
Company: HBWS		Company: GAMP		
Date/Time: 2/26/2026		Date/Time: 2/26/2026		
Relinquished by		Received by		
Custody Seals Intact: Δ Yes Δ No		Cooler Temperature(s) °C and Other Remarks: 6°C/20°F = 2.2°F		



## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-196020-1  
SDG Number: PFAS: Ka'amilo Wells P2

**Login Number: 196020**  
**List Number: 1**  
**Creator: Ngo, Theodore**

**List Source: Eurofins Pomona**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

