

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 1/5/2026 5:11:49 PM

JOB DESCRIPTION

RED-HILL
PFAS: Ka'amilo Wells P2

JOB NUMBER

380-190141-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
1/5/2026 5:11:49 PM

Table of Contents

Cover Page 1

Table of Contents 3

Definitions/Glossary 4

Case Narrative 5

Detection Summary 6

Client Sample Results 7

Action Limit Summary 11

Surrogate Summary 12

Isotope Dilution Summary 13

QC Sample Results 14

QC Association Summary 25

Lab Chronicle 26

Certification Summary 27

Method Summary 28

Sample Summary 29

Chain of Custody 30

Receipt Checklists 31



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-190141-1

Job ID: 380-190141-1

Eurofins Pomona

Job Narrative 380-190141-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 12/31/2025 9:57 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.3°C.

PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Ka'amilo Wells P2 (380-190141-1) (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2

Lab Sample ID: 380-190141-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanesulfonic acid (PFBS)	3.3		2.0	ng/L	1		533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0	ng/L	1		533	Total/NA
Perfluorohexanoic acid (PFHxA)	3.8		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.4		2.0	ng/L	1		533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		2.0	ng/L	1		533	Total/NA
Perfluoropentanoic acid (PFPeA)	4.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.2		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanoic acid (PFHxA)	3.8		2.0	ng/L	1		537.1	Total/NA
Perfluorooctanoic acid (PFOA)	4.5		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.9		2.0	ng/L	1		537.1	Total/NA
Perfluorobutanesulfonic acid (PFBS)	3.6		2.0	ng/L	1		537.1	Total/NA
Perfluoroheptanoic acid (PFHpA)	2.1		2.0	ng/L	1		537.1	Total/NA

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2

Lab Sample ID: 380-190141-1

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorobutanesulfonic acid (PFBS)	3.3		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorohexanoic acid (PFHxA)	3.8		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorooctanesulfonic acid (PFOS)	4.4		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorooctanoic acid (PFOA)	3.7		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoropentanoic acid (PFPeA)	4.2		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:35	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	88		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C6 PFDA	99		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C5 PFHxA	96		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C4 PFHpA	98		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C8 PFOA	103		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C9 PFNA	103		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C7 PFUnA	102		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C2 PFDoA	95		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C4 PFBA	98		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C5 PFPeA	99		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C3 PFBS	95		50 - 200	01/02/26 17:53	01/03/26 16:35	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2

Lab Sample ID: 380-190141-1

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	97		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C8 PFOS	102		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C2-4:2-FTS	119		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C2-6:2-FTS	108		50 - 200	01/02/26 17:53	01/03/26 16:35	1
13C2-8:2-FTS	108		50 - 200	01/02/26 17:53	01/03/26 16:35	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorooctanesulfonic acid (PFOS)	5.2		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorohexanoic acid (PFHxA)	3.8		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorooctanoic acid (PFOA)	4.5		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorohexanesulfonic acid (PFHxS)	3.9		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorobutanesulfonic acid (PFBS)	3.6		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluoroheptanoic acid (PFHpA)	2.1		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:14	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	113		70 - 130	01/02/26 02:20	01/02/26 18:14	1
13C2 PFHxA	124		70 - 130	01/02/26 02:20	01/02/26 18:14	1
13C2 PFDA	117		70 - 130	01/02/26 02:20	01/02/26 18:14	1
13C3-GenX	120		70 - 130	01/02/26 02:20	01/02/26 18:14	1

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/02/26 17:53	01/03/26 16:45	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	86		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C6 PFDA	102		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C5 PFHxA	100		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C4 PFHpA	99		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C8 PFOA	104		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C9 PFNA	103		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C7 PFUnA	103		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C2 PFDoA	101		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C4 PFBA	100		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C5 PFPeA	106		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C3 PFBS	93		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C3 PFHxS	100		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C8 PFOS	102		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C2-4:2-FTS	113		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C2-6:2-FTS	114		50 - 200	01/02/26 17:53	01/03/26 16:45	1
13C2-8:2-FTS	109		50 - 200	01/02/26 17:53	01/03/26 16:45	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 02:20	01/02/26 18:24	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	104		70 - 130			01/02/26 02:20	01/02/26 18:24	1
13C2 PFHxA	116		70 - 130			01/02/26 02:20	01/02/26 18:24	1
13C2 PFDA	113		70 - 130			01/02/26 02:20	01/02/26 18:24	1
13C3-GenX	105		70 - 130			01/02/26 02:20	01/02/26 18:24	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2

Lab Sample ID: 380-190141-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	4.4		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	3.7		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	5.2		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	4.5		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.9		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-190078-B-1-A MS	Matrix Spike	106	118	109	109
380-190078-C-1-A MSD	Matrix Spike Duplicate	113	114	108	109
380-190141-1	Ka'amilo Wells P2	113	124	117	120
380-190141-2	FB: Ka'amilo Wells P2	104	116	113	105
LCS 380-195448/22-A	Lab Control Sample	102	105	107	95
MBL 380-195448/20-A	Method Blank	106	110	111	104
MRL 380-195448/21-A	Lab Control Sample	110	119	109	109

Surrogate Legend

d5NEFOS = d5-NEtFOSAA

PFHxA = 13C2 PFHxA

PFDA = 13C2 PFDA

GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-189915-B-1-A MS	Matrix Spike	95	104	106	103	106	107	105	104
380-189915-B-1-B MSD	Matrix Spike Duplicate	91	101	99	96	103	100	98	100
380-190141-1	Ka'amilo Wells P2	88	99	96	98	103	103	102	95
380-190141-2	FB: Ka'amilo Wells P2	86	102	100	99	104	103	103	101
LCS 380-195601/22-A	Lab Control Sample	94	107	107	102	103	105	104	105
MBL 380-195601/20-A	Method Blank	89	103	104	101	105	106	100	102
MRL 380-195601/21-A	Lab Control Sample	89	100	106	101	100	106	101	103

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-189915-B-1-A MS	Matrix Spike	101	102	100	103	103	111	109	107
380-189915-B-1-B MSD	Matrix Spike Duplicate	103	103	99	101	102	112	114	105
380-190141-1	Ka'amilo Wells P2	98	99	95	97	102	119	108	108
380-190141-2	FB: Ka'amilo Wells P2	100	106	93	100	102	113	114	109
LCS 380-195601/22-A	Lab Control Sample	99	99	103	100	102	109	106	105
MBL 380-195601/20-A	Method Blank	103	105	95	107	105	120	114	107
MRL 380-195601/21-A	Lab Control Sample	104	103	100	101	104	112	112	108

Surrogate Legend

HFPODA = 13C3 HFPO-DA
C6PFDA = 13C6 PFDA
13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDaA = 13C2 PFDaA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-195601/20-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 195601

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		01/02/26 17:53	01/03/26 13:24	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	89		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C6 PFDA	103		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C5 PFHxA	104		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C4 PFHpA	101		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C8 PFOA	105		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C9 PFNA	106		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C7 PFUnA	100		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C2 PFDoA	102		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C4 PFBA	103		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C5 PFPeA	105		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C3 PFBS	95		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C3 PFHxS	107		50 - 200	01/02/26 17:53	01/03/26 13:24	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-195601/20-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 195601

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	105		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C2-4:2-FTS	120		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C2-6:2-FTS	114		50 - 200	01/02/26 17:53	01/03/26 13:24	1
13C2-8:2-FTS	107		50 - 200	01/02/26 17:53	01/03/26 13:24	1

Lab Sample ID: LCS 380-195601/22-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	101		ng/L		84	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	120	107		ng/L		89	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	112		ng/L		93	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	114		ng/L		95	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	111		ng/L		92	70 - 130
Perfluorodecanoic acid (PFDA)	120	109		ng/L		90	70 - 130
Perfluorododecanoic acid (PFDoA)	120	110		ng/L		91	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	106		ng/L		88	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	108		ng/L		90	70 - 130
Perfluorohexanoic acid (PFHxA)	120	111		ng/L		92	70 - 130
Perfluorononanoic acid (PFNA)	120	110		ng/L		91	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	110		ng/L		92	70 - 130
Perfluorooctanoic acid (PFOA)	120	114		ng/L		94	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	112		ng/L		93	70 - 130
Perfluorobutanoic acid (PFBA)	120	114		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	113		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	110		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	117		ng/L		97	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	114		ng/L		95	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	120	115		ng/L		96	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	116		ng/L		96	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	97.1		ng/L		81	70 - 130
Perfluoropentanoic acid (PFPeA)	120	119		ng/L		99	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	115		ng/L		95	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-195601/22-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	110		ng/L		91	70 - 130
		LCS	LCS				
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	94		50 - 200				
13C6 PFDA	107		50 - 200				
13C5 PFHxA	107		50 - 200				
13C4 PFHpA	102		50 - 200				
13C8 PFOA	103		50 - 200				
13C9 PFNA	105		50 - 200				
13C7 PFUnA	104		50 - 200				
13C2 PFDoA	105		50 - 200				
13C4 PFBA	99		50 - 200				
13C5 PFPeA	99		50 - 200				
13C3 PFBS	103		50 - 200				
13C3 PFHxS	100		50 - 200				
13C8 PFOS	102		50 - 200				
13C2-4:2-FTS	109		50 - 200				
13C2-6:2-FTS	106		50 - 200				
13C2-8:2-FTS	105		50 - 200				

Lab Sample ID: MRL 380-195601/21-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.87	J	ng/L		93	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.01	2.01	J	ng/L		100	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.87	J	ng/L		93	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	2.05	J	ng/L		102	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.09	J	ng/L		104	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.06	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.02	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	1.95	J	ng/L		97	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.92	J	ng/L		95	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.97	J	ng/L		98	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.95	J	ng/L		97	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.08	J	ng/L		104	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.05	J	ng/L		102	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.06	J	ng/L		103	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-195601/21-A

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	2.10	J	ng/L		104	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.17	J	ng/L		108	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.08	J	ng/L		104	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	2.07	J	ng/L		103	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	2.01	2.28	J	ng/L		113	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.90	J	ng/L		95	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.61	J	ng/L		80	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	2.14	J	ng/L		106	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.96	J	ng/L		98	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.93	J	ng/L		96	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	89		50 - 200
13C6 PFDA	100		50 - 200
13C5 PFHxA	106		50 - 200
13C4 PFHpA	101		50 - 200
13C8 PFOA	100		50 - 200
13C9 PFNA	106		50 - 200
13C7 PFUnA	101		50 - 200
13C2 PFDoA	103		50 - 200
13C4 PFBA	104		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	100		50 - 200
13C3 PFHxS	101		50 - 200
13C8 PFOS	104		50 - 200
13C2-4:2-FTS	112		50 - 200
13C2-6:2-FTS	112		50 - 200
13C2-8:2-FTS	108		50 - 200

Lab Sample ID: 380-189915-B-1-A MS

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	56.0		ng/L		93	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		60.2	56.8		ng/L		94	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	60.0		ng/L		100	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-189915-B-1-A MS

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		60.2	57.7		ng/L		96	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.2	61.9		ng/L		102	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		60.2	59.1		ng/L		98	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		60.2	58.9		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	56.8		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.2	56.9		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		60.2	57.3		ng/L		94	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		60.2	56.8		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.2	57.7		ng/L		93	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		60.2	59.0		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	57.9		ng/L		96	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		60.2	59.2		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	57.6		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	61.4		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	61.5		ng/L		102	70 - 130
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		60.2	53.7		ng/L		89	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		60.2	67.1		ng/L		111	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	59.9		ng/L		99	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	48.1		ng/L		80	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		60.2	59.3		ng/L		97	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	60.6		ng/L		101	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	58.4		ng/L		97	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	95		50 - 200
13C6 PFDA	104		50 - 200
13C5 PFHxA	106		50 - 200
13C4 PFHpA	103		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	107		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	104		50 - 200
13C4 PFBA	101		50 - 200
13C5 PFPeA	102		50 - 200
13C3 PFBS	100		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	103		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-189915-B-1-A MS

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195601

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	111		50 - 200
13C2-6:2-FTS	109		50 - 200
13C2-8:2-FTS	107		50 - 200

Lab Sample ID: 380-189915-B-1-B MSD

Matrix: Water

Analysis Batch: 195630

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 195601

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		60.2	55.0		ng/L		91	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		60.2	57.3		ng/L		95	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		60.2	58.0		ng/L		96	70 - 130	3	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		60.2	58.7		ng/L		97	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		60.2	60.2		ng/L		99	70 - 130	3	30
Perfluorodecanoic acid (PFDA)	<2.0		60.2	57.9		ng/L		96	70 - 130	2	30
Perfluorododecanoic acid (PFDaA)	<2.0		60.2	58.6		ng/L		97	70 - 130	1	30
Perfluoroheptanoic acid (PFHpA)	<2.0		60.2	58.6		ng/L		97	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		60.2	59.3		ng/L		97	70 - 130	4	30
Perfluorohexanoic acid (PFHxA)	<2.0		60.2	58.0		ng/L		95	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		60.2	59.3		ng/L		98	70 - 130	4	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		60.2	57.9		ng/L		93	70 - 130	0	30
Perfluorooctanoic acid (PFOA)	<2.0		60.2	57.4		ng/L		95	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		60.2	59.5		ng/L		99	70 - 130	3	30
Perfluorobutanoic acid (PFBA)	<2.0		60.2	58.3		ng/L		97	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		60.2	60.4		ng/L		100	70 - 130	5	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		60.2	61.6		ng/L		102	70 - 130	0	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		60.2	58.9		ng/L		98	70 - 130	4	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		60.2	60.7		ng/L		101	70 - 130	12	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		60.2	68.2		ng/L		113	70 - 130	2	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		60.2	56.0		ng/L		93	70 - 130	7	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		60.2	47.8		ng/L		79	70 - 130	0	30
Perfluoropentanoic acid (PFPeA)	<2.0		60.2	55.8		ng/L		91	70 - 130	6	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		60.2	60.2		ng/L		100	70 - 130	1	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		60.2	58.0		ng/L		96	70 - 130	1	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD %Recovery	MSD Qualifier	Limits
13C3 HFPO-DA	91		50 - 200
13C6 PFDA	101		50 - 200
13C5 PFHxA	99		50 - 200
13C4 PFHpA	96		50 - 200
13C8 PFOA	103		50 - 200
13C9 PFNA	100		50 - 200
13C7 PFUnA	98		50 - 200
13C2 PFDoA	100		50 - 200
13C4 PFBA	103		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	99		50 - 200
13C3 PFHxS	101		50 - 200
13C8 PFOS	102		50 - 200
13C2-4:2-FTS	112		50 - 200
13C2-6:2-FTS	114		50 - 200
13C2-8:2-FTS	105		50 - 200

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-195448/20-A
Matrix: Water
Analysis Batch: 195551

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195448

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/02/26 02:20	01/02/26 16:04	1
Surrogate	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	106		70 - 130			01/02/26 02:20	01/02/26 16:04	1
13C2 PFHxA	110		70 - 130			01/02/26 02:20	01/02/26 16:04	1
13C2 PFDA	111		70 - 130			01/02/26 02:20	01/02/26 16:04	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MBL 380-195448/20-A
Matrix: Water
Analysis Batch: 195551

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195448

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	104		70 - 130	01/02/26 02:20	01/02/26 16:04	1

Lab Sample ID: LCS 380-195448/22-A
Matrix: Water
Analysis Batch: 195551

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195448

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	25.1	22.1		ng/L		88	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	25.1	27.0		ng/L		108	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	26.0		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	24.4		ng/L		97	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	23.9		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	24.7		ng/L		99	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	25.9		ng/L		103	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	24.8		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	24.4		ng/L		97	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	28.2		ng/L		113	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	28.6		ng/L		114	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	24.8		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	25.1	25.4		ng/L		102	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	22.5		ng/L		90	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	26.3		ng/L		105	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	25.1	26.4		ng/L		105	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	24.4		ng/L		97	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	22.9		ng/L		92	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	102		70 - 130
13C2 PFHxA	105		70 - 130
13C2 PFDA	107		70 - 130
13C3-GenX	95		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MRL 380-195448/21-A

Matrix: Water

Analysis Batch: 195551

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195448

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	2.04	J	ng/L		102	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	2.22	J	ng/L		110	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.25	J	ng/L		112	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	2.07	J	ng/L		103	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	2.11	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.16	J	ng/L		107	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.14	J	ng/L		107	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.05	J	ng/L		102	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.18	J	ng/L		109	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.19	J	ng/L		109	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.25	J	ng/L		112	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.18	J	ng/L		108	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	1.51	J	ng/L		75	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	2.11	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	2.06	J	ng/L		102	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.78	J	ng/L		88	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	2.11	J	ng/L		105	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	110		70 - 130
13C2 PFHxA	119		70 - 130
13C2 PFDA	109		70 - 130
13C3-GenX	109		70 - 130

Lab Sample ID: 380-190078-B-1-A MS

Matrix: Water

Analysis Batch: 195551

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195448

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	24.6		ng/L		98	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	26.6		ng/L		106	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	26.0		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	24.5		ng/L		97	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 380-190078-B-1-A MS

Matrix: Water

Analysis Batch: 195551

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195448

Table 1: PFAS Data - 10/20/2023											
Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec		
	Result	Qualifier	Added	Result	Qualifier			Limits	Limits		
N-ethylperfluorooctanesulfonami doacetic acid (NEtFOSAA)	<2.0		25.1	25.6		ng/L		102	70 - 130		
Perfluorohexanoic acid (PFHxA)	<2.0		25.1	27.0		ng/L		108	70 - 130		
Perfluorododecanoic acid (PFDoA)	<2.0		25.1	26.9		ng/L		107	70 - 130		
Perfluorooctanoic acid (PFOA)	<2.0		25.1	25.8		ng/L		103	70 - 130		
Perfluorodecanoic acid (PFDA)	<2.0		25.1	25.2		ng/L		101	70 - 130		
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.1	26.4		ng/L		105	70 - 130		
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.1	27.4		ng/L		109	70 - 130		
Perfluoroheptanoic acid (PFHpA)	<2.0		25.1	26.4		ng/L		105	70 - 130		
Perfluorononanoic acid (PFNA)	<2.0		25.1	26.6		ng/L		106	70 - 130		
Perfluorotetradecanoic acid (PFTA)	<2.0		25.1	22.2		ng/L		88	70 - 130		
Perfluorotridecanoic acid (PFTrDA)	<2.0		25.1	26.9		ng/L		107	70 - 130		
9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.1	25.7		ng/L		102	70 - 130		
11-Chloroeicosafluoro-3-oxaund ecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.1	23.3		ng/L		93	70 - 130		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.1	26.0		ng/L		104	70 - 130		

Lab Sample ID: 380-190078-C-1-A MSD

Matrix: Water

Analysis Batch: 195551

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 195448

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	24.7		ng/L		98	70 - 130	0	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	27.3		ng/L		109	70 - 130	3	30
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	26.9		ng/L		107	70 - 130	4	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	26.3		ng/L		105	70 - 130	7	30
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		25.1	27.9		ng/L		111	70 - 130	9	30
Perfluorohexanoic acid (PFHxA)	<2.0		25.1	26.9		ng/L		107	70 - 130	0	30
Perfluorododecanoic acid (PFDoA)	<2.0		25.1	27.2		ng/L		108	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		25.1	27.1		ng/L		108	70 - 130	5	30
Perfluorodecanoic acid (PFDA)	<2.0		25.1	25.9		ng/L		103	70 - 130	3	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 380-190078-C-1-A MSD

Matrix: Water

Analysis Batch: 195551

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 195448

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.1	28.4		ng/L		113	70 - 130	7	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.1	29.1		ng/L		116	70 - 130	6	30
Perfluoroheptanoic acid (PFHpA)	<2.0		25.1	27.2		ng/L		109	70 - 130	3	30
Perfluorononanoic acid (PFNA)	<2.0		25.1	27.2		ng/L		108	70 - 130	2	30
Perfluorotetradecanoic acid (PFTA)	<2.0		25.1	23.4		ng/L		93	70 - 130	5	30
Perfluorotridecanoic acid (PFTTrDA)	<2.0		25.1	27.7		ng/L		110	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.1	26.7		ng/L		106	70 - 130	4	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.1	23.9		ng/L		95	70 - 130	3	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.1	26.4		ng/L		105	70 - 130	1	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	113		70 - 130								
13C2 PFHxA	114		70 - 130								
13C2 PFDA	108		70 - 130								
13C3-GenX	109		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

LCMS

Prep Batch: 195448

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190141-1	Ka'amilo Wells P2	Total/NA	Water	537.1 DW	
380-190141-2	FB: Ka'amilo Wells P2	Total/NA	Water	537.1 DW	
MBL 380-195448/20-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-195448/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-195448/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-190078-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-190078-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 195551

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190141-1	Ka'amilo Wells P2	Total/NA	Water	537.1	195448
380-190141-2	FB: Ka'amilo Wells P2	Total/NA	Water	537.1	195448
MBL 380-195448/20-A	Method Blank	Total/NA	Water	537.1	195448
LCS 380-195448/22-A	Lab Control Sample	Total/NA	Water	537.1	195448
MRL 380-195448/21-A	Lab Control Sample	Total/NA	Water	537.1	195448
380-190078-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	195448
380-190078-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	195448

Prep Batch: 195601

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190141-1	Ka'amilo Wells P2	Total/NA	Water	533	
380-190141-2	FB: Ka'amilo Wells P2	Total/NA	Water	533	
MBL 380-195601/20-A	Method Blank	Total/NA	Water	533	
LCS 380-195601/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-195601/21-A	Lab Control Sample	Total/NA	Water	533	
380-189915-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-189915-B-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 195630

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190141-1	Ka'amilo Wells P2	Total/NA	Water	533	195601
380-190141-2	FB: Ka'amilo Wells P2	Total/NA	Water	533	195601
MBL 380-195601/20-A	Method Blank	Total/NA	Water	533	195601
LCS 380-195601/22-A	Lab Control Sample	Total/NA	Water	533	195601
MRL 380-195601/21-A	Lab Control Sample	Total/NA	Water	533	195601
380-189915-B-1-A MS	Matrix Spike	Total/NA	Water	533	195601
380-189915-B-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	533	195601

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Client Sample ID: Ka'amilo Wells P2

Lab Sample ID: 380-190141-1

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195601	N8NE	EA POM	01/02/26 17:53
Total/NA	Analysis	533		1	195630	SZ9R	EA POM	01/03/26 16:35
Total/NA	Prep	537.1 DW			195448	G9MN	EA POM	01/02/26 02:20
Total/NA	Analysis	537.1		1	195551	SZ9R	EA POM	01/02/26 18:14

Client Sample ID: FB: Ka'amilo Wells P2

Lab Sample ID: 380-190141-2

Date Collected: 12/29/25 12:38

Matrix: Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195601	N8NE	EA POM	01/02/26 17:53
Total/NA	Analysis	533		1	195630	SZ9R	EA POM	01/03/26 16:45
Total/NA	Prep	537.1 DW			195448	G9MN	EA POM	01/02/26 02:20
Total/NA	Analysis	537.1		1	195551	SZ9R	EA POM	01/02/26 18:24

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190141-1
SDG: PFAS: Ka'amilo Wells P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-190141-1	Ka'amilo Wells P2	Water	12/29/25 12:38	12/31/25 09:57	Hawaii
380-190141-2	FB: Ka'amilo Wells P2	Water	12/29/25 12:38	12/31/25 09:57	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

750 Royal Oaks Drive Suite 100
Monrovia, CA 91016
Phone (626) 386-1100

Client Information		Lab PM: Arada, Rachelle		Carrier Tracking No(s): 380-27941-2757.2		COC No: 380-27941-2757.2	
Client Contact: Kirk Iwamoto		E-Mail: Rachelle.Arada@et.euronisus.com		State of Origin:		Page 2 of 2	
City & County of Honolulu		PWSID:		Analysis Requested		Job #:	
Address: 6330 South Beretania Street, Chemistry Lab Honolulu State, Zip: HI, 96843		Due Date Requested: TAT Requested (days): RUSH		Total Number of Containers		Preservation Codes: M - Hexane N - None O - AsNaO2 P - Na2OAS Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecaldehyde U - Acetone V - MCAA W - pH 4-5 Y - Trisoma Z - other (specify) Other:	
Phone: 808-748-5840 (tel)		Compliance Project: ☐ No		533 - All Analytes		Special Instructions/Note:	
Email: Kiwamoto@hbws.org		PO #: C20525101 exp 05312023		537 1, DW, PREC - 537.1 Full Lot		538 2, PREC - (MOD) 526plus PLUS TICs	
Project #: 38001111		WFO #:		539 1, DW, PREC - 537.1 Full Lot		540 2, PREC - (MOD) 526plus PLUS TICs	
Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill		Sample Date		541 1, DW, PREC - 537.1 Full Lot		542 2, PREC - (MOD) 526plus PLUS TICs	
Site:		Sample Time		543 1, DW, PREC - 537.1 Full Lot		544 2, PREC - (MOD) 526plus PLUS TICs	
Sample Identification		Sample Type (C=Comp, G=grab)		545 1, DW, PREC - 537.1 Full Lot		546 2, PREC - (MOD) 526plus PLUS TICs	
Ka'amilo Wells P2		Matrix (Water, Specific, Organic, Inorganic)		547 1, DW, PREC - 537.1 Full Lot		548 2, PREC - (MOD) 526plus PLUS TICs	
FB: Ka'amilo Wells P2		Preservation Code:		549 1, DW, PREC - 537.1 Full Lot		550 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		551 1, DW, PREC - 537.1 Full Lot		552 2, PREC - (MOD) 526plus PLUS TICs	
		1238		553 1, DW, PREC - 537.1 Full Lot		554 2, PREC - (MOD) 526plus PLUS TICs	
		Water		555 1, DW, PREC - 537.1 Full Lot		556 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		557 1, DW, PREC - 537.1 Full Lot		558 2, PREC - (MOD) 526plus PLUS TICs	
		1238		559 1, DW, PREC - 537.1 Full Lot		560 2, PREC - (MOD) 526plus PLUS TICs	
		Water		561 1, DW, PREC - 537.1 Full Lot		562 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		563 1, DW, PREC - 537.1 Full Lot		564 2, PREC - (MOD) 526plus PLUS TICs	
		1238		565 1, DW, PREC - 537.1 Full Lot		566 2, PREC - (MOD) 526plus PLUS TICs	
		Water		567 1, DW, PREC - 537.1 Full Lot		568 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		569 1, DW, PREC - 537.1 Full Lot		570 2, PREC - (MOD) 526plus PLUS TICs	
		1238		571 1, DW, PREC - 537.1 Full Lot		572 2, PREC - (MOD) 526plus PLUS TICs	
		Water		573 1, DW, PREC - 537.1 Full Lot		574 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		575 1, DW, PREC - 537.1 Full Lot		576 2, PREC - (MOD) 526plus PLUS TICs	
		1238		577 1, DW, PREC - 537.1 Full Lot		578 2, PREC - (MOD) 526plus PLUS TICs	
		Water		579 1, DW, PREC - 537.1 Full Lot		580 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		581 1, DW, PREC - 537.1 Full Lot		582 2, PREC - (MOD) 526plus PLUS TICs	
		1238		583 1, DW, PREC - 537.1 Full Lot		584 2, PREC - (MOD) 526plus PLUS TICs	
		Water		585 1, DW, PREC - 537.1 Full Lot		586 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		587 1, DW, PREC - 537.1 Full Lot		588 2, PREC - (MOD) 526plus PLUS TICs	
		1238		589 1, DW, PREC - 537.1 Full Lot		590 2, PREC - (MOD) 526plus PLUS TICs	
		Water		591 1, DW, PREC - 537.1 Full Lot		592 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		593 1, DW, PREC - 537.1 Full Lot		594 2, PREC - (MOD) 526plus PLUS TICs	
		1238		595 1, DW, PREC - 537.1 Full Lot		596 2, PREC - (MOD) 526plus PLUS TICs	
		Water		597 1, DW, PREC - 537.1 Full Lot		598 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		599 1, DW, PREC - 537.1 Full Lot		600 2, PREC - (MOD) 526plus PLUS TICs	
		1238		601 1, DW, PREC - 537.1 Full Lot		602 2, PREC - (MOD) 526plus PLUS TICs	
		Water		603 1, DW, PREC - 537.1 Full Lot		604 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		605 1, DW, PREC - 537.1 Full Lot		606 2, PREC - (MOD) 526plus PLUS TICs	
		1238		607 1, DW, PREC - 537.1 Full Lot		608 2, PREC - (MOD) 526plus PLUS TICs	
		Water		609 1, DW, PREC - 537.1 Full Lot		610 2, PREC - (MOD) 526plus PLUS TICs	
		29-Dec-2025		611 1, DW, PREC - 537.1 Full Lot		612 2, PREC - (MOD) 526plus PLUS TICs	
		1238		613 1, DW, PREC - 537.1 Full Lot		614 2, PREC - (MOD) 526plus PLUS TICs	
		Water		615 1, DW, PREC - 537.1 Full Lot			

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-190141-1

SDG Number: PFAS: Ka'amilo Wells P2

Login Number: 190141

List Number: 1

Creator: Hernandez, Orlando

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	