

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 2/23/2026 3:53:13 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Shaft Viewing Pool

JOB NUMBER

380-199002-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
2/23/2026 3:53:13 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	14
QC Association Summary	25
Lab Chronicle	26
Certification Summary	27
Method Summary	28
Sample Summary	29
Chain of Custody	30
Receipt Checklists	31

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-199002-1

Job ID: 380-199002-1

Eurofins Pomona

Job Narrative 380-199002-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 2/19/2026 9:29 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.6°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-199002-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	3.1		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.4		2.0	ng/L	1		EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0	ng/L	1		EPA 537.1 V2	Total/NA

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona



Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-199002-1

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorohexanesulfonic acid (PFHxS)	3.1		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorooctanesulfonic acid (PFOS)	3.2		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 15:00	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	108		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C6 PFDA	111		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C5 PFHxA	116		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C4 PFHpA	112		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C8 PFOA	116		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C9 PFNA	114		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C7 PFUnA	111		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C2 PFDoA	117		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C4 PFBA	113		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C5 PFPeA	116		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C3 PFBS	110		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C3 PFHxS	113		50 - 200	02/20/26 05:41	02/20/26 15:00	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-199002-1

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	106		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C2-4:2-FTS	124		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C2-6:2-FTS	123		50 - 200	02/20/26 05:41	02/20/26 15:00	1
13C2-8:2-FTS	114		50 - 200	02/20/26 05:41	02/20/26 15:00	1

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorooctanesulfonic acid (PFOS)	3.4		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorohexanesulfonic acid (PFHxS)	3.6		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:37	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	112		70 - 130			02/20/26 02:00	02/20/26 20:37	1
13C2 PFHxA	113		70 - 130			02/20/26 02:00	02/20/26 20:37	1
13C2 PFDA	108		70 - 130			02/20/26 02:00	02/20/26 20:37	1
13C3-GenX	105		70 - 130			02/20/26 02:00	02/20/26 20:37	1

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		02/20/26 05:41	02/20/26 18:18	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	102		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C6 PFDA	108		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C5 PFHxA	111		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C4 PFHpA	111		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C8 PFOA	112		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C9 PFNA	110		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C7 PFUnA	107		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C2 PFDoA	108		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C4 PFBA	108		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C5 PFPeA	109		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C3 PFBS	109		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C3 PFHxS	109		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C8 PFOS	105		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C2-4:2-FTS	118		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C2-6:2-FTS	118		50 - 200			02/20/26 05:41	02/20/26 18:18	1
13C2-8:2-FTS	109		50 - 200			02/20/26 05:41	02/20/26 18:18	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		02/20/26 02:00	02/21/26 14:36	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	109		70 - 130			02/20/26 02:00	02/21/26 14:36	1
13C2 PFHxA	120		70 - 130			02/20/26 02:00	02/21/26 14:36	1
13C2 PFDA	112		70 - 130			02/20/26 02:00	02/21/26 14:36	1
13C3-GenX	110		70 - 130			02/20/26 02:00	02/21/26 14:36	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-199002-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.1		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.2		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	3.4		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	3.6		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	EPA 537.1 V2	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	EPA 537.1 V2	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-199002-1
 SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-199002-1	Halawa Shaft Viewing Pool	112	113	108	105
380-199002-1 MS	Halawa Shaft Viewing Pool	107	111	106	107
380-199002-1 MSD	Halawa Shaft Viewing Pool	108	110	104	104
380-199002-2	Halawa Shaft Viewing Pool	109	120	112	110
	Blank				
LCS 380-207477/20-A	Lab Control Sample	106	93	104	92
MBL 380-207477/18-A	Method Blank	105	94	109	95
MRL 380-207477/19-A	Lab Control Sample	100	95	100	92

Surrogate Legend

- d5NEFOS = d5-NEtFOSAA
- PFHxA = 13C2 PFHxA
- PFDA = 13C2 PFDA
- GenX = 13C3-GenX



Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFD _o A (50-200)
380-199002-1	Halawa Shaft Viewing Pool	108	111	116	112	116	114	111	117
380-199002-1 MS	Halawa Shaft Viewing Pool	109	113	109	110	112	112	112	117
380-199002-1 MSD	Halawa Shaft Viewing Pool	114	114	112	110	114	114	114	114
380-199002-2	Halawa Shaft Viewing Pool	102	108	111	111	112	110	107	108
	Blank								
LCS 380-207478/22-A	Lab Control Sample	109	112	109	107	111	111	112	113
MBL 380-207478/20-A	Method Blank	98	103	108	103	106	103	102	107
MRL 380-207478/21-A	Lab Control Sample	106	114	116	111	115	112	111	112

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-199002-1	Halawa Shaft Viewing Pool	113	116	110	113	106	124	123	114
380-199002-1 MS	Halawa Shaft Viewing Pool	105	111	101	103	101	104	109	108
380-199002-1 MSD	Halawa Shaft Viewing Pool	112	115	111	112	107	116	119	114
380-199002-2	Halawa Shaft Viewing Pool	108	109	109	109	105	118	118	109
	Blank								
LCS 380-207478/22-A	Lab Control Sample	105	106	106	106	105	103	113	111
MBL 380-207478/20-A	Method Blank	103	106	103	104	97	111	112	104
MRL 380-207478/21-A	Lab Control Sample	111	109	110	113	107	117	120	114

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFD_oA = 13C2 PFD_oA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-207478/20-A
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 207478

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		02/20/26 05:41	02/20/26 14:31	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	98		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C6 PFDA	103		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C5 PFHxA	108		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C4 PFHpA	103		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C8 PFOA	106		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C9 PFNA	103		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C7 PFUnA	102		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C2 PFDoA	107		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C4 PFBA	103		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C5 PFPeA	106		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C3 PFBS	103		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C3 PFHxS	104		50 - 200	02/20/26 05:41	02/20/26 14:31	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-207478/20-A
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 207478

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	97		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C2-4:2-FTS	111		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C2-6:2-FTS	112		50 - 200	02/20/26 05:41	02/20/26 14:31	1
13C2-8:2-FTS	104		50 - 200	02/20/26 05:41	02/20/26 14:31	1

Lab Sample ID: LCS 380-207478/22-A
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 207478

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	114		ng/L		95	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	112		ng/L		93	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	109		ng/L		91	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	111		ng/L		92	70 - 130
Perfluorodecanoic acid (PFDA)	120	112		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	120	114		ng/L		95	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	111		ng/L		93	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	114		ng/L		95	70 - 130
Perfluorohexanoic acid (PFHxA)	120	112		ng/L		93	70 - 130
Perfluorononanoic acid (PFNA)	120	112		ng/L		93	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	115		ng/L		96	70 - 130
Perfluorooctanoic acid (PFOA)	120	108		ng/L		90	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	114		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	120	114		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	114		ng/L		95	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	115		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	108		ng/L		90	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	115		ng/L		96	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	120	108		ng/L		90	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	116		ng/L		96	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	113		ng/L		94	70 - 130
Perfluoropentanoic acid (PFPeA)	120	115		ng/L		95	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	116		ng/L		97	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-207478/22-A
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 207478

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	112		ng/L		93	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	109		50 - 200				
13C6 PFDA	112		50 - 200				
13C5 PFHxA	109		50 - 200				
13C4 PFHpA	107		50 - 200				
13C8 PFOA	111		50 - 200				
13C9 PFNA	111		50 - 200				
13C7 PFUnA	112		50 - 200				
13C2 PFDoA	113		50 - 200				
13C4 PFBA	105		50 - 200				
13C5 PFPeA	106		50 - 200				
13C3 PFBS	106		50 - 200				
13C3 PFHxS	106		50 - 200				
13C8 PFOS	105		50 - 200				
13C2-4:2-FTS	103		50 - 200				
13C2-6:2-FTS	113		50 - 200				
13C2-8:2-FTS	111		50 - 200				

Lab Sample ID: MRL 380-207478/21-A
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 207478

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	2.11	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.09	J	ng/L		104	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.01	J	ng/L		100	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.95	J	ng/L		97	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.95	J	ng/L		97	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	1.88	J	ng/L		94	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	1.87	J	ng/L		93	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.90	J	ng/L		95	50 - 150
Perfluorononanoic acid (PFNA)	2.00	1.96	J	ng/L		98	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.99	J	ng/L		99	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	1.87	J	ng/L		93	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-207478/21-A

Matrix: Water

Analysis Batch: 207654

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 207478

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.18	J	ng/L		109	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.19	J	ng/L		109	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.60	J	ng/L		130	50 - 150
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	2.00	1.92	J	ng/L		96	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.00	1.95	J	ng/L		97	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	1.90	J	ng/L		95	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	1.93	J	ng/L		96	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.03	J	ng/L		101	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	2.08	J	ng/L		104	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	1.95	J	ng/L		97	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	106		50 - 200
13C6 PFDA	114		50 - 200
13C5 PFHxA	116		50 - 200
13C4 PFHpA	111		50 - 200
13C8 PFOA	115		50 - 200
13C9 PFNA	112		50 - 200
13C7 PFUnA	111		50 - 200
13C2 PFDoA	112		50 - 200
13C4 PFBA	111		50 - 200
13C5 PFPeA	109		50 - 200
13C3 PFBS	110		50 - 200
13C3 PFHxS	113		50 - 200
13C8 PFOS	107		50 - 200
13C2-4:2-FTS	117		50 - 200
13C2-6:2-FTS	120		50 - 200
13C2-8:2-FTS	114		50 - 200

Lab Sample ID: 380-199002-1 MS

Matrix: Water

Analysis Batch: 207654

Client Sample ID: Halawa Shaft Viewing Pool

Prep Type: Total/NA

Prep Batch: 207478

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	120		ng/L		99	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	119		ng/L		98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	116		ng/L		95	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-199002-1 MS

Client Sample ID: Halawa Shaft Viewing Pool

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 207654

Prep Batch: 207478

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Hexafluoropropylene Oxide	<2.0		121	118		ng/L		97	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	122		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		121	118		ng/L		98	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		121	119		ng/L		98	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		121	116		ng/L		96	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	3.1		121	123		ng/L		99	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		121	116		ng/L		95	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		121	119		ng/L		98	70 - 130
Perfluorooctanesulfonic acid (PFOS)	3.2		121	127		ng/L		102	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		121	119		ng/L		97	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		121	119		ng/L		98	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		121	118		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	120		ng/L		99	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	126		ng/L		104	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	116		ng/L		96	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	115		ng/L		95	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	118		ng/L		97	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	121		ng/L		100	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	115		ng/L		95	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		121	117		ng/L		96	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	121		ng/L		100	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	122		ng/L		101	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	109		50 - 200
13C6 PFDA	113		50 - 200
13C5 PFHxA	109		50 - 200
13C4 PFHpA	110		50 - 200
13C8 PFOA	112		50 - 200
13C9 PFNA	112		50 - 200
13C7 PFUnA	112		50 - 200
13C2 PFDoA	117		50 - 200
13C4 PFBA	105		50 - 200
13C5 PFPeA	111		50 - 200
13C3 PFBS	101		50 - 200
13C3 PFHxS	103		50 - 200
13C8 PFOS	101		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-199002-1 MS
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Halawa Shaft Viewing Pool
Prep Type: Total/NA
Prep Batch: 207478

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C2-4:2-FTS	104		50 - 200
13C2-6:2-FTS	109		50 - 200
13C2-8:2-FTS	108		50 - 200

Lab Sample ID: 380-199002-1 MSD
Matrix: Water
Analysis Batch: 207654

Client Sample ID: Halawa Shaft Viewing Pool
Prep Type: Total/NA
Prep Batch: 207478

Analyte	Sample	Sample	Spike	MSD		Unit	D	%Rec	%Rec	Limits	RPD	RPD	Limit
	Result	Qualifier		Result	Qualifier								
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	116		ng/L		96	70 - 130	4	30		
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	114		ng/L		94	70 - 130	4	30		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	111		ng/L		92	70 - 130	4	30		
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		121	109		ng/L		90	70 - 130	8	30		
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	114		ng/L		94	70 - 130	6	30		
Perfluorodecanoic acid (PFDA)	<2.0		121	112		ng/L		92	70 - 130	6	30		
Perfluorododecanoic acid (PFDoA)	<2.0		121	113		ng/L		93	70 - 130	6	30		
Perfluoroheptanoic acid (PFHpA)	<2.0		121	113		ng/L		93	70 - 130	3	30		
Perfluorohexanesulfonic acid (PFHxS)	3.1		121	113		ng/L		91	70 - 130	8	30		
Perfluorohexanoic acid (PFHxA)	<2.0		121	110		ng/L		90	70 - 130	5	30		
Perfluorononanoic acid (PFNA)	<2.0		121	110		ng/L		91	70 - 130	8	30		
Perfluorooctanesulfonic acid (PFOS)	3.2		121	120		ng/L		97	70 - 130	5	30		
Perfluorooctanoic acid (PFOA)	<2.0		121	111		ng/L		91	70 - 130	7	30		
Perfluoroundecanoic acid (PFUnA)	<2.0		121	111		ng/L		92	70 - 130	6	30		
Perfluorobutanoic acid (PFBA)	<2.0		121	111		ng/L		92	70 - 130	6	30		
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	115		ng/L		95	70 - 130	4	30		
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	109		ng/L		90	70 - 130	15	30		
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	107		ng/L		88	70 - 130	8	30		
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	111		ng/L		92	70 - 130	3	30		
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	108		ng/L		89	70 - 130	8	30		
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	112		ng/L		93	70 - 130	8	30		
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	111		ng/L		92	70 - 130	4	30		
Perfluoropentanoic acid (PFPeA)	<2.0		121	111		ng/L		91	70 - 130	5	30		
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	117		ng/L		97	70 - 130	4	30		
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	113		ng/L		93	70 - 130	8	30		

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	114		50 - 200
13C6 PFDA	114		50 - 200
13C5 PFHxA	112		50 - 200
13C4 PFHpA	110		50 - 200
13C8 PFOA	114		50 - 200
13C9 PFNA	114		50 - 200
13C7 PFUnA	114		50 - 200
13C2 PFDoA	114		50 - 200
13C4 PFBA	112		50 - 200
13C5 PFPeA	115		50 - 200
13C3 PFBS	111		50 - 200
13C3 PFHxS	112		50 - 200
13C8 PFOS	107		50 - 200
13C2-4:2-FTS	116		50 - 200
13C2-6:2-FTS	119		50 - 200
13C2-8:2-FTS	114		50 - 200

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020

Lab Sample ID: MBL 380-207477/18-A
Matrix: Water
Analysis Batch: 207620

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 207477

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		02/20/26 02:00	02/20/26 20:07	1
Surrogate	MBL MBL		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
d5-NEtFOSAA	105		70 - 130			02/20/26 02:00	02/20/26 20:07	1
13C2 PFHxA	94		70 - 130			02/20/26 02:00	02/20/26 20:07	1
13C2 PFDA	109		70 - 130			02/20/26 02:00	02/20/26 20:07	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MBL 380-207477/18-A
Matrix: Water
Analysis Batch: 207620

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 207477

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	95		70 - 130	02/20/26 02:00	02/20/26 20:07	1

Lab Sample ID: LCS 380-207477/20-A
Matrix: Water
Analysis Batch: 207620

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 207477

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	50.1	43.5		ng/L		87	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	50.1	52.8		ng/L		105	70 - 130
Perfluoroundecanoic acid (PFUnA)	50.1	52.3		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.1	50.7		ng/L		101	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.1	50.2		ng/L		100	70 - 130
Perfluorohexanoic acid (PFHxA)	50.1	44.2		ng/L		88	70 - 130
Perfluorododecanoic acid (PFDoA)	50.1	50.0		ng/L		100	70 - 130
Perfluorooctanoic acid (PFOA)	50.1	49.9		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	50.1	50.0		ng/L		100	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.1	51.6		ng/L		103	70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.1	37.3		ng/L		74	70 - 130
Perfluoroheptanoic acid (PFHpA)	50.1	50.6		ng/L		101	70 - 130
Perfluorononanoic acid (PFNA)	50.1	50.1		ng/L		100	70 - 130
Perfluorotetradecanoic acid (PFTA)	50.1	46.7		ng/L		93	70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.1	54.7		ng/L		109	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	50.1	51.3		ng/L		102	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.1	49.8		ng/L		99	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.1	51.0		ng/L		102	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	106		70 - 130
13C2 PFHxA	93		70 - 130
13C2 PFDA	104		70 - 130
13C3-GenX	92		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method: EPA 537.1 V2 - EPA 537.1 Ver. 2.0 March 2020 (Continued)

Lab Sample ID: MRL 380-207477/19-A
Matrix: Water
Analysis Batch: 207620

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 207477

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	1.74	J	ng/L		87	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	1.96	J	ng/L		98	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.10	J	ng/L		105	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.97	J	ng/L		98	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	1.89	J	ng/L		94	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	1.84	J	ng/L		92	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.02	J	ng/L		101	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	1.97	J	ng/L		99	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.01	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.00	J	ng/L		100	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	1.63	J	ng/L		82	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorononanoic acid (PFNA)	2.00	1.99	J	ng/L		99	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.94	J	ng/L		97	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.16	J	ng/L		108	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	1.94	J	ng/L		97	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.94	J	ng/L		97	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.92	J	ng/L		96	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	100		70 - 130
13C2 PFHxA	95		70 - 130
13C2 PFDA	100		70 - 130
13C3-GenX	92		70 - 130

Lab Sample ID: 380-199002-1 MS
Matrix: Water
Analysis Batch: 207620

Client Sample ID: Halawa Shaft Viewing Pool
Prep Type: Total/NA
Prep Batch: 207477

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		50.1	49.6		ng/L		99	70 - 130
Perfluorooctanesulfonic acid (PFOS)	3.4		50.1	52.3		ng/L		98	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		50.1	51.3		ng/L		102	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		50.1	51.5		ng/L		103	70 - 130

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-199002-1
 SDG: PFAS: Halawa Shaft Viewing Pool

LCMS

Prep Batch: 207477

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-199002-1	Halawa Shaft Viewing Pool	Total/NA	Water	537.1 DW	
380-199002-2	Halawa Shaft Viewing Pool Blank	Total/NA	Water	537.1 DW	
MBL 380-207477/18-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-207477/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-207477/19-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-199002-1 MS	Halawa Shaft Viewing Pool	Total/NA	Water	537.1 DW	
380-199002-1 MSD	Halawa Shaft Viewing Pool	Total/NA	Water	537.1 DW	

Prep Batch: 207478

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-199002-1	Halawa Shaft Viewing Pool	Total/NA	Water	533	
380-199002-2	Halawa Shaft Viewing Pool Blank	Total/NA	Water	533	
MBL 380-207478/20-A	Method Blank	Total/NA	Water	533	
LCS 380-207478/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-207478/21-A	Lab Control Sample	Total/NA	Water	533	
380-199002-1 MS	Halawa Shaft Viewing Pool	Total/NA	Water	533	
380-199002-1 MSD	Halawa Shaft Viewing Pool	Total/NA	Water	533	

Analysis Batch: 207620

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-199002-1	Halawa Shaft Viewing Pool	Total/NA	Water	EPA 537.1 V2	207477
MBL 380-207477/18-A	Method Blank	Total/NA	Water	EPA 537.1 V2	207477
LCS 380-207477/20-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	207477
MRL 380-207477/19-A	Lab Control Sample	Total/NA	Water	EPA 537.1 V2	207477
380-199002-1 MS	Halawa Shaft Viewing Pool	Total/NA	Water	EPA 537.1 V2	207477
380-199002-1 MSD	Halawa Shaft Viewing Pool	Total/NA	Water	EPA 537.1 V2	207477

Analysis Batch: 207654

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-199002-1	Halawa Shaft Viewing Pool	Total/NA	Water	533	207478
380-199002-2	Halawa Shaft Viewing Pool Blank	Total/NA	Water	533	207478
MBL 380-207478/20-A	Method Blank	Total/NA	Water	533	207478
LCS 380-207478/22-A	Lab Control Sample	Total/NA	Water	533	207478
MRL 380-207478/21-A	Lab Control Sample	Total/NA	Water	533	207478
380-199002-1 MS	Halawa Shaft Viewing Pool	Total/NA	Water	533	207478
380-199002-1 MSD	Halawa Shaft Viewing Pool	Total/NA	Water	533	207478

Analysis Batch: 207809

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-199002-2	Halawa Shaft Viewing Pool Blank	Total/NA	Water	EPA 537.1 V2	207477

Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-199002-1
 SDG: PFAS: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-199002-1

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			207478	XTD8	EA POM	02/20/26 05:41
Total/NA	Analysis	533		1	207654	Y5FM	EA POM	02/20/26 15:00
Total/NA	Prep	537.1 DW			207477	G9MN	EA POM	02/20/26 02:00
Total/NA	Analysis	EPA 537.1 V2		1	207620	Y5FM	EA POM	02/20/26 20:37

Client Sample ID: Halawa Shaft Viewing Pool Blank

Lab Sample ID: 380-199002-2

Date Collected: 02/17/26 09:35

Matrix: Water

Date Received: 02/19/26 09:29

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			207478	XTD8	EA POM	02/20/26 05:41
Total/NA	Analysis	533		1	207654	Y5FM	EA POM	02/20/26 18:18
Total/NA	Prep	537.1 DW			207477	G9MN	EA POM	02/20/26 02:00
Total/NA	Analysis	EPA 537.1 V2		1	207809	M7ML	EA POM	02/21/26 14:36

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26 *

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

* Accreditation/Certification renewal pending - accreditation/certification considered valid.

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
EPA 537.1 V2	EPA 537.1 Ver. 2.0 March 2020	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-199002-1
SDG: PFAS: Halawa Shaft Viewing Pool

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-199002-1	Halawa Shaft Viewing Pool	Water	02/17/26 09:35	02/19/26 09:29	Hawaii
380-199002-2	Halawa Shaft Viewing Pool Blank	Water	02/17/26 09:35	02/19/26 09:29	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)
 750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100

Chain of Custody Record



Client Information Client Contact: Kirk Iwamoto Company: City & County of Honolulu Address: 630 South Beretania Street, Chemistry Lab City: Honolulu State: HI, Zip: 96843 Phone: 808-748-5840 (tel) Email: kiwamoto@hbws.org Project Name: RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill Site:		Lab PM: Arada, Rachelle E-Mail: Rachelle.Arada@et.euronisus.com PWSID:		Sampler: Jason Rakofsky Phone: +1 808 748 5840		Carrier Tracking No(s): State of Origin:		COC No: 380-27941-2757.2 Page: Page 2 of 2 Job #:	
Due Date Requested: TAT Requested (days): RUSH Compliance Project: Δ No PO #: C20525101 exp 05312023 WO #: 380-199002 COC Project #: 38001111 SSOWN#:		Analysis Requested Perform MS/MSD (Yes or No) <input checked="" type="checkbox"/> Field Filtered Sample (Yes or No) <input checked="" type="checkbox"/> SUBCONTRACT - 625 PAH Physia LL (EAL) + TICs 8016B_GRO_LL (MOD) GRO 8016B_DRO_LL_CB - HNL Ranges: C10-C24/C24-C36/C8-C18 625.2_PREC - (MOD) 626plus PLUS TICs 637.1_DW_PREC - 637.1 Full List 633 - All Analytes		Total Number of Containers:		Preservation Codes: A - HCL B - NaOH C - Zn Acetate D - Nitric Acid E - MeOH F - Amchlor G - Ascorbic Acid H - Ice I - DI Water J - EDTA K - EDTA L - EDA Other: M - Hexane N - None O - AsNaO2 P - Na2O4S Q - Na2SO3 R - Na2S2O3 S - H2SO4 T - TSP Dodecahydrate U - Acetone V - MCAA W - pH 4-5 Y - Trizma Z - other (specify)		Special Instructions/Note:	
Sample Identification Halawa Shaft Viewing Pool Halawa Shaft Viewing Pool Blank		Sample Date: 17-Feb-2026 Sample Time: 0935 Sample Type (C=Comp, G=grab): G Matrix (Residue, Swab, On-surface, BT=Trace, A=Air): Water		R A Q O A Y I 3 3 1 1		Special Instructions/Note:			
Possible Hazard Identification <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		Deliverable Requested: I, II, III, IV, Other (specify)		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month) <input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months		Special Instructions/QC Requirements:			
Empty Kit Relinquished by:		Date: 2/18/26 1100 Company: HBWS		Date/Time: 2/19/26 979 Company: ELAP		Method of Shipment: FedEx 888882082594			
Relinquished by:		Date/Time:		Date/Time:		Date/Time:			
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks: (31A) 3.4 + 0.2 = 3.6 gel-frozen		Ver 01/16/2019			



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-199002-1
SDG Number: PFAS: Halawa Shaft Viewing Pool

Login Number: 199002

List Number: 1

Creator: Tran, Kristine

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

