

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 12/22/2025 5:59:46 PM

JOB DESCRIPTION

RED-HILL
PFAS: Halawa Wells Units 1&2 P1
RUSH Weekly Red Hill

JOB NUMBER

380-188408-1

Eurofins Eaton Analytical Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
12/22/2025 5:59:46 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	33

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-188408-1

Job ID: 380-188408-1

Eurofins Eaton Analytical Pomona

Job Narrative 380-188408-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 12/17/2025 10:30 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.5°C.

PFAS

EPA 537.1 and EPA 533 are two distinct methods for the analysis of PFAS in drinking water. The analyses are conducted on differing instrumentation, with calibrations, extraction solvents and sample preservatives being dissimilar among the two methods. Therefore it is probable and not unexpected to see the methods having slight variations in analytical results: Halawa Wells Units 1 & 2 (331-206-TP065) (380-188408-1). (XWB4)

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.2		2.0	ng/L	1		533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.5		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanoic acid (PFHxA)	2.2		2.0	ng/L	1		537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.6		2.0	ng/L	1		537.1	Total/NA

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-1

Date Collected: 12/15/25 10:35

Matrix: Drinking Water

Date Received: 12/17/25 10:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorohexanesulfonic acid (PFHxS)	2.4		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorooctanesulfonic acid (PFOS)	2.2		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:32	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	97		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C6 PFDA	109		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C5 PFHxA	106		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C4 PFHpA	110		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C8 PFOA	110		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C9 PFNA	112		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C7 PFUnA	108		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C2 PFDoA	106		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C4 PFBA	109		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C5 PFPeA	110		50 - 200			12/19/25 14:06	12/22/25 11:32	1
13C3 PFBS	108		50 - 200			12/19/25 14:06	12/22/25 11:32	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-1

Date Collected: 12/15/25 10:35

Matrix: Drinking Water

Date Received: 12/17/25 10:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 PFHxS	110		50 - 200	12/19/25 14:06	12/22/25 11:32	1
13C8 PFOS	112		50 - 200	12/19/25 14:06	12/22/25 11:32	1
13C2-4:2-FTS	139		50 - 200	12/19/25 14:06	12/22/25 11:32	1
13C2-6:2-FTS	127		50 - 200	12/19/25 14:06	12/22/25 11:32	1
13C2-8:2-FTS	118		50 - 200	12/19/25 14:06	12/22/25 11:32	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorooctanesulfonic acid (PFOS)	2.5		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
N-methylperfluorooctanesulfonamideacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
N-ethylperfluorooctanesulfonamideacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorohexanoic acid (PFHxA)	2.2		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorohexanesulfonic acid (PFHxS)	2.6		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:44	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	94		70 - 130	12/18/25 18:54	12/19/25 11:44	1
13C2 PFHxA	98		70 - 130	12/18/25 18:54	12/19/25 11:44	1
13C2 PFDA	102		70 - 130	12/18/25 18:54	12/19/25 11:44	1
13C3-GenX	102		70 - 130	12/18/25 18:54	12/19/25 11:44	1

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

Date Collected: 12/15/25 10:35

Matrix: Water

Date Received: 12/17/25 10:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

Date Collected: 12/15/25 10:35

Matrix: Water

Date Received: 12/17/25 10:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6 2 FTS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/19/25 14:06	12/22/25 11:42	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	78		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C6 PFDA	79		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C5 PFHxA	88		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C4 PFHpA	87		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C8 PFOA	85		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C9 PFNA	80		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C7 PFUnA	81		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C2 PFDoA	84		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C4 PFBA	98		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C5 PFPeA	95		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C3 PFBS	109		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C3 PFHxS	110		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C8 PFOS	112		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C2-4:2-FTS	129		50 - 200	12/19/25 14:06	12/22/25 11:42	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

Date Collected: 12/15/25 10:35

Matrix: Water

Date Received: 12/17/25 10:30

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C2-6:2-FTS	126		50 - 200	12/19/25 14:06	12/22/25 11:42	1
13C2-8:2-FTS	117		50 - 200	12/19/25 14:06	12/22/25 11:42	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

<i>Analyte</i>	<i>Result</i>	<i>Qualifier</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/18/25 18:54	12/19/25 11:53	1

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
d5-NEtFOSAA	88		70 - 130	12/18/25 18:54	12/19/25 11:53	1
13C2 PFHxA	97		70 - 130	12/18/25 18:54	12/19/25 11:53	1
13C2 PFDA	103		70 - 130	12/18/25 18:54	12/19/25 11:53	1
13C3-GenX	99		70 - 130	12/18/25 18:54	12/19/25 11:53	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.4		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.2		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	2.5		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	2.6		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Surrogate Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-188408-1
 SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-188408-1	HALAWA WELLS UNITS 1 & 2 (94	98	102	102

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-188407-B-1-A MS	Matrix Spike	96	98	96	96
380-188407-C-1-A MSD	Matrix Spike Duplicate	96	107	105	101
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	88	97	103	99
LCS 380-193379/23-A	Lab Control Sample	90	99	101	104
MBL 380-193379/21-A	Method Blank	90	97	101	101
MRL 380-193379/22-A	Lab Control Sample	91	98	101	103

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
 PFHxA = 13C2 PFHxA
 PFDA = 13C2 PFDA
 GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-188408-1	HALAWA WELLS UNITS 1 & 2 (97	109	106	110	110	112	108	106

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-188408-1	HALAWA WELLS UNITS 1 & 2 (109	110	108	110	112	139	127	118

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDaA = 13C2 PFDaA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-188407-E-1-A MS	Matrix Spike	103	106	102	106	107	106	105	105
380-188407-F-1-A MSD	Matrix Spike Duplicate	106	104	102	106	108	106	105	104
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	78	79	88	87	85	80	81	84
LCS 380-193693/22-A	Lab Control Sample	98	103	101	101	105	105	103	102
MBL 380-193693/20-A	Method Blank	96	100	102	105	102	103	100	100
MRL 380-193693/21-A	Lab Control Sample	100	102	107	110	106	108	105	102

Percent Isotope Dilution Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-188407-E-1-A MS	Matrix Spike	108	107	107	107	108	113	111	106
380-188407-F-1-A MSD	Matrix Spike Duplicate	106	106	108	110	105	120	116	111
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	98	95	109	110	112	129	126	117
LCS 380-193693/22-A	Lab Control Sample	104	101	107	108	108	113	110	102
MBL 380-193693/20-A	Method Blank	107	105	108	108	111	120	117	104
MRL 380-193693/21-A	Lab Control Sample	108	108	108	107	110	125	117	105

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

13C5PHA = 13C5 PFHxA
C4PFHA = 13C4 PFHpA
C8PFOA = 13C8 PFOA
C9PFNA = 13C9 PFNA
13C7PUA = 13C7 PFUnA
PFDoA = 13C2 PFDoA
PFBA = 13C4 PFBA
PFPeA = 13C5 PFPeA
C3PFBS = 13C3 PFBS
C3PFHS = 13C3 PFHxS
C8PFOS = 13C8 PFOS
42FTS = 13C2-4:2-FTS
62FTS = 13C2-6:2-FTS
82FTS = 13C2-8:2-FTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-193693/20-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 193693

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosfluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<0.25		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		12/19/25 14:06	12/22/25 09:00	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	96		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C6 PFDA	100		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C5 PFHxA	102		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C4 PFHpA	105		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C8 PFOA	102		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C9 PFNA	103		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C7 PFUnA	100		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C2 PFDoA	100		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C4 PFBA	107		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C5 PFPeA	105		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C3 PFBS	108		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C3 PFHxS	108		50 - 200	12/19/25 14:06	12/22/25 09:00	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-193693/20-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 193693

<i>Isotope Dilution</i>	<i>MBL %Recovery</i>	<i>MBL Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C8 PFOS	111		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C2-4:2-FTS	120		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C2-6:2-FTS	117		50 - 200	12/19/25 14:06	12/22/25 09:00	1
13C2-8:2-FTS	104		50 - 200	12/19/25 14:06	12/22/25 09:00	1

Lab Sample ID: LCS 380-193693/22-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193693

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	60.1	55.2		ng/L		92	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	60.1	57.0		ng/L		95	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	60.1	59.2		ng/L		99	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	60.1	57.9		ng/L		96	70 - 130
Perfluorobutanesulfonic acid (PFBS)	60.1	59.6		ng/L		99	70 - 130
Perfluorodecanoic acid (PFDA)	60.1	60.0		ng/L		100	70 - 130
Perfluorododecanoic acid (PFDoA)	60.1	59.8		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	60.1	60.7		ng/L		101	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	60.1	60.1		ng/L		100	70 - 130
Perfluorohexanoic acid (PFHxA)	60.1	59.2		ng/L		98	70 - 130
Perfluorononanoic acid (PFNA)	60.1	58.5		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	60.1	57.3		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	60.1	57.6		ng/L		96	70 - 130
Perfluoroundecanoic acid (PFUnA)	60.1	59.9		ng/L		100	70 - 130
Perfluorobutanoic acid (PFBA)	60.1	59.1		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	60.1	61.5		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	60.1	61.1		ng/L		102	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	60.1	65.2		ng/L		108	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	60.1	57.3		ng/L		95	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	60.1	61.4		ng/L		102	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	60.1	58.1		ng/L		97	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	60.1	59.2		ng/L		99	70 - 130
Perfluoropentanoic acid (PFPeA)	60.1	61.0		ng/L		101	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	60.1	58.2		ng/L		97	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-193693/22-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193693

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	60.1	59.7		ng/L		99	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	98		50 - 200				
13C6 PFDA	103		50 - 200				
13C5 PFHxA	101		50 - 200				
13C4 PFHpA	101		50 - 200				
13C8 PFOA	105		50 - 200				
13C9 PFNA	105		50 - 200				
13C7 PFUnA	103		50 - 200				
13C2 PFDoA	102		50 - 200				
13C4 PFBA	104		50 - 200				
13C5 PFPeA	101		50 - 200				
13C3 PFBS	107		50 - 200				
13C3 PFHxS	108		50 - 200				
13C8 PFOS	108		50 - 200				
13C2-4:2-FTS	113		50 - 200				
13C2-6:2-FTS	110		50 - 200				
13C2-8:2-FTS	102		50 - 200				

Lab Sample ID: MRL 380-193693/21-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193693

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.81	J	ng/L		90	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.84	J	ng/L		92	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.93	J	ng/L		96	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	1.87	J	ng/L		93	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.02	J	ng/L		100	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	2.04	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.94	J	ng/L		97	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.03	J	ng/L		101	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.98	J	ng/L		99	50 - 150
Perfluorononanoic acid (PFNA)	2.01	1.96	J	ng/L		97	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.98	J	ng/L		99	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.08	J	ng/L		103	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.95	J	ng/L		97	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.04	J	ng/L		102	50 - 150

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-193693/21-A
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193693

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	2.04	J	ng/L		102	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.13	J	ng/L		106	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.21	J	ng/L		110	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	2.02	J	ng/L		101	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	2.06	J	ng/L		103	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	2.04	J	ng/L		101	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	1.98	J	ng/L		98	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.84	J	ng/L		92	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	2.03	J	ng/L		101	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	100		50 - 200
13C6 PFDA	102		50 - 200
13C5 PFHxA	107		50 - 200
13C4 PFHpA	110		50 - 200
13C8 PFOA	106		50 - 200
13C9 PFNA	108		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	102		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	108		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	110		50 - 200
13C2-4:2-FTS	125		50 - 200
13C2-6:2-FTS	117		50 - 200
13C2-8:2-FTS	105		50 - 200

Lab Sample ID: 380-188407-E-1-A MS
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 193693

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	112		ng/L		92	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	112		ng/L		93	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	113		ng/L		93	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-188407-E-1-A MS
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 193693

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		121	116		ng/L		96	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	118		ng/L		98	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		121	113		ng/L		94	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		121	114		ng/L		95	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		121	113		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	119		ng/L		98	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		121	115		ng/L		95	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		121	113		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	115		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		121	114		ng/L		95	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		121	115		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		121	117		ng/L		97	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	118		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	116		ng/L		96	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	119		ng/L		99	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	118		ng/L		98	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	117		ng/L		97	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	116		ng/L		96	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	117		ng/L		97	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		121	119		ng/L		98	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	113		ng/L		94	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	119		ng/L		98	70 - 130

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>
13C3 HFPO-DA	103		50 - 200
13C6 PFDA	106		50 - 200
13C5 PFHxA	102		50 - 200
13C4 PFHpA	106		50 - 200
13C8 PFOA	107		50 - 200
13C9 PFNA	106		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	105		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	107		50 - 200
13C3 PFBS	107		50 - 200
13C3 PFHxS	107		50 - 200
13C8 PFOS	108		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-188407-E-1-A MS
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 193693

<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>MS MS Qualifier</i>	<i>Limits</i>
13C2-4:2-FTS	113		50 - 200
13C2-6:2-FTS	111		50 - 200
13C2-8:2-FTS	106		50 - 200

Lab Sample ID: 380-188407-F-1-A MSD
Matrix: Water
Analysis Batch: 193904

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 193693

<i>Analyte</i>	<i>Sample Result</i>	<i>Sample Qualifier</i>	<i>Spike Added</i>	<i>MSD Result</i>	<i>MSD Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>	<i>RPD</i>	<i>RPD Limit</i>
11-Chloroeicosfluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	115		ng/L		95	70 - 130	3	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	116		ng/L		96	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	115		ng/L		95	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		121	114		ng/L		95	70 - 130	1	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	116		ng/L		96	70 - 130	2	30
Perfluorodecanoic acid (PFDA)	<2.0		121	120		ng/L		100	70 - 130	6	30
Perfluorododecanoic acid (PFDoA)	<2.0		121	116		ng/L		96	70 - 130	2	30
Perfluoroheptanoic acid (PFHpA)	<2.0		121	116		ng/L		96	70 - 130	3	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	118		ng/L		97	70 - 130	1	30
Perfluorohexanoic acid (PFHxA)	<2.0		121	117		ng/L		97	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		121	114		ng/L		94	70 - 130	0	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	116		ng/L		96	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		121	117		ng/L		97	70 - 130	2	30
Perfluoroundecanoic acid (PFUnA)	<2.0		121	119		ng/L		99	70 - 130	3	30
Perfluorobutanoic acid (PFBA)	<2.0		121	118		ng/L		98	70 - 130	1	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	113		ng/L		93	70 - 130	5	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	113		ng/L		93	70 - 130	3	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	122		ng/L		101	70 - 130	2	30
Nonfluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	121		ng/L		100	70 - 130	2	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		121	116		ng/L		96	70 - 130	1	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	119		ng/L		99	70 - 130	3	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	117		ng/L		97	70 - 130	0	30
Perfluoropentanoic acid (PFPeA)	<2.0		121	119		ng/L		98	70 - 130	0	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	119		ng/L		99	70 - 130	5	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	115		ng/L		95	70 - 130	3	30

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

<i>Isotope Dilution</i>	<i>MSD</i>	<i>MSD</i>	<i>Limits</i>
<u>Isotope Dilution</u>	<u>%Recovery</u>	<u>Qualifier</u>	<u>Limits</u>
13C3 HFPO-DA	106		50 - 200
13C6 PFDA	104		50 - 200
13C5 PFHxA	102		50 - 200
13C4 PFHpA	106		50 - 200
13C8 PFOA	108		50 - 200
13C9 PFNA	106		50 - 200
13C7 PFUnA	105		50 - 200
13C2 PFDoA	104		50 - 200
13C4 PFBA	106		50 - 200
13C5 PFPeA	106		50 - 200
13C3 PFBS	108		50 - 200
13C3 PFHxS	110		50 - 200
13C8 PFOS	105		50 - 200
13C2-4:2-FTS	120		50 - 200
13C2-6:2-FTS	116		50 - 200
13C2-8:2-FTS	111		50 - 200

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-193379/21-A
Matrix: Water
Analysis Batch: 193571

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 193379

<u>Analyte</u>	<u>MBL</u>	<u>MBL</u>	<u>RL</u>	<u>Unit</u>	<u>D</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
<u>Analyte</u>	<u>Result</u>	<u>Qualifier</u>	<u>RL</u>	<u>Unit</u>	<u>D</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
Perfluorotridecanoic acid (PFTrDA)	<0.36		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		12/18/25 18:54	12/19/25 09:59	1
<u>Surrogate</u>	<u>MBL</u>	<u>MBL</u>	<u>Limits</u>			<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
<u>Surrogate</u>	<u>%Recovery</u>	<u>Qualifier</u>	<u>Limits</u>			<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
d5-NEtFOSAA	90		70 - 130			12/18/25 18:54	12/19/25 09:59	1
13C2 PFHxA	97		70 - 130			12/18/25 18:54	12/19/25 09:59	1
13C2 PFDA	101		70 - 130			12/18/25 18:54	12/19/25 09:59	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MBL 380-193379/21-A
Matrix: Water
Analysis Batch: 193571

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 193379

<i>Surrogate</i>	<i>MBL</i>	<i>MBL</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3-GenX	101	Qualifier	70 - 130	12/18/25 18:54	12/19/25 09:59	1

Lab Sample ID: LCS 380-193379/23-A
Matrix: Water
Analysis Batch: 193571

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193379

<i>Analyte</i>	<i>Spike Added</i>	<i>LCS Result</i>	<i>LCS Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	50.1	54.6	Qualifier	ng/L	-	109	70 - 130
Perfluorooctanesulfonic acid (PFOS)	50.1	52.7	Qualifier	ng/L	-	105	70 - 130
Perfluoroundecanoic acid (PFUnA)	50.1	54.9	Qualifier	ng/L	-	110	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	50.1	51.4	Qualifier	ng/L	-	103	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	50.1	47.7	Qualifier	ng/L	-	95	70 - 130
Perfluorohexanoic acid (PFHxA)	50.1	48.9	Qualifier	ng/L	-	98	70 - 130
Perfluorododecanoic acid (PFDoA)	50.1	54.9	Qualifier	ng/L	-	110	70 - 130
Perfluorooctanoic acid (PFOA)	50.1	53.5	Qualifier	ng/L	-	107	70 - 130
Perfluorodecanoic acid (PFDA)	50.1	55.2	Qualifier	ng/L	-	110	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	50.1	51.2	Qualifier	ng/L	-	102	70 - 130
Perfluorobutanesulfonic acid (PFBS)	50.1	50.8	Qualifier	ng/L	-	101	70 - 130
Perfluoroheptanoic acid (PFHpA)	50.1	52.9	Qualifier	ng/L	-	106	70 - 130
Perfluorononanoic acid (PFNA)	50.1	54.0	Qualifier	ng/L	-	108	70 - 130
Perfluorotetradecanoic acid (PFTA)	50.1	50.7	Qualifier	ng/L	-	101	70 - 130
Perfluorotridecanoic acid (PFTrDA)	50.1	52.2	Qualifier	ng/L	-	104	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	50.1	52.0	Qualifier	ng/L	-	104	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	50.1	49.2	Qualifier	ng/L	-	98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	50.1	50.5	Qualifier	ng/L	-	101	70 - 130

<i>Surrogate</i>	<i>LCS</i>	<i>LCS</i>	<i>Limits</i>
d5-NEtFOSAA	90	Qualifier	70 - 130
13C2 PFHxA	99	Qualifier	70 - 130
13C2 PFDA	101	Qualifier	70 - 130
13C3-GenX	104	Qualifier	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MRL 380-193379/22-A
Matrix: Water
Analysis Batch: 193571

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 193379

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.20	J	ng/L		110	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.03	J	ng/L		101	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.17	J	ng/L		108	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.00	1.97	J	ng/L		98	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.30	J	ng/L		115	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.30	J	ng/L		115	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.06	J	ng/L		103	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.05	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.12	J	ng/L		106	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.20	J	ng/L		110	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.00	1.98	J	ng/L		99	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.00	2.03	J	ng/L		101	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	2.00	2.04	J	ng/L		102	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.87	J	ng/L		93	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	2.05	J	ng/L		103	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	91		70 - 130
13C2 PFHxA	98		70 - 130
13C2 PFDA	101		70 - 130
13C3-GenX	103		70 - 130

Lab Sample ID: 380-188407-B-1-A MS
Matrix: Water
Analysis Batch: 193741

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 193379

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	25.5		ng/L		102	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.1	27.2		ng/L		108	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	27.0		ng/L		107	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	26.2		ng/L		104	70 - 130

Eurofins Eaton Analytical Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

LCMS

Prep Batch: 193379

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-188408-1	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	537.1 DW	
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Water	537.1 DW	
MBL 380-193379/21-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-193379/23-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-193379/22-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-188407-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-188407-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 193571

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-188408-1	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	537.1	193379
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Water	537.1	193379
MBL 380-193379/21-A	Method Blank	Total/NA	Water	537.1	193379
LCS 380-193379/23-A	Lab Control Sample	Total/NA	Water	537.1	193379
MRL 380-193379/22-A	Lab Control Sample	Total/NA	Water	537.1	193379

Prep Batch: 193693

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-188408-1	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	533	
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Water	533	
MBL 380-193693/20-A	Method Blank	Total/NA	Water	533	
LCS 380-193693/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-193693/21-A	Lab Control Sample	Total/NA	Water	533	
380-188407-E-1-A MS	Matrix Spike	Total/NA	Water	533	
380-188407-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 193741

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-188407-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	193379
380-188407-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	193379

Analysis Batch: 193904

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-188408-1	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	533	193693
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Water	533	193693
MBL 380-193693/20-A	Method Blank	Total/NA	Water	533	193693
LCS 380-193693/22-A	Lab Control Sample	Total/NA	Water	533	193693
MRL 380-193693/21-A	Lab Control Sample	Total/NA	Water	533	193693
380-188407-E-1-A MS	Matrix Spike	Total/NA	Water	533	193693
380-188407-F-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	193693

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

**Client Sample ID: HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-1

Date Collected: 12/15/25 10:35

Matrix: Drinking Water

Date Received: 12/17/25 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			193693	N8NE	EA POM	12/19/25 14:06
Total/NA	Analysis	533		1	193904	SZ9R	EA POM	12/22/25 11:32
Total/NA	Prep	537.1 DW			193379	N8NE	EA POM	12/18/25 18:54
Total/NA	Analysis	537.1		1	193571	Y5FM	EA POM	12/19/25 11:44

**Client Sample ID: FB HALAWA WELLS UNITS 1 & 2
(331-206-TP065)**

Lab Sample ID: 380-188408-2

Date Collected: 12/15/25 10:35

Matrix: Water

Date Received: 12/17/25 10:30

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			193693	N8NE	EA POM	12/19/25 14:06
Total/NA	Analysis	533		1	193904	SZ9R	EA POM	12/22/25 11:42
Total/NA	Prep	537.1 DW			193379	N8NE	EA POM	12/18/25 18:54
Total/NA	Analysis	537.1		1	193571	Y5FM	EA POM	12/19/25 11:53

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Laboratory: Eurofins Eaton Analytical Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-188408-1
SDG: PFAS: Halawa Wells Units 1&2 P1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-188408-1	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Drinking Water	12/15/25 10:35	12/17/25 10:30	Hawaii
380-188408-2	FB HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Water	12/15/25 10:35	12/17/25 10:30	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Monrovia, CA (Suite 100)

750 Royal Oaks Drive Suite 100
 Monrovia, CA 91016
 Phone (626) 386-1100

Chain of Custody Record

Client Information		Sampler bailey		Lab PM. Lopez, Maria		Carrier Tracking No(s):		COC No					
Client Contact: kirk iwamoto		Phone +1 808 748 5840		E-Mail Maria.Lopez@et.eurofinsus.com		State of Origin:		Page Page 1 of 1					
Company City & County of Honolulu		PWSID:		Analysis Requested						Job #:			
Address: 630 South Beretania Street, Chemistry Lab		Due Date Requested:		Field Filtered Sample (Yes or No) Perform MS/MSD (Yes or No) SUBCONTRACT - 625 PAH Physis LL (EAL) + TICs 8018B_GRO_LL - (MOD) GRO 8018B_DRO_LL_CS - HNL Ranges. C10-C24/C24-C36/C8-C16 525.2_PREC - (MOD) 525plus PLUS TICs 537.1_DW_PREC - 537.1 Full List 533 - All Analyses						Preservation Codes:			
City Honolulu		TAT Requested (days):								A - HCL		M - Hexane	
State, Zip: HI, 96843		Compliance Project. Δ No								B - NaOH		N - None	
Phone: 808-748-5840 (tel)		PO #: C20525101 exp 05312023								C - Zn Acetate		O - AsNaO2	
Email: kiwamoto@hbws.org		WO #:								D - Nitric Acid		P - Na2O4S	
Project Name RED-HILL/HBWS sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111		E - NaHSO4		Q - Na2SO3							
Site		SSOW#:		F - MeOH		R Na2S2O3							
				G - Archlor		S - H2SO4							
				H - Ascorbic Acid		T - TSP Dodecahydrate							
				I - Ice		U - Acetone							
				J - DI Water		V - MCAA							
				K - EDTA		W - pH 4-5							
				L - EDA		Y - Trizma							
				Other:		Z - other (specify)							
Sample Identification		Sample Date		Sample Time		Sample Type (C=comp, G=grab)		Matrix (W=water, S=solid, O=waste/Oil, BT=Trizma, A=All)		Total Number of Containers		Special Instructions/Note:	
Halawa Wells Units 1&2 P1		15-Dec-2025		1035		G		Water		R		A Q QA Y I 3 3	
FB Halawa Wells Units 1&2		15-Dec-2025		1035						1 1		380-188408 COC	
Possible Hazard Identification						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)							
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months							
Deliverable Requested I, II, III, IV, Other (specify)						Special Instructions/QC Requirements							
Empty Kit Relinquished by:		Date		Time:		Method of Shipment: FedEx 8871 5104 9470							
Relinquished by:		Date/Time: 16 December 2025 1400		Company: HBWS		Received by: Smolok Drake B		Date/Time: 12-22-25 10 30		Company: FEAP			
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:			
Custody Seals Intact Δ Yes Δ No		Custody Seal No.		Cooler Temperature(s) °C and Other Remarks: 401A) 1.7-0.2=1.5 Gel frozen									

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-188408-1

SDG Number: PFAS: Halawa Wells Units 1&2 P1

Login Number: 188408

List Number: 1

Creator: Ngo, Theodore

List Source: Eurofins Eaton Analytical Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	