

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

# ANALYTICAL REPORT

## PREPARED FOR

Attn: Mr. Erwin Kawata  
City & County of Honolulu  
630 South Beretania Street  
Public Service Bldg. Room 310  
Honolulu, Hawaii 96843

Generated 12/15/2025 4:55:28 PM

## JOB DESCRIPTION

RED-HILL  
Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea Wells Pumps 1&2 P2  
RUSH Weekly Red Hill

## JOB NUMBER

380-185035-1

# Eurofins Eaton Analytical Pomona

## Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

## Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

## Authorization



Authorized for release by  
Maria Lopez, Project Manager  
[Maria.Lopez@et.eurofinsus.com](mailto:Maria.Lopez@et.eurofinsus.com)  
(626)386-1100

Generated  
12/15/2025 4:55:28 PM



# Table of Contents

Cover Page . . . . .	1
Table of Contents . . . . .	3
Definitions/Glossary . . . . .	4
Case Narrative . . . . .	5
Detection Summary . . . . .	6
Client Sample Results . . . . .	7
Action Limit Summary . . . . .	14
Surrogate Summary . . . . .	15
QC Sample Results . . . . .	18
QC Association Summary . . . . .	34
Lab Chronicle . . . . .	36
Certification Summary . . . . .	38
Method Summary . . . . .	40
Sample Summary . . . . .	41
Chain of Custody . . . . .	42
Receipt Checklists . . . . .	46

# Definitions/Glossary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Qualifiers

### GC/MS Semi VOA

Qualifier	Qualifier Description
*+	LCS and/or LCSD is outside acceptance limits, high biased.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

### GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

### GC Semi VOA

Qualifier	Qualifier Description
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

# Case Narrative

Client: City & County of Honolulu  
Project: RED-HILL

Job ID: 380-185035-1

**Job ID: 380-185035-1**

**Eurofins Eaton Analytical Pomona**

## Job Narrative 380-185035-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

### Receipt

The samples were received on 11/28/2025 10:12 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperatures of the 2 coolers at receipt time were 3.3°C and 3.6°C.

### Receipt Exceptions

One or more containers for the following samples were received broken or leaking: HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP065) (380-185035-1) and AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400) (380-185035-2).

1 of 2 625 containers for site Halawa Wells received broken, enough volume was available for analysis.

1 of 3 8015\_GRO vials for site AIEA WELLS received broken, enough volume was available for analysis.

### GC/MS Semi VOA

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

### Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

### Diesel Range Organics

Method 8015B\_DRO\_LL\_CS: Method 8015B\_DRO results for HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP065) (380-185035-1), HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1[MS]), HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1[MSD]) and AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400) (380-185035-2) collected on 11/24/25 are a resample for job # 380-183491 collected on 11/17/25. (XWB4)

Method 8015B\_DRO\_LL\_CS: The method reporting limit check (MRL) for preparation batch 570-663324 and analytical batch 570-666060 recovered outside control limits for the following analytes: C10-C28. These analytes were biased high in the MRL and were not detected in the associated samples; therefore, the data have been reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

# Detection Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1  
(331-206-TP065)**  
PWSID Number: HI0000331

**Lab Sample ID: 380-185035-1**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Dieldrin	0.032		0.0097	ug/L	1		525.2	Total/NA

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2  
(331-203-TP400)**  
PWSID Number: HI0000331

**Lab Sample ID: 380-185035-2**

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Dieldrin	0.013		0.0097	ug/L	1		525.2	Total/NA

**Client Sample ID: TB: HALAWA WELLS UNITS 1&2  
(331-206-TP065)**

**Lab Sample ID: 380-185035-3**

No Detections.

**Client Sample ID: TB: AIEA WELLS PUMPS 1&2 (260)  
(331-203-TP400)**

**Lab Sample ID: 380-185035-4**

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1  
(331-206-TP065)**

**Lab Sample ID: 380-185035-1**

**Date Collected: 11/24/25 10:58  
Date Received: 11/28/25 10:12**

**Matrix: Drinking Water  
PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2,4' DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2,4-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2,6-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
2-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
4,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
4,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
4,4'-DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Acenaphthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Acenaphthylene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Acetochlor	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Alachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
alpha-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
alpha-Chlordane	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Anthracene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:24	1
Atrazine	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Benz(a)anthracene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Benzo[a]pyrene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:24	1
Benzo[b]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:24	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Benzo[k]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:24	1
beta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Bis(2-ethylhexyl) phthalate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 16:24	1
Bromacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Butachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Butylbenzylphthalate	<0.49		0.49	ug/L		12/02/25 08:58	12/03/25 16:24	1
Chlorobenzilate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Chloroneb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Chlorothalonil (Draconil, Bravo)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Chlorpyrifos	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Chrysene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:24	1
delta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Di(2-ethylhexyl)adipate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 16:24	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
<b>Dieldrin</b>	<b>0.032</b>		0.0097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Diethylphthalate	<0.49	+	0.49	ug/L		12/02/25 08:58	12/03/25 16:24	1
Dimethylphthalate	<0.49		0.49	ug/L		12/02/25 08:58	12/03/25 16:24	1
Di-n-butyl phthalate	<0.97		0.97	ug/L		12/02/25 08:58	12/03/25 16:24	1
Di-n-octyl phthalate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Endosulfan I (Alpha)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Endosulfan II (Beta)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Endosulfan sulfate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Endrin	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Endrin aldehyde	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1  
(331-206-TP065)**

**Lab Sample ID: 380-185035-1**

**Date Collected: 11/24/25 10:58**

**Matrix: Drinking Water**

**Date Received: 11/28/25 10:12**

**PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
EPTC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Fluoranthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Fluorene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
gamma-Chlordane	0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Heptachlor	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Heptachlor epoxide (isomer B)	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Hexachlorobenzene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Isophorone	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Lindane	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Malathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Methoxychlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Metolachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Molinate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Naphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Parathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Pendimethalin (Penoxaline)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Phenanthrene	<0.039		0.039	ug/L		12/02/25 08:58	12/03/25 16:24	1
Propachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Pyrene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Simazine	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Terbacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Terbutylazine	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Thiobencarb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1
Total Permethrin (mixed isomers)	<0.19		0.19	ug/L		12/02/25 08:58	12/03/25 16:24	1
trans-Nonachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 16:24	1
Trifluralin	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:24	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	12/02/25 08:58	12/03/25 16:24	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	97		70 - 130	12/02/25 08:58	12/03/25 16:24	1
Perylene-d12	97		70 - 130	12/02/25 08:58	12/03/25 16:24	1
Triphenylphosphate	108		70 - 130	12/02/25 08:58	12/03/25 16:24	1

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
2-Methylnaphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Acenaphthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Acenaphthylene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Benzo[a]anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Benzo[a]pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1  
(331-206-TP065)**

**Lab Sample ID: 380-185035-1**

**Date Collected: 11/24/25 10:58**

**Matrix: Drinking Water**

**Date Received: 11/28/25 10:12**

**PWSID Number: HI0000331**

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Benzo[k]fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Chrysene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Fluorene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Naphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Phenanthrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1
Pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:08	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	65		28 - 127	12/01/25 08:21	12/11/25 11:08	1
2-Fluorobiphenyl (Surr)	62		31 - 120	12/01/25 08:21	12/11/25 11:08	1
2-Fluorophenol (Surr)	37		17 - 120	12/01/25 08:21	12/11/25 11:08	1
Nitrobenzene-d5 (Surr)	62		27 - 120	12/01/25 08:21	12/11/25 11:08	1
Phenol-d6 (Surr)	23		10 - 120	12/01/25 08:21	12/11/25 11:08	1
p-Terphenyl-d14 (Surr)	79		45 - 120	12/01/25 08:21	12/11/25 11:08	1

**Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)**

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	12/01/25 08:21	12/11/25 13:23	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	57		33 - 139	12/01/25 08:21	12/11/25 13:23	1
2-Fluorobiphenyl (Surr)	67		33 - 126	12/01/25 08:21	12/11/25 13:23	1
2-Fluorophenol (Surr)	36		12 - 120	12/01/25 08:21	12/11/25 13:23	1
Nitrobenzene-d5 (Surr)	67		36 - 120	12/01/25 08:21	12/11/25 13:23	1
Phenol-d6 (Surr)	21		10 - 120	12/01/25 08:21	12/11/25 13:23	1
p-Terphenyl-d14 (Surr)	81		47 - 131	12/01/25 08:21	12/11/25 13:23	1

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			12/07/25 17:44	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	86		38 - 134		12/07/25 17:44	1

**Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		12/01/25 16:05	12/07/25 02:38	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		12/01/25 16:05	12/07/25 02:38	1
C8-C18	<26		26	ug/L		12/01/25 16:05	12/07/25 02:38	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	113		60 - 130	12/01/25 16:05	12/07/25 02:38	1

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2  
(331-203-TP400)**

**Lab Sample ID: 380-185035-2**

**Date Collected: 11/24/25 12:44**

**Matrix: Drinking Water**

**Date Received: 11/28/25 10:12**

**PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2,4' DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2,4-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2,6-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
2-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
4,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
4,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
4,4'-DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Acenaphthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Acenaphthylene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Acetochlor	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Alachlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
alpha-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
alpha-Chlordane	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Anthracene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:44	1
Atrazine	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Benz(a)anthracene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Benzo[a]pyrene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:44	1
Benzo[b]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:44	1
Benzo[g,h,i]perylene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Benzo[k]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:44	1
beta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Bis(2-ethylhexyl) phthalate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 16:44	1
Bromacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Butachlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Butylbenzylphthalate	<0.48		0.48	ug/L		12/02/25 08:58	12/03/25 16:44	1
Chlorobenzilate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Chloroneb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Chlorothalonil (Draconil, Bravo)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Chlorpyrifos	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Chrysene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 16:44	1
delta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Di(2-ethylhexyl)adipate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 16:44	1
Dibenz(a,h)anthracene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Diclorvos (DDVP)	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
<b>Dieldrin</b>	<b>0.013</b>		0.0097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Diethylphthalate	<0.48	+	0.48	ug/L		12/02/25 08:58	12/03/25 16:44	1
Dimethylphthalate	<0.48		0.48	ug/L		12/02/25 08:58	12/03/25 16:44	1
Di-n-butyl phthalate	<0.97		0.97	ug/L		12/02/25 08:58	12/03/25 16:44	1
Di-n-octyl phthalate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Endosulfan I (Alpha)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Endosulfan II (Beta)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Endosulfan sulfate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Endrin	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Endrin aldehyde	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2  
(331-203-TP400)**

**Lab Sample ID: 380-185035-2**

**Date Collected: 11/24/25 12:44**

**Matrix: Drinking Water**

**Date Received: 11/28/25 10:12**

**PWSID Number: HI0000331**

**Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
EPTC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Fluoranthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Fluorene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
gamma-Chlordane	0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Heptachlor	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Heptachlor epoxide (isomer B)	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Hexachlorobenzene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Hexachlorocyclopentadiene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Indeno[1,2,3-cd]pyrene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Isophorone	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Lindane	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Malathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Methoxychlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Metolachlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Molinate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Naphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Parathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Pendimethalin (Penoxaline)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Phenanthrene	<0.039		0.039	ug/L		12/02/25 08:58	12/03/25 16:44	1
Propachlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Pyrene	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Simazine	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Terbacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Terbutylazine	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Thiobencarb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1
Total Permethrin (mixed isomers)	<0.19		0.19	ug/L		12/02/25 08:58	12/03/25 16:44	1
trans-Nonachlor	<0.048		0.048	ug/L		12/02/25 08:58	12/03/25 16:44	1
Trifluralin	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 16:44	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	12/02/25 08:58	12/03/25 16:44	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	97		70 - 130	12/02/25 08:58	12/03/25 16:44	1
Perylene-d12	97		70 - 130	12/02/25 08:58	12/03/25 16:44	1
Triphenylphosphate	105		70 - 130	12/02/25 08:58	12/03/25 16:44	1

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
2-Methylnaphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Acenaphthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Acenaphthylene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Benzo[a]anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Benzo[a]pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1

Eurofins Eaton Analytical Pomona

# Client Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2  
(331-203-TP400)**

**Lab Sample ID: 380-185035-2**

**Date Collected: 11/24/25 12:44**

**Matrix: Drinking Water**

**Date Received: 11/28/25 10:12**

**PWSID Number: HI0000331**

**Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Benzo[k]fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Chrysene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Fluoranthene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Fluorene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Naphthalene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Phenanthrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1
Pyrene	<0.19		0.19	ug/L		12/01/25 08:21	12/11/25 11:30	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	67		28 - 127	12/01/25 08:21	12/11/25 11:30	1
2-Fluorobiphenyl (Surr)	56		31 - 120	12/01/25 08:21	12/11/25 11:30	1
2-Fluorophenol (Surr)	32		17 - 120	12/01/25 08:21	12/11/25 11:30	1
Nitrobenzene-d5 (Surr)	56		27 - 120	12/01/25 08:21	12/11/25 11:30	1
Phenol-d6 (Surr)	21		10 - 120	12/01/25 08:21	12/11/25 11:30	1
p-Terphenyl-d14 (Surr)	73		45 - 120	12/01/25 08:21	12/11/25 11:30	1

**Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)**

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	12/01/25 08:21	12/11/25 13:47	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	57		33 - 139	12/01/25 08:21	12/11/25 13:47	1
2-Fluorobiphenyl (Surr)	60		33 - 126	12/01/25 08:21	12/11/25 13:47	1
2-Fluorophenol (Surr)	33		12 - 120	12/01/25 08:21	12/11/25 13:47	1
Nitrobenzene-d5 (Surr)	61		36 - 120	12/01/25 08:21	12/11/25 13:47	1
Phenol-d6 (Surr)	17		10 - 120	12/01/25 08:21	12/11/25 13:47	1
p-Terphenyl-d14 (Surr)	79		47 - 131	12/01/25 08:21	12/11/25 13:47	1

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			12/07/25 18:50	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	91		38 - 134		12/07/25 18:50	1

**Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		12/01/25 16:05	12/07/25 05:51	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		12/01/25 16:05	12/07/25 05:51	1
C8-C18	<26		26	ug/L		12/01/25 16:05	12/07/25 05:51	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	111		60 - 130	12/01/25 16:05	12/07/25 05:51	1

# Client Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

**Client Sample ID: TB: HALAWA WELLS UNITS 1&2  
 (331-206-TP065)**

**Lab Sample ID: 380-185035-3**

Date Collected: 11/24/25 10:58  
 Date Received: 11/28/25 10:12

Matrix: Water

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			12/08/25 01:02	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	94		38 - 134				12/08/25 01:02	1

**Client Sample ID: TB: AIEA WELLS PUMPS 1&2 (260)  
 (331-203-TP400)**

**Lab Sample ID: 380-185035-4**

Date Collected: 11/24/25 12:44  
 Date Received: 11/28/25 10:12

Matrix: Water

**Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)**

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			12/08/25 01:24	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	93		38 - 134				12/08/25 01:24	1

# Action Limit Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1**  
**(331-206-TP065)**  
**PWSID Number: HI0000331**

**Lab Sample ID: 380-185035-1**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.019		ug/L	0.2	0.019	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.58		ug/L	6	0.58	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.58		ug/L	400	0.58	525.2	Total/NA
Endrin	<0.0097		ug/L	2	0.0097	525.2	Total/NA
Heptachlor	<0.0097		ug/L	0.4	0.0097	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.0097		ug/L	0.2	0.0097	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.0097		ug/L	0.2	0.0097	525.2	Total/NA
Methoxychlor	<0.049		ug/L	40	0.049	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2**  
**(331-203-TP400)**  
**PWSID Number: HI0000331**

**Lab Sample ID: 380-185035-2**

## Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.048		ug/L	2	0.048	525.2	Total/NA
Atrazine	<0.048		ug/L	3	0.048	525.2	Total/NA
Benzo[a]pyrene	<0.019		ug/L	0.2	0.019	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.58		ug/L	6	0.58	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.58		ug/L	400	0.58	525.2	Total/NA
Endrin	<0.0097		ug/L	2	0.0097	525.2	Total/NA
Heptachlor	<0.0097		ug/L	0.4	0.0097	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.0097		ug/L	0.2	0.0097	525.2	Total/NA
Hexachlorobenzene	<0.048		ug/L	1	0.048	525.2	Total/NA
Hexachlorocyclopentadiene	<0.048		ug/L	50	0.048	525.2	Total/NA
Lindane	<0.0097		ug/L	0.2	0.0097	525.2	Total/NA
Methoxychlor	<0.048		ug/L	40	0.048	525.2	Total/NA
Simazine	<0.048		ug/L	4	0.048	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-185035-1	HALAWA WELLS UNITS 1 & 2 F	97	97	108
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	97	97	105

### Surrogate Legend

2NMX = 2-Nitro-m-xylene  
PRY = Perylene-d12  
TPP = Triphenylphosphate

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-185034-I-1-A MS	Matrix Spike	97	99	109
380-185034-J-1-A MSD	Matrix Spike Duplicate	98	98	112
LCS 380-189440/22-A	Lab Control Sample	97	95	108
MB 380-189440/20-A	Method Blank	95	89	104
MRL 380-189440/21-A	Lab Control Sample	97	94	106

### Surrogate Legend

2NMX = 2-Nitro-m-xylene  
PRY = Perylene-d12  
TPP = Triphenylphosphate

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-185035-1	HALAWA WELLS UNITS 1 & 2 F	57	67	36	67	21	81
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	57	60	33	61	17	79

### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)  
FBP = 2-Fluorobiphenyl (Surr)  
2FP = 2-Fluorophenol (Surr)  
NBZ = Nitrobenzene-d5 (Surr)  
PHL6 = Phenol-d6 (Surr)  
TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
MB 570-663011/1-A	Method Blank	75	84	53	89	33	83

# Surrogate Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)  
FBP = 2-Fluorobiphenyl (Surr)  
2FP = 2-Fluorophenol (Surr)  
NBZ = Nitrobenzene-d5 (Surr)  
PHL6 = Phenol-d6 (Surr)  
TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-185035-1	HALAWA WELLS UNITS 1 & 2 F	65	62	37	62	23	79
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	83	86	61	71	40	84
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	81	79	53	65	35	84
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	67	56	32	56	21	73

### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)  
FBP = 2-Fluorobiphenyl (Surr)  
2FP = 2-Fluorophenol (Surr)  
NBZ = Nitrobenzene-d5 (Surr)  
PHL6 = Phenol-d6 (Surr)  
TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
LCS 570-663011/2-A	Lab Control Sample	85	84	63	71	43	85
LCSD 570-663011/3-A	Lab Control Sample Dup	81	86	65	70	43	82
MB 570-663011/1-A	Method Blank	88	80	52	82	34	76

### Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)  
FBP = 2-Fluorobiphenyl (Surr)  
2FP = 2-Fluorophenol (Surr)  
NBZ = Nitrobenzene-d5 (Surr)  
PHL6 = Phenol-d6 (Surr)  
TPHd14 = p-Terphenyl-d14 (Surr)

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)
		BFB1 (38-134)
380-185035-1	HALAWA WELLS UNITS 1 & 2 F	86
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	91

# Surrogate Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

## Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (	91
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	91

**Surrogate Legend**

BFB = 4-Bromofluorobenzene (Surr)

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-185035-3	TB: HALAWA WELLS UNITS 1&	94
380-185035-4	TB: AIEA WELLS PUMPS 1&2 (260) (331-203-TP400)	93
LCS 570-666187/4	Lab Control Sample	83
LCSD 570-666187/5	Lab Control Sample Dup	85
MB 570-666187/6	Method Blank	87
MRL 570-666187/3	Lab Control Sample	81

**Surrogate Legend**

BFB = 4-Bromofluorobenzene (Surr)

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Drinking Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-185035-1	HALAWA WELLS UNITS 1 & 2 F	113
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	112
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	113
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	111

**Surrogate Legend**

OTCSN = n-Octacosane (Surr)

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

### Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
LCS 570-663324/2-A	Lab Control Sample	109
LCSD 570-663324/3-A	Lab Control Sample Dup	106
MB 570-663324/1-A	Method Blank	99
MRL 570-663324/4-A	Lab Control Sample	117

**Surrogate Legend**

OTCSN = n-Octacosane (Surr)

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

**Lab Sample ID: MB 380-189440/20-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2,4'-DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2,4-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2,6-Dinitrotoluene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
2-Methylnaphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
4,4'-DDD	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
4,4'-DDE	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
4,4'-DDT	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Acenaphthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Acenaphthylene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Acetochlor	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Alachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
alpha-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
alpha-Chlordane	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Anthracene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 11:42	1
Atrazine	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Benz(a)anthracene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Benzo[a]pyrene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 11:42	1
Benzo[b]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 11:42	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Benzo[k]fluoranthene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 11:42	1
beta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Bis(2-ethylhexyl) phthalate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 11:42	1
Bromacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Butachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Butylbenzylphthalate	<0.49		0.49	ug/L		12/02/25 08:58	12/03/25 11:42	1
Chlorobenzilate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Chloroneb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Chlorothalonil (Draconil, Bravo)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Chlorpyrifos	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Chrysene	<0.019		0.019	ug/L		12/02/25 08:58	12/03/25 11:42	1
delta-BHC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Di(2-ethylhexyl)adipate	<0.58		0.58	ug/L		12/02/25 08:58	12/03/25 11:42	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Dieldrin	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Diethylphthalate	<0.49		0.49	ug/L		12/02/25 08:58	12/03/25 11:42	1
Dimethylphthalate	<0.49		0.49	ug/L		12/02/25 08:58	12/03/25 11:42	1
Di-n-butyl phthalate	<0.97		0.97	ug/L		12/02/25 08:58	12/03/25 11:42	1
Di-n-octyl phthalate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Endosulfan I (Alpha)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Endosulfan II (Beta)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Endosulfan sulfate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Endrin	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Endrin aldehyde	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MB 380-189440/20-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
EPTC	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Fluoranthene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Fluorene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
gamma-Chlordane	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Heptachlor	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Heptachlor epoxide (isomer B)	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Hexachlorobenzene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Isophorone	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Lindane	<0.0097		0.0097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Malathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Methoxychlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Metolachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Molinate	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Naphthalene	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Parathion	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Pendimethalin (Penoxaline)	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Phenanthrene	<0.039		0.039	ug/L		12/02/25 08:58	12/03/25 11:42	1
Propachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Pyrene	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Simazine	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Terbacil	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Terbutylazine	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Thiobencarb	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1
Total Permethrin (mixed isomers)	<0.19		0.19	ug/L		12/02/25 08:58	12/03/25 11:42	1
trans-Nonachlor	<0.049		0.049	ug/L		12/02/25 08:58	12/03/25 11:42	1
Trifluralin	<0.097		0.097	ug/L		12/02/25 08:58	12/03/25 11:42	1

<i>Tentatively Identified Compound</i>	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Unknown	0.533	T J	ug/L		3.83	N/A	12/02/25 08:58	12/03/25 11:42	1
9-Octadecenamide, (Z)-	5.08	T J N	ug/L		7.79	301-02-0	12/02/25 08:58	12/03/25 11:42	1
Octadecanamide	0.583	T J N	ug/L		7.89	124-26-5	12/02/25 08:58	12/03/25 11:42	1
Unknown	0.851	T J	ug/L		14.75	N/A	12/02/25 08:58	12/03/25 11:42	1
Unknown	0.499	T J	ug/L		15.61	N/A	12/02/25 08:58	12/03/25 11:42	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	95		70 - 130	12/02/25 08:58	12/03/25 11:42	1
Perylene-d12	89		70 - 130	12/02/25 08:58	12/03/25 11:42	1
Triphenylphosphate	104		70 - 130	12/02/25 08:58	12/03/25 11:42	1

**Lab Sample ID: LCS 380-189440/22-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	1.95	1.77		ug/L		91	70 - 130

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-189440/22-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2,4'-DDD	1.95	2.00		ug/L		103	70 - 130
2,4'-DDE	1.95	1.94		ug/L		100	70 - 130
2,4'-DDT	1.95	1.97		ug/L		101	70 - 130
2,4-Dinitrotoluene	1.95	1.93		ug/L		99	70 - 130
2,6-Dinitrotoluene	1.95	1.91		ug/L		98	70 - 130
2-Methylnaphthalene	1.95	1.79		ug/L		92	70 - 130
4,4'-DDD	1.95	2.02		ug/L		104	70 - 130
4,4'-DDE	1.95	1.93		ug/L		99	70 - 130
4,4'-DDT	1.95	1.82		ug/L		93	70 - 130
Acenaphthene	1.95	1.91		ug/L		98	70 - 130
Acenaphthylene	1.95	1.94		ug/L		100	70 - 130
Acetochlor	1.95	2.08		ug/L		107	70 - 130
Alachlor	1.95	2.06		ug/L		106	70 - 130
alpha-BHC	1.95	1.92		ug/L		99	70 - 130
alpha-Chlordane	1.95	2.13		ug/L		110	70 - 130
Anthracene	1.95	1.72		ug/L		88	70 - 130
Atrazine	1.95	2.14		ug/L		110	70 - 130
Benz(a)anthracene	1.95	1.80		ug/L		92	70 - 130
Benzo[a]pyrene	1.95	1.89		ug/L		97	70 - 130
Benzo[b]fluoranthene	1.95	1.90		ug/L		98	70 - 130
Benzo[g,h,i]perylene	1.95	1.88		ug/L		96	70 - 130
Benzo[k]fluoranthene	1.95	1.88		ug/L		97	70 - 130
beta-BHC	1.95	1.94		ug/L		99	70 - 130
Bis(2-ethylhexyl) phthalate	1.95	1.79		ug/L		92	70 - 130
Bromacil	1.95	1.87		ug/L		96	70 - 130
Butachlor	1.95	2.03		ug/L		104	70 - 130
Butylbenzylphthalate	1.95	2.20		ug/L		113	70 - 130
Chlorobenzilate	1.95	1.80		ug/L		93	70 - 130
Chloroneb	1.95	1.83		ug/L		94	70 - 130
Chlorothalonil (Draconil, Bravo)	1.95	2.25		ug/L		115	70 - 130
Chlorpyrifos	1.95	2.13		ug/L		109	70 - 130
Chrysene	1.95	1.84		ug/L		94	70 - 130
delta-BHC	1.95	1.87		ug/L		96	70 - 130
Di(2-ethylhexyl)adipate	1.95	2.00		ug/L		103	70 - 130
Dibenz(a,h)anthracene	1.95	1.84		ug/L		95	70 - 130
Diclorvos (DDVP)	1.95	1.88		ug/L		96	70 - 130
Dieldrin	1.95	1.96		ug/L		101	70 - 130
Diethylphthalate	1.95	7.68	*+	ug/L		394	70 - 130
Dimethylphthalate	1.95	2.01		ug/L		103	70 - 130
Di-n-butyl phthalate	3.90	4.17		ug/L		107	70 - 130
Di-n-octyl phthalate	1.95	1.80		ug/L		92	70 - 130
Endosulfan I (Alpha)	1.95	1.87		ug/L		96	70 - 130
Endosulfan II (Beta)	1.95	1.93		ug/L		99	70 - 130
Endosulfan sulfate	1.95	2.14		ug/L		110	70 - 130
Endrin	1.95	2.04		ug/L		105	70 - 130
Endrin aldehyde	1.95	1.77		ug/L		91	60 - 130
EPTC	1.95	2.02		ug/L		104	70 - 130
Fluoranthene	1.95	2.04		ug/L		105	70 - 130

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: LCS 380-189440/22-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Fluorene	1.95	1.91		ug/L		98	70 - 130
gamma-Chlordane	1.95	2.15		ug/L		110	70 - 130
Heptachlor	1.95	2.01		ug/L		103	70 - 130
Heptachlor epoxide (isomer B)	1.95	1.99		ug/L		102	70 - 130
Hexachlorobenzene	1.95	2.00		ug/L		103	70 - 130
Hexachlorocyclopentadiene	1.95	1.90		ug/L		97	70 - 130
Indeno[1,2,3-cd]pyrene	1.95	1.87		ug/L		96	70 - 130
Isophorone	1.95	1.98		ug/L		102	70 - 130
Lindane	1.95	1.89		ug/L		97	70 - 130
Malathion	1.95	1.89		ug/L		97	70 - 130
Methoxychlor	1.95	2.01		ug/L		103	70 - 130
Metolachlor	1.95	1.93		ug/L		99	70 - 130
Molinate	1.95	2.02		ug/L		104	70 - 130
Naphthalene	1.95	1.84		ug/L		94	70 - 130
Parathion	1.95	2.27		ug/L		116	70 - 130
Pendimethalin (Penoxaline)	1.95	2.03		ug/L		104	70 - 130
Phenanthrene	1.95	1.89		ug/L		97	70 - 130
Propachlor	1.95	2.10		ug/L		108	70 - 130
Pyrene	1.95	2.00		ug/L		103	70 - 130
Simazine	1.95	1.92		ug/L		99	70 - 130
Terbacil	1.95	2.08		ug/L		107	70 - 130
Terbutylazine	1.95	2.19		ug/L		113	70 - 130
Thiobencarb	1.95	2.16		ug/L		111	70 - 130
trans-Nonachlor	1.95	2.00		ug/L		103	70 - 130
Trifluralin	1.95	2.03		ug/L		104	70 - 130

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	97		70 - 130
Perylene-d12	95		70 - 130
Triphenylphosphate	108		70 - 130

**Lab Sample ID: MRL 380-189440/21-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0972	0.122		ug/L		126	50 - 150
2,4'-DDD	0.0972	0.0879	J	ug/L		90	50 - 150
2,4'-DDE	0.0972	0.103		ug/L		106	50 - 150
2,4'-DDT	0.0972	0.0939	J	ug/L		97	50 - 150
2,4-Dinitrotoluene	0.0972	0.0956	J	ug/L		98	50 - 150
2,6-Dinitrotoluene	0.0972	0.112		ug/L		115	50 - 150
2-Methylnaphthalene	0.0972	0.113		ug/L		116	50 - 150
4,4'-DDD	0.0972	0.0970		ug/L		100	50 - 150
4,4'-DDE	0.0972	0.0979		ug/L		101	50 - 150
4,4'-DDT	0.0972	0.112		ug/L		116	50 - 150
Acenaphthene	0.0972	0.0941	J	ug/L		97	50 - 150

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-189440/21-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Acenaphthylene	0.0972	0.0997		ug/L		103	50 - 150
Acetochlor	0.0972	0.119		ug/L		122	50 - 150
Alachlor	0.0486	0.0582		ug/L		120	50 - 150
alpha-BHC	0.0972	0.103		ug/L		106	50 - 150
alpha-Chlordane	0.0243	<0.028		ug/L		103	50 - 150
Anthracene	0.0194	0.0229		ug/L		118	50 - 150
Atrazine	0.0486	0.0488	J	ug/L		100	50 - 150
Benz(a)anthracene	0.0486	0.0538		ug/L		111	50 - 150
Benzo[a]pyrene	0.0194	0.0216		ug/L		111	50 - 150
Benzo[b]fluoranthene	0.0194	0.0215		ug/L		111	50 - 150
Benzo[g,h,i]perylene	0.0486	0.0498		ug/L		102	50 - 150
Benzo[k]fluoranthene	0.0194	0.0224		ug/L		115	50 - 150
beta-BHC	0.0972	0.112		ug/L		115	50 - 150
Bis(2-ethylhexyl) phthalate	0.583	0.618		ug/L		106	50 - 150
Bromacil	0.0972	0.102		ug/L		105	50 - 150
Butachlor	0.0486	0.0632		ug/L		130	50 - 150
Butylbenzylphthalate	0.486	0.639		ug/L		132	50 - 150
Chlorobenzilate	0.0972	0.0946	J	ug/L		97	50 - 150
Chloroneb	0.0972	0.105		ug/L		108	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0972	0.0952	J	ug/L		98	50 - 150
Chlorpyrifos	0.0486	0.0510		ug/L		105	50 - 150
Chrysene	0.0194	0.0182	J	ug/L		94	50 - 150
delta-BHC	0.0972	0.0984		ug/L		101	50 - 150
Di(2-ethylhexyl)adipate	0.583	0.743		ug/L		127	50 - 150
Dibenz(a,h)anthracene	0.0486	0.0479	J	ug/L		99	50 - 150
Diclorvos (DDVP)	0.0486	0.0558		ug/L		115	50 - 150
Dieldrin	0.00972	0.00849	J	ug/L		87	50 - 150
Diethylphthalate	0.486	0.569		ug/L		117	50 - 150
Dimethylphthalate	0.486	0.534		ug/L		110	50 - 150
Di-n-butyl phthalate	0.486	0.532	J	ug/L		110	49 - 243
Di-n-octyl phthalate	0.0972	0.0954	J	ug/L		98	50 - 150
Endosulfan I (Alpha)	0.0972	0.105		ug/L		108	50 - 150
Endosulfan II (Beta)	0.0972	0.112		ug/L		116	50 - 150
Endosulfan sulfate	0.0972	0.105		ug/L		108	50 - 150
Endrin	0.00972	0.00964	J	ug/L		99	50 - 150
Endrin aldehyde	0.0972	0.111		ug/L		114	50 - 150
EPTC	0.0972	0.106		ug/L		109	50 - 150
Fluoranthene	0.0972	0.101		ug/L		104	50 - 150
Fluorene	0.0486	0.0530		ug/L		109	50 - 150
gamma-Chlordane	0.0243	0.0236	J	ug/L		97	50 - 150
Heptachlor	0.00972	0.0109		ug/L		112	50 - 150
Heptachlor epoxide (isomer B)	0.00972	0.0121		ug/L		125	50 - 150
Hexachlorobenzene	0.0486	0.0551		ug/L		113	50 - 150
Hexachlorocyclopentadiene	0.0486	0.0464	J	ug/L		96	50 - 150
Indeno[1,2,3-cd]pyrene	0.0486	0.0514		ug/L		106	50 - 150
Isophorone	0.0972	0.120		ug/L		124	50 - 150
Lindane	0.00972	0.00865	J	ug/L		89	50 - 150
Malathion	0.0972	0.0984		ug/L		101	50 - 150

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: MRL 380-189440/21-A**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Methoxychlor	0.0486	0.0522		ug/L		107	50 - 150
Metolachlor	0.0486	0.0616		ug/L		127	50 - 150
Molinate	0.0972	0.107		ug/L		110	50 - 150
Naphthalene	0.0972	0.103		ug/L		106	50 - 150
Parathion	0.0972	0.0924	J	ug/L		95	50 - 150
Pendimethalin (Penoxaline)	0.0972	0.0933	J	ug/L		96	50 - 150
Phenanthrene	0.0389	0.0394		ug/L		101	50 - 150
Propachlor	0.0486	0.0568		ug/L		117	50 - 150
Pyrene	0.0486	0.0512		ug/L		105	50 - 150
Simazine	0.0486	0.0419	J	ug/L		86	50 - 150
Terbacil	0.0972	0.110		ug/L		113	50 - 150
Terbutylazine	0.0972	0.104		ug/L		107	50 - 150
Thiobencarb	0.0972	0.108		ug/L		111	50 - 150
trans-Nonachlor	0.0243	<0.025		ug/L		102	50 - 150
Trifluralin	0.0972	0.108		ug/L		111	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	97		70 - 130
Perylene-d12	94		70 - 130
Triphenylphosphate	106		70 - 130

**Lab Sample ID: 380-185034-I-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.098		1.96	1.79		ug/L		91	70 - 130
2,4'-DDD	<0.098		1.96	1.93		ug/L		99	70 - 130
2,4'-DDE	<0.098		1.96	1.84		ug/L		94	70 - 130
2,4'-DDT	<0.098		1.96	1.87		ug/L		96	70 - 130
2,4-Dinitrotoluene	<0.098		1.96	2.04		ug/L		105	70 - 130
2,6-Dinitrotoluene	<0.098		1.96	2.04		ug/L		104	70 - 130
2-Methylnaphthalene	<0.098		1.96	1.83		ug/L		93	70 - 130
4,4'-DDD	<0.098		1.96	1.98		ug/L		101	70 - 130
4,4'-DDE	<0.098		1.96	1.83		ug/L		93	70 - 130
4,4'-DDT	<0.098		1.96	1.71		ug/L		87	70 - 130
Acenaphthene	<0.098		1.96	1.91		ug/L		98	70 - 130
Acenaphthylene	<0.098		1.96	1.95		ug/L		100	70 - 130
Acetochlor	<0.098		1.96	2.09		ug/L		107	70 - 130
Alachlor	<0.049		1.96	2.06		ug/L		106	70 - 130
alpha-BHC	<0.098		1.96	1.92		ug/L		98	70 - 130
alpha-Chlordane	<0.049		1.96	2.07		ug/L		106	70 - 130
Anthracene	<0.020		1.96	1.42		ug/L		73	70 - 130
Atrazine	<0.049		1.96	2.07		ug/L		106	70 - 130
Benz(a)anthracene	<0.049		1.96	1.74		ug/L		89	70 - 130
Benzo[a]pyrene	<0.020		1.96	1.78		ug/L		91	70 - 130
Benzo[b]fluoranthene	<0.020		1.96	1.85		ug/L		95	70 - 130

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-185034-I-1-A MS**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Matrix Spike**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Benzo[g,h,i]perylene	<0.049		1.96	1.96		ug/L		100	70 - 130
Benzo[k]fluoranthene	<0.020		1.96	1.89		ug/L		97	70 - 130
beta-BHC	<0.098		1.96	1.92		ug/L		98	70 - 130
Bis(2-ethylhexyl) phthalate	<0.59		1.96	1.83		ug/L		94	70 - 130
Bromacil	<0.098		1.96	2.06		ug/L		105	70 - 130
Butachlor	<0.049		1.96	2.09		ug/L		107	70 - 130
Butylbenzylphthalate	<0.49		1.96	2.25		ug/L		115	70 - 130
Chlorobenzilate	<0.098		1.96	1.85		ug/L		95	70 - 130
Chloroneb	<0.098		1.96	1.86		ug/L		95	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.098		1.96	2.37		ug/L		121	70 - 130
Chlorpyrifos	<0.049		1.96	2.14		ug/L		109	70 - 130
Chrysene	<0.020		1.96	1.88		ug/L		96	70 - 130
delta-BHC	<0.098		1.96	1.89		ug/L		96	70 - 130
Di(2-ethylhexyl)adipate	<0.59		1.96	1.92		ug/L		98	70 - 130
Dibenz(a,h)anthracene	<0.049		1.96	1.90		ug/L		97	70 - 130
Diclorvos (DDVP)	<0.049		1.96	1.94		ug/L		99	70 - 130
Dieldrin	<0.0098		1.96	1.99		ug/L		102	70 - 130
Diethylphthalate	<0.49	*+	1.96	2.10		ug/L		107	70 - 130
Dimethylphthalate	<0.49		1.96	2.05		ug/L		105	70 - 130
Di-n-butyl phthalate	<0.98		3.91	4.19		ug/L		107	70 - 130
Di-n-octyl phthalate	<0.098		1.96	1.82		ug/L		93	70 - 130
Endosulfan I (Alpha)	<0.098		1.96	1.84		ug/L		94	70 - 130
Endosulfan II (Beta)	<0.098		1.96	1.94		ug/L		99	70 - 130
Endosulfan sulfate	<0.098		1.96	2.14		ug/L		110	70 - 130
Endrin	<0.0098		1.96	2.09		ug/L		107	70 - 130
Endrin aldehyde	<0.098		1.96	1.51		ug/L		77	60 - 130
EPTC	<0.098		1.96	2.04		ug/L		104	70 - 130
Fluoranthene	<0.098		1.96	2.07		ug/L		106	70 - 130
Fluorene	<0.049		1.96	1.93		ug/L		99	70 - 130
gamma-Chlordane	<0.049		1.96	2.07		ug/L		106	70 - 130
Heptachlor	<0.0098		1.96	2.00		ug/L		102	70 - 130
Heptachlor epoxide (isomer B)	<0.0098		1.96	2.00		ug/L		102	70 - 130
Hexachlorobenzene	<0.049		1.96	2.02		ug/L		104	70 - 130
Hexachlorocyclopentadiene	<0.049		1.96	1.99		ug/L		102	70 - 130
Indeno[1,2,3-cd]pyrene	<0.049		1.96	1.96		ug/L		100	70 - 130
Isophorone	<0.098		1.96	2.02		ug/L		103	70 - 130
Lindane	<0.0098		1.96	1.90		ug/L		97	70 - 130
Malathion	<0.098		1.96	1.94		ug/L		99	70 - 130
Methoxychlor	<0.049		1.96	2.10		ug/L		107	70 - 130
Metolachlor	<0.049		1.96	1.95		ug/L		100	70 - 130
Molinate	<0.098		1.96	2.05		ug/L		105	70 - 130
Naphthalene	<0.098		1.96	1.86		ug/L		95	70 - 130
Parathion	<0.098		1.96	2.27		ug/L		116	70 - 130
Pendimethalin (Penoxaline)	<0.098		1.96	2.19		ug/L		112	70 - 130
Phenanthrene	<0.039		1.96	1.92		ug/L		98	70 - 130
Propachlor	<0.049		1.96	2.14		ug/L		110	70 - 130
Pyrene	<0.049		1.96	2.04		ug/L		104	70 - 130
Simazine	<0.049		1.96	1.98		ug/L		101	70 - 130

Eurofins Eaton Analytical Pomona



# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

**Lab Sample ID: 380-185034-J-1-A MSD**  
**Matrix: Water**  
**Analysis Batch: 189744**

**Client Sample ID: Matrix Spike Duplicate**  
**Prep Type: Total/NA**  
**Prep Batch: 189440**

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec		RPD	Limit
	Result	Qualifier	Added	Result	Qualifier				Limits	RPD		
Chlorpyrifos	<0.049		1.95	2.13		ug/L		109	70 - 130	1	20	
Chrysene	<0.020		1.95	1.85		ug/L		95	70 - 130	2	20	
delta-BHC	<0.098		1.95	1.91		ug/L		98	70 - 130	2	20	
Di(2-ethylhexyl)adipate	<0.59		1.95	1.91		ug/L		98	70 - 130	1	20	
Dibenz(a,h)anthracene	<0.049		1.95	1.93		ug/L		99	70 - 130	2	20	
Diclorvos (DDVP)	<0.049		1.95	1.93		ug/L		99	70 - 130	0	20	
Dieldrin	<0.0098		1.95	2.04		ug/L		105	70 - 130	2	20	
Diethylphthalate	<0.49	*+	1.95	2.14		ug/L		110	70 - 130	2	20	
Dimethylphthalate	<0.49		1.95	2.07		ug/L		106	70 - 130	1	20	
Di-n-butyl phthalate	<0.98		3.89	4.25		ug/L		109	70 - 130	1	20	
Di-n-octyl phthalate	<0.098		1.95	1.79		ug/L		92	70 - 130	2	20	
Endosulfan I (Alpha)	<0.098		1.95	1.90		ug/L		98	70 - 130	3	20	
Endosulfan II (Beta)	<0.098		1.95	1.96		ug/L		101	70 - 130	1	20	
Endosulfan sulfate	<0.098		1.95	2.23		ug/L		115	70 - 130	4	20	
Endrin	<0.0098		1.95	2.15		ug/L		111	70 - 130	3	20	
Endrin aldehyde	<0.098		1.95	1.74		ug/L		90	60 - 130	15	20	
EPTC	<0.098		1.95	2.07		ug/L		107	70 - 130	2	20	
Fluoranthene	<0.098		1.95	2.08		ug/L		107	70 - 130	0	20	
Fluorene	<0.049		1.95	1.90		ug/L		98	70 - 130	2	20	
gamma-Chlordane	<0.049		1.95	2.13		ug/L		109	70 - 130	3	20	
Heptachlor	<0.0098		1.95	1.99		ug/L		102	70 - 130	1	20	
Heptachlor epoxide (isomer B)	<0.0098		1.95	2.07		ug/L		106	70 - 130	3	20	
Hexachlorobenzene	<0.049		1.95	2.00		ug/L		103	70 - 130	1	20	
Hexachlorocyclopentadiene	<0.049		1.95	1.92		ug/L		99	70 - 130	4	20	
Indeno[1,2,3-cd]pyrene	<0.049		1.95	2.02		ug/L		104	70 - 130	3	20	
Isophorone	<0.098		1.95	2.04		ug/L		105	70 - 130	1	20	
Lindane	<0.0098		1.95	1.96		ug/L		101	70 - 130	3	20	
Malathion	<0.098		1.95	1.99		ug/L		102	70 - 130	3	20	
Methoxychlor	<0.049		1.95	2.07		ug/L		107	70 - 130	1	20	
Metolachlor	<0.049		1.95	2.01		ug/L		103	70 - 130	3	20	
Molinate	<0.098		1.95	2.04		ug/L		105	70 - 130	1	20	
Naphthalene	<0.098		1.95	1.86		ug/L		95	70 - 130	0	20	
Parathion	<0.098		1.95	2.36		ug/L		121	70 - 130	4	20	
Pendimethalin (Penoxaline)	<0.098		1.95	2.21		ug/L		114	70 - 130	1	20	
Phenanthrene	<0.039		1.95	1.90		ug/L		98	70 - 130	1	20	
Propachlor	<0.049		1.95	2.16		ug/L		111	70 - 130	1	20	
Pyrene	<0.049		1.95	2.00		ug/L		103	70 - 130	2	20	
Simazine	<0.049		1.95	2.01		ug/L		103	70 - 130	2	20	
Terbacil	<0.098		1.95	2.22		ug/L		114	70 - 130	3	20	
Terbutylazine	<0.098		1.95	2.21		ug/L		113	70 - 130	2	20	
Thiobencarb	<0.098		1.95	2.19		ug/L		112	70 - 130	1	20	
trans-Nonachlor	<0.049		1.95	1.94		ug/L		100	70 - 130	3	20	
Trifluralin	<0.098		1.95	2.15		ug/L		111	70 - 130	3	20	

Surrogate	MSD %Recovery	MSD Qualifier	Limits
2-Nitro-m-xylene	98		70 - 130
Perylene-d12	98		70 - 130

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-185034-J-1-A MSD  
Matrix: Water  
Analysis Batch: 189744

Client Sample ID: Matrix Spike Duplicate  
Prep Type: Total/NA  
Prep Batch: 189440

Surrogate	MSD %Recovery	MSD Qualifier	Limits
Triphenylphosphate	112		70 - 130

## Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-663011/1-A  
Matrix: Water  
Analysis Batch: 668191

Client Sample ID: Method Blank  
Prep Type: Total/NA  
Prep Batch: 663011

Tentatively Identified Compound	Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	12/01/25 08:21	12/11/25 08:33	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	75		33 - 139	12/01/25 08:21	12/11/25 08:33	1
2-Fluorobiphenyl (Surr)	84		33 - 126	12/01/25 08:21	12/11/25 08:33	1
2-Fluorophenol (Surr)	53		12 - 120	12/01/25 08:21	12/11/25 08:33	1
Nitrobenzene-d5 (Surr)	89		36 - 120	12/01/25 08:21	12/11/25 08:33	1
Phenol-d6 (Surr)	33		10 - 120	12/01/25 08:21	12/11/25 08:33	1
p-Terphenyl-d14 (Surr)	83		47 - 131	12/01/25 08:21	12/11/25 08:33	1

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-663011/1-A  
Matrix: Water  
Analysis Batch: 668356

Client Sample ID: Method Blank  
Prep Type: Total/NA  
Prep Batch: 663011

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
2-Methylnaphthalene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Acenaphthene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Acenaphthylene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Anthracene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Benzo[a]anthracene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Benzo[a]pyrene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Chrysene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Fluoranthene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Fluorene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Naphthalene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Phenanthrene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1
Pyrene	<0.20		0.20	ug/L		12/01/25 08:21	12/11/25 08:13	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	88		28 - 127	12/01/25 08:21	12/11/25 08:13	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: MB 570-663011/1-A**  
**Matrix: Water**  
**Analysis Batch: 668356**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Surrogate	MB MB	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Fluorobiphenyl (Surr)		80		31 - 120	12/01/25 08:21	12/11/25 08:13	1
2-Fluorophenol (Surr)		52		17 - 120	12/01/25 08:21	12/11/25 08:13	1
Nitrobenzene-d5 (Surr)		82		27 - 120	12/01/25 08:21	12/11/25 08:13	1
Phenol-d6 (Surr)		34		10 - 120	12/01/25 08:21	12/11/25 08:13	1
p-Terphenyl-d14 (Surr)		76		45 - 120	12/01/25 08:21	12/11/25 08:13	1

**Lab Sample ID: LCS 570-663011/2-A**  
**Matrix: Water**  
**Analysis Batch: 668356**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	20.0	13.4		ug/L		67	47 - 120
2-Methylnaphthalene	20.0	13.5		ug/L		68	43 - 120
Acenaphthene	20.0	16.4		ug/L		82	60 - 132
Acenaphthylene	20.0	15.7		ug/L		79	54 - 126
Anthracene	20.0	16.5		ug/L		82	43 - 120
Benzo[a]anthracene	20.0	16.7		ug/L		84	42 - 133
Benzo[a]pyrene	20.0	17.6		ug/L		88	32 - 148
Benzo[b]fluoranthene	20.0	17.8		ug/L		89	42 - 140
Benzo[g,h,i]perylene	20.0	16.6		ug/L		83	1 - 195
Benzo[k]fluoranthene	20.0	16.6		ug/L		83	25 - 146
Chrysene	20.0	16.5		ug/L		83	44 - 140
Dibenz(a,h)anthracene	20.0	17.9		ug/L		89	1 - 200
Fluoranthene	20.0	17.1		ug/L		86	43 - 121
Fluorene	20.0	16.6		ug/L		83	70 - 120
Indeno[1,2,3-cd]pyrene	20.0	17.2		ug/L		86	1 - 151
Naphthalene	20.0	13.2		ug/L		66	36 - 120
Phenanthrene	20.0	16.9		ug/L		84	65 - 120
Pyrene	20.0	17.7		ug/L		89	70 - 120

Surrogate	LCS LCS	%Recovery	Qualifier	Limits
2,4,6-Tribromophenol (Surr)		85		28 - 127
2-Fluorobiphenyl (Surr)		84		31 - 120
2-Fluorophenol (Surr)		63		17 - 120
Nitrobenzene-d5 (Surr)		71		27 - 120
Phenol-d6 (Surr)		43		10 - 120
p-Terphenyl-d14 (Surr)		85		45 - 120

**Lab Sample ID: LCSD 570-663011/3-A**  
**Matrix: Water**  
**Analysis Batch: 668356**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	20.0	13.6		ug/L		68	47 - 120	1	20
2-Methylnaphthalene	20.0	13.6		ug/L		68	43 - 120	1	20
Acenaphthene	20.0	16.5		ug/L		82	60 - 132	0	29
Acenaphthylene	20.0	15.8		ug/L		79	54 - 126	0	45

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: LCSD 570-663011/3-A**  
**Matrix: Water**  
**Analysis Batch: 668356**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Anthracene	20.0	16.2		ug/L		81	43 - 120	2	40
Benzo[a]anthracene	20.0	16.5		ug/L		82	42 - 133	2	32
Benzo[a]pyrene	20.0	17.2		ug/L		86	32 - 148	2	43
Benzo[b]fluoranthene	20.0	17.6		ug/L		88	42 - 140	1	43
Benzo[g,h,i]perylene	20.0	16.5		ug/L		83	1 - 195	1	61
Benzo[k]fluoranthene	20.0	16.4		ug/L		82	25 - 146	1	38
Chrysene	20.0	16.5		ug/L		82	44 - 140	1	53
Dibenz(a,h)anthracene	20.0	17.5		ug/L		87	1 - 200	2	75
Fluoranthene	20.0	16.7		ug/L		84	43 - 121	2	40
Fluorene	20.0	16.7		ug/L		84	70 - 120	1	23
Indeno[1,2,3-cd]pyrene	20.0	17.0		ug/L		85	1 - 151	1	60
Naphthalene	20.0	12.9		ug/L		65	36 - 120	2	39
Phenanthrene	20.0	16.5		ug/L		83	65 - 120	2	24
Pyrene	20.0	17.5		ug/L		88	70 - 120	1	30

Surrogate	LCSD		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	81		28 - 127
2-Fluorobiphenyl (Surr)	86		31 - 120
2-Fluorophenol (Surr)	65		17 - 120
Nitrobenzene-d5 (Surr)	70		27 - 120
Phenol-d6 (Surr)	43		10 - 120
p-Terphenyl-d14 (Surr)	82		45 - 120

**Lab Sample ID: 380-185035-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 668356**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.19		19.2	13.7		ug/L		72	36 - 120
2-Methylnaphthalene	<0.19		19.2	13.8		ug/L		72	32 - 124
Acenaphthene	<0.19		19.2	16.4		ug/L		86	47 - 145
Acenaphthylene	<0.19		19.2	15.8		ug/L		83	33 - 145
Anthracene	<0.19		19.2	15.7		ug/L		82	27 - 133
Benzo[a]anthracene	<0.19		19.2	16.6		ug/L		87	33 - 143
Benzo[a]pyrene	<0.19		19.2	16.8		ug/L		88	17 - 163
Benzo[b]fluoranthene	<0.19		19.2	17.1		ug/L		89	24 - 159
Benzo[g,h,i]perylene	<0.19		19.2	16.3		ug/L		85	1 - 219
Benzo[k]fluoranthene	<0.19		19.2	16.6		ug/L		87	11 - 162
Chrysene	<0.19		19.2	16.4		ug/L		86	17 - 168
Dibenz(a,h)anthracene	<0.19		19.2	17.2		ug/L		90	1 - 227
Fluoranthene	<0.19		19.2	16.7		ug/L		87	26 - 137
Fluorene	<0.19		19.2	16.4		ug/L		86	59 - 121
Indeno[1,2,3-cd]pyrene	<0.19		19.2	16.9		ug/L		88	1 - 171
Naphthalene	<0.19		19.2	13.2		ug/L		69	21 - 133
Phenanthrene	<0.19		19.2	16.4		ug/L		86	54 - 120
Pyrene	<0.19		19.2	17.6		ug/L		92	52 - 120

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

**Lab Sample ID: 380-185035-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 668356**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Surrogate	MS %Recovery	MS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	83		28 - 127
2-Fluorobiphenyl (Surr)	86		31 - 120
2-Fluorophenol (Surr)	61		17 - 120
Nitrobenzene d5 (Surr)	71		27 - 120
Phenol-d6 (Surr)	40		10 - 120
p-Terphenyl-d14 (Surr)	84		45 - 120

**Lab Sample ID: 380-185035-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 668356**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**  
**Prep Batch: 663011**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec		RPD	Limit
									Limits	RPD		
1-Methylnaphthalene	<0.19		19.3	12.8		ug/L		66	36 - 120	7	30	
2-Methylnaphthalene	<0.19		19.3	13.0		ug/L		67	32 - 124	6	30	
Acenaphthene	<0.19		19.3	15.9		ug/L		82	47 - 145	3	48	
Acenaphthylene	<0.19		19.3	15.0		ug/L		78	33 - 145	5	74	
Anthracene	<0.19		19.3	16.0		ug/L		83	27 - 133	2	66	
Benzo[a]anthracene	<0.19		19.3	16.8		ug/L		87	33 - 143	1	53	
Benzo[a]pyrene	<0.19		19.3	17.1		ug/L		89	17 - 163	2	72	
Benzo[b]fluoranthene	<0.19		19.3	17.3		ug/L		90	24 - 159	1	71	
Benzo[g,h,i]perylene	<0.19		19.3	16.5		ug/L		85	1 - 219	1	97	
Benzo[k]fluoranthene	<0.19		19.3	16.5		ug/L		86	11 - 162	1	63	
Chrysene	<0.19		19.3	16.6		ug/L		86	17 - 168	1	87	
Dibenz(a,h)anthracene	<0.19		19.3	17.6		ug/L		91	1 - 227	2	126	
Fluoranthene	<0.19		19.3	16.8		ug/L		87	26 - 137	1	66	
Fluorene	<0.19		19.3	15.6		ug/L		81	59 - 121	5	38	
Indeno[1,2,3-cd]pyrene	<0.19		19.3	17.0		ug/L		88	1 - 171	1	99	
Naphthalene	<0.19		19.3	12.0		ug/L		62	21 - 133	9	65	
Phenanthrene	<0.19		19.3	16.3		ug/L		84	54 - 120	1	39	
Pyrene	<0.19		19.3	17.9		ug/L		93	52 - 120	2	49	

Surrogate	MSD %Recovery	MSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	81		28 - 127
2-Fluorobiphenyl (Surr)	79		31 - 120
2-Fluorophenol (Surr)	53		17 - 120
Nitrobenzene-d5 (Surr)	65		27 - 120
Phenol-d6 (Surr)	35		10 - 120
p-Terphenyl-d14 (Surr)	84		45 - 120

## Method: 8015B GRO LL - Gasoline Range Organics - (GC)

**Lab Sample ID: MB 570-666187/6**  
**Matrix: Water**  
**Analysis Batch: 666187**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			12/07/25 16:30	1

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

**Lab Sample ID: MB 570-666187/6**  
**Matrix: Water**  
**Analysis Batch: 666187**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**

Surrogate	MB MB %Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	87		38 - 134		12/07/25 16:30	1

**Lab Sample ID: LCS 570-666187/4**  
**Matrix: Water**  
**Analysis Batch: 666187**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	400	345		ug/L		86	78 - 120

Surrogate	LCS LCS %Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	83		38 - 134

**Lab Sample ID: LCSD 570-666187/5**  
**Matrix: Water**  
**Analysis Batch: 666187**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	400	377		ug/L		94	78 - 120	9	10

Surrogate	LCSD LCSD %Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	85		38 - 134

**Lab Sample ID: MRL 570-666187/3**  
**Matrix: Water**  
**Analysis Batch: 666187**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	10.0	12.3		ug/L		123	50 - 150

Surrogate	MRL MRL %Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	81		38 - 134

**Lab Sample ID: 380-185035-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 666187**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	<10		400	400		ug/L		100	68 - 122

Surrogate	MS MS %Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	91		38 - 134

# QC Sample Results

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

**Lab Sample ID: 380-185035-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 666187**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	<10		400	396		ug/L		99	68 - 122	1	18
<b>Surrogate</b>	<b>%Recovery</b>	<b>MSD Qualifier</b>	<b>MSD Limits</b>								
4-Bromofluorobenzene (Surr)	91		38 - 134								

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

**Lab Sample ID: MB 570-663324/1-A**  
**Matrix: Water**  
**Analysis Batch: 666060**

**Client Sample ID: Method Blank**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		12/01/25 16:05	12/06/25 23:26	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		12/01/25 16:05	12/06/25 23:26	1
C8-C18	<25		25	ug/L		12/01/25 16:05	12/06/25 23:26	1
<b>Surrogate</b>	<b>%Recovery</b>	<b>MB Qualifier</b>	<b>MB Limits</b>	<b>Prepared</b>	<b>Analyzed</b>	<b>Dil Fac</b>		
n-Octacosane (Surr)	99		60 - 130	12/01/25 16:05	12/06/25 23:26	1		

**Lab Sample ID: LCS 570-663324/2-A**  
**Matrix: Water**  
**Analysis Batch: 666060**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	1600	1660		ug/L		104	56 - 127		
<b>Surrogate</b>	<b>%Recovery</b>	<b>LCS Qualifier</b>	<b>LCS Limits</b>						
n-Octacosane (Surr)	109		60 - 130						

**Lab Sample ID: LCSD 570-663324/3-A**  
**Matrix: Water**  
**Analysis Batch: 666060**

**Client Sample ID: Lab Control Sample Dup**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	1600	1610		ug/L		101	56 - 127	3	23
<b>Surrogate</b>	<b>%Recovery</b>	<b>LCSD Qualifier</b>	<b>LCSD Limits</b>						
n-Octacosane (Surr)	106		60 - 130						

**Lab Sample ID: MRL 570-663324/4-A**  
**Matrix: Water**  
**Analysis Batch: 666060**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	0.0200	0.0303	^3+	mg/L		152	50 - 150		

Eurofins Eaton Analytical Pomona

# QC Sample Results

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

## Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

**Lab Sample ID: MRL 570-663324/4-A**  
**Matrix: Water**  
**Analysis Batch: 666060**

**Client Sample ID: Lab Control Sample**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

<i>Surrogate</i>	<i>MRL</i> %Recovery	<i>MRL</i> Qualifier	<i>Limits</i>
<i>n-Octacosane (Surr)</i>	117		60 - 130

**Lab Sample ID: 380-185035-1 MS**  
**Matrix: Drinking Water**  
**Analysis Batch: 666060**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

<i>Analyte</i>	<i>Sample</i> Result	<i>Sample</i> Qualifier	<i>Spike</i> Added	<i>MS</i> Result	<i>MS</i> Qualifier	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i> Limits
C10-C28	<26	^3+	1650	1780		ug/L		108	70 - 130
<i>Surrogate</i>	<i>MS</i> %Recovery	<i>MS</i> Qualifier	<i>Limits</i>						
<i>n-Octacosane (Surr)</i>	112		60 - 130						

**Lab Sample ID: 380-185035-1 MSD**  
**Matrix: Drinking Water**  
**Analysis Batch: 666060**

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 (331-206-TP065)**  
**Prep Type: Total/NA**  
**Prep Batch: 663324**

<i>Analyte</i>	<i>Sample</i> Result	<i>Sample</i> Qualifier	<i>Spike</i> Added	<i>MSD</i> Result	<i>MSD</i> Qualifier	<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec</i> Limits	<i>RPD</i> RPD	<i>RPD</i> Limit
C10-C28	<26	^3+	1690	1880		ug/L		111	70 - 130	5	20
<i>Surrogate</i>	<i>MSD</i> %Recovery	<i>MSD</i> Qualifier	<i>Limits</i>								
<i>n-Octacosane (Surr)</i>	113		60 - 130								

# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## GC/MS Semi VOA

### Prep Batch: 189440

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	525.2	
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	525.2	
MB 380-189440/20-A	Method Blank	Total/NA	Water	525.2	
LCS 380-189440/22-A	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-189440/21-A	Lab Control Sample	Total/NA	Water	525.2	
380-185034-I-1-A MS	Matrix Spike	Total/NA	Water	525.2	
380-185034-J-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	525.2	

### Analysis Batch: 189744

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	525.2	189440
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	525.2	189440
MB 380-189440/20-A	Method Blank	Total/NA	Water	525.2	189440
LCS 380-189440/22-A	Lab Control Sample	Total/NA	Water	525.2	189440
MRL 380-189440/21-A	Lab Control Sample	Total/NA	Water	525.2	189440
380-185034-I-1-A MS	Matrix Spike	Total/NA	Water	525.2	189440
380-185034-J-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	525.2	189440

### Prep Batch: 663011

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	625.1	
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	625.1	
MB 570-663011/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-663011/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-663011/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	625.1	
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	625.1	

### Analysis Batch: 668191

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	625.1	663011
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	625.1	663011
MB 570-663011/1-A	Method Blank	Total/NA	Water	625.1	663011

### Analysis Batch: 668356

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	625.1 SIM	663011
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	625.1 SIM	663011
MB 570-663011/1-A	Method Blank	Total/NA	Water	625.1 SIM	663011
LCS 570-663011/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	663011
LCSD 570-663011/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	663011
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	625.1 SIM	663011
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	625.1 SIM	663011

## GC VOA

### Analysis Batch: 666187

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP01)	Total/NA	Drinking Water	8015B GRO LL	
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP01)	Total/NA	Drinking Water	8015B GRO LL	

# QC Association Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## GC VOA (Continued)

### Analysis Batch: 666187 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-3	TB: HALAWA WELLS UNITS 1&2 (331-206-TP06	Total/NA	Water	8015B GRO LL	
380-185035-4	TB: AIEA WELLS PUMPS 1&2 (260) (331-203-TF	Total/NA	Water	8015B GRO LL	
MB 570-666187/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-666187/4	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-666187/5	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-666187/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	8015B GRO LL	
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	8015B GRO LL	

## GC Semi VOA

### Prep Batch: 663324

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP06	Total/NA	Drinking Water	3510C	
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP	Total/NA	Drinking Water	3510C	
MB 570-663324/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-663324/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-663324/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-663324/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	3510C	
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	3510C	

### Analysis Batch: 666060

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP06	Total/NA	Drinking Water	8015B	663324
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP	Total/NA	Drinking Water	8015B	663324
MB 570-663324/1-A	Method Blank	Total/NA	Water	8015B	663324
LCS 570-663324/2-A	Lab Control Sample	Total/NA	Water	8015B	663324
LCSD 570-663324/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	663324
MRL 570-663324/4-A	Lab Control Sample	Total/NA	Water	8015B	663324
380-185035-1 MS	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	8015B	663324
380-185035-1 MSD	HALAWA WELLS UNITS 1 & 2 (331-206-TP065)	Total/NA	Drinking Water	8015B	663324

# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Client Sample ID: HALAWA WELLS UNITS 1 & 2 P1  
(331-206-TP065)**

**Lab Sample ID: 380-185035-1**

**Date Collected: 11/24/25 10:58  
Date Received: 11/28/25 10:12**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			189440	KRD3	EA POM	12/02/25 08:58
Total/NA	Analysis	525.2		1	189744	Q8LA	EA POM	12/03/25 16:24
Total/NA	Prep	625.1			663011	OAJ3	EET CAL 4	12/01/25 08:21
Total/NA	Analysis	625.1		1	668191	J7WE	EET CAL 4	12/11/25 13:23
Total/NA	Prep	625.1			663011	OAJ3	EET CAL 4	12/01/25 08:21
Total/NA	Analysis	625.1 SIM		1	668356	PQS1	EET CAL 4	12/11/25 11:08
Total/NA	Analysis	8015B GRO LL		1	666187	YD9V	EET CAL 4	12/07/25 17:44
Total/NA	Prep	3510C			663324	TVD6	EET CAL 4	12/01/25 16:05
Total/NA	Analysis	8015B		1	666060	H6FE	EET CAL 4	12/07/25 02:38

**Client Sample ID: AIEA WELLS PUMPS 1&2 (260) P2  
(331-203-TP400)**

**Lab Sample ID: 380-185035-2**

**Date Collected: 11/24/25 12:44  
Date Received: 11/28/25 10:12**

**Matrix: Drinking Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			189440	KRD3	EA POM	12/02/25 08:58
Total/NA	Analysis	525.2		1	189744	Q8LA	EA POM	12/03/25 16:44
Total/NA	Prep	625.1			663011	OAJ3	EET CAL 4	12/01/25 08:21
Total/NA	Analysis	625.1		1	668191	J7WE	EET CAL 4	12/11/25 13:47
Total/NA	Prep	625.1			663011	OAJ3	EET CAL 4	12/01/25 08:21
Total/NA	Analysis	625.1 SIM		1	668356	PQS1	EET CAL 4	12/11/25 11:30
Total/NA	Analysis	8015B GRO LL		1	666187	YD9V	EET CAL 4	12/07/25 18:50
Total/NA	Prep	3510C			663324	TVD6	EET CAL 4	12/01/25 16:05
Total/NA	Analysis	8015B		1	666060	H6FE	EET CAL 4	12/07/25 05:51

**Client Sample ID: TB: HALAWA WELLS UNITS 1&2  
(331-206-TP065)**

**Lab Sample ID: 380-185035-3**

**Date Collected: 11/24/25 10:58  
Date Received: 11/28/25 10:12**

**Matrix: Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	666187	YD9V	EET CAL 4	12/08/25 01:02

**Client Sample ID: TB: AIEA WELLS PUMPS 1&2 (260)  
(331-203-TP400)**

**Lab Sample ID: 380-185035-4**

**Date Collected: 11/24/25 12:44  
Date Received: 11/28/25 10:12**

**Matrix: Water**

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	666187	YD9V	EET CAL 4	12/08/25 01:24

# Lab Chronicle

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

**Laboratory References:**

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100  
EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

# Accreditation/Certification Summary

Client: City & County of Honolulu  
 Project/Site: RED-HILL

Job ID: 380-185035-1  
 SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
 Wells Pumps 1&2 P2

## Laboratory: Eurofins Eaton Analytical Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	1-Methylnaphthalene
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	2-Methylnaphthalene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4' DDE
525.2	525.2	Drinking Water	4,4'-DDT
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Isophorone
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans Nonachlor

## Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-26
Arizona	State	AZ0830	11-17-26
California	Los Angeles County Sanitation Districts	9257304	07-31-26

Eurofins Eaton Analytical Pomona

# Accreditation/Certification Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

## Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	3082	07-31-26
Kansas	NELAP	E-10420	07-31-26
Nevada	State	CA00111	07-31-26
Oregon	NELAP	4175	02-02-26
USDA	US Federal Programs	525-23-159-97150	06-08-26
Utah	NELAP	CA00111	02-28-26
Washington	State	C916	10-11-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

# Method Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

#### Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

#### Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

# Sample Summary

Client: City & County of Honolulu  
Project/Site: RED-HILL

Job ID: 380-185035-1  
SDG: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea  
Wells Pumps 1&2 P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-185035-1	HALAWA WELLS UNITS 1 & 2 P1 (331-206-TP065)	Drinking Water	11/24/25 10:58	11/28/25 10:12	HI0000331
380-185035-2	AIEA WELLS PUMPS 1&2 (260) P2 (331-203-TP400)	Drinking Water	11/24/25 12:44	11/28/25 10:12	HI0000331
380-185035-3	TB: HALAWA WELLS UNITS 1&2 (331-206-TP065)	Water	11/24/25 10:58	11/28/25 10:12	
380-185035-4	TB: AIEA WELLS PUMPS 1&2 (260) (331-203-TP400)	Water	11/24/25 12:44	11/28/25 10:12	

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16



**Eurofins Eaton Analytical Pomona**

941 Corporate Center Drive  
 Pomona, CA 91768-2642  
 Phone (626) 386-1100

**Chain of Custody Record**

**eurofins** | Environment Testing

<b>Client Information</b>		Sampler bailey		Lab PM Lopez, Maria		Carrier Tracking No(s)		COC No: 380-28005-2757 1																																																																																																																								
Client Contact: Kirk Iwamoto		Phone: +1 808 748 5840		E-Mail Maria.Lopez@et.eurofinsus.com		State of Origin:		Page: Page 1 of 1																																																																																																																								
Company: City & County of Honolulu		PWSID:		<b>Analysis Requested</b>						Job #:																																																																																																																						
Address: 630 South Beretania Street Chemistry Lab		Due Date Requested:		<table border="1"> <tr> <th>Field Filled Sample (Yes or No)</th> <th>Perform MS/MSD (Yes or No)</th> <th>625.1, 625.1_SIM</th> <th>8015B_GRO_LL - (MOD) GRO</th> <th>8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18</th> <th>625.2_PREC - (MOD) 625plus Plus TICs</th> <th>537_DW_PREC - 537 1 Full List</th> <th>533 - All Analytes</th> </tr> <tr> <td>X</td> <td>X</td> <td>4</td> <td>5</td> <td>4</td> <td>2</td> <td></td> <td></td> </tr> </table>						Field Filled Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 625.1_SIM	8015B_GRO_LL - (MOD) GRO	8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18	625.2_PREC - (MOD) 625plus Plus TICs	537_DW_PREC - 537 1 Full List	533 - All Analytes	X	X	4	5	4	2			Preservation Codes: R - NaThioSO4 RA - NaThio/HCl Q - Na2SO3 QA - Na2SO3/HCl Y - Trizma I - NH4 Acetate																																																																																																						
Field Filled Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 625.1_SIM	8015B_GRO_LL - (MOD) GRO							8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18	625.2_PREC - (MOD) 625plus Plus TICs	537_DW_PREC - 537 1 Full List	533 - All Analytes																																																																																																																			
X	X	4	5							4	2																																																																																																																					
City: Honolulu		TAT Requested (days):								 <p>380-185035 COC</p>		Other:																																																																																																																				
State, Zip: HI, 96843		Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No																																																																																																																														
Phone: 808-748-5840 (Tel)		PO #: C20525101 exp 05312023																																																																																																																														
Email: kiwamoto@hbws.org		WO #:		Total Number of Containers						Special Instructions/Note:																																																																																																																						
Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111		<table border="1"> <thead> <tr> <th>Sample Identification</th> <th>Sample Date</th> <th>Sample Time</th> <th>Sample Type (C=comp, G=grab)</th> <th>Matrix (Water, Soil, or Waste/Sl, or Tissue, Air)</th> <th>Field Filled Sample (Yes or No)</th> <th>Perform MS/MSD (Yes or No)</th> <th>625.1, 625.1_SIM</th> <th>8015B_GRO_LL - (MOD) GRO</th> <th>8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18</th> <th>625.2_PREC - (MOD) 625plus Plus TICs</th> <th>537_DW_PREC - 537 1 Full List</th> <th>533 - All Analytes</th> </tr> </thead> <tbody> <tr> <td>Halawa Wells Units 1 &amp; 2 P1</td> <td>24-Nov-2025</td> <td>1058</td> <td>G</td> <td>Water</td> <td>X</td> <td>X</td> <td>4</td> <td>5</td> <td>4</td> <td>2</td> <td></td> <td></td> </tr> <tr> <td>Halawa Wells Units 1 &amp; 2 (Matrix Spike)</td> <td></td> <td></td> <td></td> <td>Water</td> <td></td> <td></td> <td>X</td> <td>X</td> <td>X</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Halawa Wells Units 1 &amp; 2 (Matrix Spike Duplicate)</td> <td></td> <td></td> <td></td> <td>Water</td> <td></td> <td></td> <td>X</td> <td>X</td> <td>X</td> <td></td> <td></td> <td></td> </tr> <tr> <td>TB: Halawa Wells Units 1 &amp; 2</td> <td>24-Nov-2025</td> <td>1058</td> <td></td> <td>Water</td> <td></td> <td></td> <td></td> <td></td> <td>2</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Aiea Wells Pumps 1&amp;2 (260) P2</td> <td>24-Nov-2025</td> <td>1244</td> <td>G</td> <td>Water</td> <td></td> <td></td> <td>2</td> <td>3</td> <td>2</td> <td>2</td> <td></td> <td></td> </tr> <tr> <td>Aiea Wells Pumps 1&amp;2 (260) (Matrix Spike)</td> <td></td> <td></td> <td></td> <td>Water</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>Aiea Wells Pumps 1&amp;2 (Matrix Spike Duplicate)</td> <td></td> <td></td> <td></td> <td>Water</td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> <td></td> </tr> <tr> <td>TB: Aiea Wells Pumps 1&amp;2 (260)</td> <td>24-Nov-2025</td> <td>1244</td> <td></td> <td>Water</td> <td></td> <td></td> <td></td> <td></td> <td>2</td> <td></td> <td></td> <td></td> </tr> </tbody> </table>								Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (Water, Soil, or Waste/Sl, or Tissue, Air)	Field Filled Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 625.1_SIM	8015B_GRO_LL - (MOD) GRO	8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18	625.2_PREC - (MOD) 625plus Plus TICs	537_DW_PREC - 537 1 Full List	533 - All Analytes	Halawa Wells Units 1 & 2 P1	24-Nov-2025	1058	G	Water	X	X	4	5	4	2			Halawa Wells Units 1 & 2 (Matrix Spike)				Water			X	X	X				Halawa Wells Units 1 & 2 (Matrix Spike Duplicate)				Water			X	X	X				TB: Halawa Wells Units 1 & 2	24-Nov-2025	1058		Water					2				Aiea Wells Pumps 1&2 (260) P2	24-Nov-2025	1244	G	Water			2	3	2	2			Aiea Wells Pumps 1&2 (260) (Matrix Spike)				Water									Aiea Wells Pumps 1&2 (Matrix Spike Duplicate)				Water									TB: Aiea Wells Pumps 1&2 (260)	24-Nov-2025	1244		Water					2			
Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (Water, Soil, or Waste/Sl, or Tissue, Air)	Field Filled Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1, 625.1_SIM	8015B_GRO_LL - (MOD) GRO	8015B_DRO_LL_CS - HHL Ranges: C10-C24/C24-C36/C8-C18	625.2_PREC - (MOD) 625plus Plus TICs	537_DW_PREC - 537 1 Full List	533 - All Analytes																																																																																																																				
Halawa Wells Units 1 & 2 P1	24-Nov-2025	1058	G	Water	X	X	4	5	4	2																																																																																																																						
Halawa Wells Units 1 & 2 (Matrix Spike)				Water			X	X	X																																																																																																																							
Halawa Wells Units 1 & 2 (Matrix Spike Duplicate)				Water			X	X	X																																																																																																																							
TB: Halawa Wells Units 1 & 2	24-Nov-2025	1058		Water					2																																																																																																																							
Aiea Wells Pumps 1&2 (260) P2	24-Nov-2025	1244	G	Water			2	3	2	2																																																																																																																						
Aiea Wells Pumps 1&2 (260) (Matrix Spike)				Water																																																																																																																												
Aiea Wells Pumps 1&2 (Matrix Spike Duplicate)				Water																																																																																																																												
TB: Aiea Wells Pumps 1&2 (260)	24-Nov-2025	1244		Water					2																																																																																																																							
<b>Possible Hazard Identification</b>		<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		<b>Sample Disposal ( A fee may be assessed if samples are retained longer than 1 month)</b>																																																																																																																												
Deliverable Requested I, II, III, IV, Other (specify)				<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months																																																																																																																												
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:																																																																																																																										
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:																																																																																																																						
Relinquished by:		Date/Time:		Company:		Received by:		Date/Time:		Company:																																																																																																																						
Custody Seals Intact <input type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.		Cooler Temperature(s) °C and Other Remarks.																																																																																																																												

**Eurofins Eaton Analytical Pomona**

941 Corporate Center Drive  
 Pomona, CA 91768-2642  
 Phone: 626-386-1100

**Chain of Custody Record**



eurofins | Environ

Loc: 380  
**185035**

<b>Client Information (Sub Contract Lab)</b>		Sampler: N/A		Lab PM: Lopez, Maria		Carrier Tracking No(s): N/A		COC No: 380-280803.1									
Client Contact: Shipping/Receiving		Phone: N/A		E-Mail: Maria.Lopez@et.eurofinsus.com		State of Origin: Hawaii		Page: Page 1 of 1									
Company: Eurofins Environment Testing Southwest,				Accreditations Required (See note): State - Hawaii				Job #: 380-185035-1									
Address: 2841 Dow Avenue, Suite 100,		Due Date Requested: 12/11/2025		<b>Analysis Requested</b>   380-185035 Chain of Custody						Preservation Codes:							
City: Tustin		TAT Requested (days): N/A															
State, Zip: CA, 92780		PO #: N/A															
Phone: 714-895-5494(Tel)		WO #: N/A															
Email: N/A		Project #: 38001111															
Project Name: RED-HILL		SSOW#: N/A															
Site: Honolulu BWS Sites																	
<b>Sample Identification - Client ID (Lab ID)</b>		<b>Sample Date</b>	<b>Sample Time</b>	<b>Sample Type (C=comp, G=grab)</b>	<b>Matrix (W=water, S=sediment, O=soil, BT=Tissue, AA=Air)</b>	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	625.1_SIME25_Prep(MOD) Extended PAM List	625.1625_Prep(MOD) Tentatively Identified Compounds (Hold)	80188_DRO_LL_CS2610C_LLHL Ranges: C-10-C24/C24-C36/C8-C18	80188_GRO_LL/6030C(MOD) GRO	Total Number	<b>Special Instructions/Note:</b>				
				Preservation Code:													
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X		MRLs are needed. Confirm any hits >RL.				
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1M)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X		MRLs are needed. Confirm any hits >RL.				
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1M)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X		MRLs are needed. Confirm any hits >RL.				
AIEA WELLS PUMPS 1&2 (260) (331-203-TP400) (380-185035-2)		11/24/25	12:44 Hawaiian	G	Water			X	X	X	X	6	MRLs are needed. Confirm any hits >RL.				
TB: HALAWA WELLS UNITS 1&2 (331-206-TP065) (380-185035-1)		11/24/25	10:58 Hawaiian	G	Water						X	2	MRLs are needed.				
TB: AIEA WELLS PUMPS 1&2 (260) (331-203-TP400) (380-185035-2)		11/24/25	12:44 Hawaiian	G	Water						X	2	MRLs are needed.				
Note: Since laboratory accreditations are subject to change, Eurofins Eaton Analytical, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/matrix being analyzed, the samples must be shipped back to the Eurofins Eaton Analytical, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Eaton Analytical, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Eaton Analytical, LLC.																	
<b>Possible Hazard Identification</b>						<b>Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)</b>											
Unconfirmed						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months											
Deliverable Requested: I, II, III, IV, Other (specify)						Primary Deliverable Rank: 2						Special Instructions/QC Requirements:					
Empty Kit Relinquished by:			Date:			Time:			Method of Shipment:								
Relinquished by: <i>[Signature]</i>			Date/Time: 11/29/25 10:05			Company: <i>een</i>			Received by: <i>[Signature]</i>			Date/Time: 11/29/25 10:05			Company: <i>WP</i>		
Relinquished by: <i>[Signature]</i>			Date/Time: 11/29/25 10:45			Company: <i>WP</i>			Received by: <i>[Signature]</i>			Date/Time: 11/29/25 10:45			Company: <i>SC</i>		
Relinquished by:			Date/Time:			Company:			Received by:			Date/Time:			Company:		
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:				Cooler Temperature(s) °C and Other Remarks: 0.9/1.0 1.3/1.4 FRY											

**Eurofins Eaton Analytical Pomona**


941 Corporate Center Drive  
Pomona, CA 91768-2642  
Phone: 626-386-1100

**Chain of Custody Record**



eurofins | Environ

Loc: 380  
**185035**

<b>Client Information (Sub Contract Lab)</b>		Sampler: N/A	Lab PM: Lopez, Maria	Carrier Tracking No(s): N/A	COC No: 380-280803.1									
Client Contact: Shipping/Receiving		Phone: N/A	E-Mail: Maria.Lopez@et.eurofinsus.com	State of Origin: Hawaii	Page: Page 1 of 1									
Company: Eurofins Environment Testing Southwest,		Accreditations Required (See note): State - Hawaii			Job #: 380-185035-1									
Address: 2841 Dow Avenue, Suite 100,		Due Date Requested: 12/11/2025	<b>Analysis Requested</b>   380-185035 Chain of Custody		Preservation Codes: -									
City: Tustin		TAT Requested (days): N/A												
State, Zip: CA, 92780		PO #: N/A												
Phone: 714-895-5494(Tel)		WO #: N/A												
Email: N/A		Project #: 38001111												
Project Name: RED-HILL		SSOW#: N/A												
Site: Honolulu BWS Sites														
<b>Sample Identification - Client ID (Lab ID)</b>		<b>Sample Date</b>	<b>Sample Time</b>	<b>Sample Type (C=comp, G=grab)</b>	<b>Matrix (W=water, S=sediment, O=soil, BT=Tissue, AA=Air)</b>	<b>Field Filtered Sample (Yes or No)</b>	<b>Perform MS/MSD (Yes or No)</b>	<b>625.1_SIME25_Prep(MOD) Extended PAM List</b>	<b>625.1625_Prep(MOD) Tentatively Identified Compounds (Hold)</b>	<b>80188_DRO_LL_CS2610C_LLHL Ranges: C-10, C24/C24-C36/C8-C18</b>	<b>80188_GRO_LL/6030C(MOD) GRO</b>	<b>Total Number</b>	<b>Special Instructions/Note:</b>	
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X	5	MRLs are needed. Confirm any hits >RL.	
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1M)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X	5	MRLs are needed. Confirm any hits >RL.	
HALAWA WELLS UNITS 1 & 2 (331-206-TP065) (380-185035-1M)		11/24/25	10:58 Hawaiian	G	Water			X	X	X	X	5	MRLs are needed. Confirm any hits >RL.	
AIEA WELLS PUMPS 1&2 (260) (331-203-TP400) (380-185035-2)		11/24/25	12:44 Hawaiian	G	Water			X	X	X	X	6	MRLs are needed. Confirm any hits >RL.	
TB: HALAWA WELLS UNITS 1&2 (331-206-TP065) (380-185035-1)		11/24/25	10:58 Hawaiian	G	Water						X	2	MRLs are needed.	
TB: AIEA WELLS PUMPS 1&2 (260) (331-203-TP400) (380-185035-2)		11/24/25	12:44 Hawaiian	G	Water						X	2	MRLs are needed.	
Note: Since laboratory accreditations are subject to change, Eurofins Eaton Analytical, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/matrix being analyzed, the samples must be shipped back to the Eurofins Eaton Analytical, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Eaton Analytical, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Eaton Analytical, LLC.														
<b>Possible Hazard Identification</b>						<b>Sample Disposal ( A fee may be assessed if samples are retained longer than 1 month)</b>								
Unconfirmed						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months								
Deliverable Requested: I, II, III, IV, Other (specify)						Primary Deliverable Rank: 2						Special Instructions/QC Requirements:		
Empty Kit Relinquished by:			Date:			Time:			Method of Shipment:					
Relinquished by: <i>[Signature]</i>			Date/Time: 11/29/25 10:05			Company: <i>een</i>			Received by: <i>[Signature]</i>			Date/Time: 11/29/25 10:05		Company: <i>WP</i>
Relinquished by: <i>[Signature]</i>			Date/Time: 11/29/25 10:45			Company: <i>WP</i>			Received by: <i>[Signature]</i>			Date/Time: 11/29/25 10:45		Company: <i>SC</i>
Relinquished by:			Date/Time:			Company:			Received by:			Date/Time:		Company:
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:				Cooler Temperature(s) °C and Other Remarks: 0.9/1.0 1.3/1.4 FRY								

## Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-185035-1

SDG Number: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea Wells Pumps 1&2 P2


**Login Number: 185035**

**List Source: Eurofins Eaton Analytical Pomona**

**List Number: 1**

**Creator: Ngo, Theodore**

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	Containers recd broken. Sufficient sample in remaining containers for analysis.
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	



# Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-185035-1

SDG Number: Weekly: Halawa Wells Units 1&2 P1 (MS/MSD), Aiea Wells Pumps 1&2 P2

**Login Number: 185035**

**List Source: Eurofins Calscience**

**List Number: 2**

**List Creation: 11/29/25 11:12 AM**

**Creator: Yu, Tiffany**

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	1.0,1.4
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	

