

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 1/5/2026 8:34:34 PM

JOB DESCRIPTION

RED-HILL
PFAS: Aiea Gulch Wells Pump 1

JOB NUMBER

380-190182-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
1/5/2026 8:34:34 PM

Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	11
Surrogate Summary	12
Isotope Dilution Summary	13
QC Sample Results	15
QC Association Summary	26
Lab Chronicle	27
Certification Summary	28
Method Summary	29
Sample Summary	30
Chain of Custody	31
Receipt Checklists	32



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-190182-1

Job ID: 380-190182-1

Eurofins Pomona

Job Narrative 380-190182-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 12/31/2025 9:57 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.3°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-1

☐ No Detections.

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

☐ No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-1

Date Collected: 12/29/25 10:50

Matrix: Drinking Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:43	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	87		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C6 PFDA	98		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C5 PFHxA	101		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C4 PFHpA	98		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C8 PFOA	100		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C9 PFNA	101		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C7 PFUnA	102		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C2 PFDoA	98		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C4 PFBA	100		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C5 PFPeA	108		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C3 PFBS	103		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C3 PFHxS	103		50 - 200	01/02/26 06:37	01/02/26 20:43	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-1

Date Collected: 12/29/25 10:50

Matrix: Drinking Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	104		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C2-4:2-FTS	114		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C2-6:2-FTS	119		50 - 200	01/02/26 06:37	01/02/26 20:43	1
13C2-8:2-FTS	106		50 - 200	01/02/26 06:37	01/02/26 20:43	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:31	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	104		70 - 130	01/02/26 02:20	01/03/26 00:31	1
13C2 PFHxA	115		70 - 130	01/02/26 02:20	01/03/26 00:31	1
13C2 PFDA	112		70 - 130	01/02/26 02:20	01/03/26 00:31	1
13C3-GenX	111		70 - 130	01/02/26 02:20	01/03/26 00:31	1

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

Date Collected: 12/29/25 10:50

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

Date Collected: 12/29/25 10:50

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		01/02/26 06:37	01/03/26 00:04	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	87		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C6 PFDA	105		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C5 PFHxA	105		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C4 PFHpA	100		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C8 PFOA	104		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C9 PFNA	104		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C7 PFUnA	106		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C2 PFDoA	100		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C4 PFBA	105		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C5 PFPeA	102		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C3 PFBS	98		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C3 PFHxS	102		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C8 PFOS	99		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C2-4:2-FTS	113		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C2-6:2-FTS	109		50 - 200	01/02/26 06:37	01/03/26 00:04	1
13C2-8:2-FTS	107		50 - 200	01/02/26 06:37	01/03/26 00:04	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

Date Collected: 12/29/25 10:50

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		01/02/26 02:20	01/03/26 00:40	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	105		70 - 130			01/02/26 02:20	01/03/26 00:40	1
13C2 PFHxA	111		70 - 130			01/02/26 02:20	01/03/26 00:40	1
13C2 PFDA	113		70 - 130			01/02/26 02:20	01/03/26 00:40	1
13C3-GenX	103		70 - 130			01/02/26 02:20	01/03/26 00:40	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL Limit	RL	Method	Prep Type
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-190182-1	AIEA GULCH WELLS PUMP 1 (331	104	115	112	111
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-190126-B-1-A MS	Matrix Spike	106	118	108	107
380-190126-C-1-A MSD	Matrix Spike Duplicate	106	119	106	112
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	105	111	113	103
LCS 380-195449/21-A	Lab Control Sample	91	95	107	89
MBL 380-195449/19-A	Method Blank	96	103	109	97
MRL 380-195449/20-A	Lab Control Sample	100	111	105	95
Surrogate Legend					
d5NEFOS = d5-NEtFOSAA					
PFHxA = 13C2 PFHxA					
PFDA = 13C2 PFDA					
GenX = 13C3-GenX					

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-190182-1	AIEA GULCH WELLS PUMP 1 (331	87	98	101	98	100	101	102	98
380-190182-1 MS	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	97	107	105	97	103	107	108	103
380-190182-1 MSD	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	94	100	97	95	96	101	103	100
		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-190182-1	AIEA GULCH WELLS PUMP 1 (331	10	108	103	103	104	114	119	106
380-190182-1 MS	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	99	101	101	102	105	116	107	105
380-190182-1 MSD	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	103	103	103	105	102	118	111	105
Surrogate Legend									
HFPODA = 13C3 HFPO-DA									
C6PFDA = 13C6 PFDA									
13C5PHA = 13C5 PFHxA									
C4PFHA = 13C4 PFHpA									
C8PFOA = 13C8 PFOA									
C9PFNA = 13C9 PFNA									
13C7PUA = 13C7 PFUnA									
PFDaA = 13C2 PFDaA									
PFBA = 13C4 PFBA									
PFPeA = 13C5 PFPeA									
C3PFBS = 13C3 PFBS									
C3PFHS = 13C3 PFHxS									
C8PFOS = 13C8 PFOS									
42FTS = 13C2-4:2-FTS									
62FTS = 13C2-6:2-FTS									
82FTS = 13C2-8:2-FTS									

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFDaA (50-200)
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (;	87	105	105	100	104	104	106	100
LCS 380-195460/24-A	Lab Control Sample	100	104	102	100	102	101	108	101
MBL 380-195460/22-A	Method Blank	87	98	102	105	104	101	104	100
MRL 380-195460/23-A	Lab Control Sample	87	99	94	99	102	101	99	98
		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (;	105	102	98	102	99	113	109	107
LCS 380-195460/24-A	Lab Control Sample	104	104	103	104	101	113	113	105
MBL 380-195460/22-A	Method Blank	102	107	105	102	102	116	112	106
MRL 380-195460/23-A	Lab Control Sample	101	105	98	102	101	117	109	106
Surrogate Legend									

Eurofins Pomona

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-195460/22-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 195460

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<0.30		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<0.30		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		01/02/26 06:37	01/02/26 20:15	1
Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	87		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C6 PFDA	98		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C5 PFHxA	102		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C4 PFHpA	105		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C8 PFOA	104		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C9 PFNA	101		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C7 PFUnA	104		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C2 PFDoA	100		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C4 PFBA	102		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C5 PFPeA	107		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C3 PFBS	105		50 - 200			01/02/26 06:37	01/02/26 20:15	1
13C3 PFHxS	102		50 - 200			01/02/26 06:37	01/02/26 20:15	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-195460/22-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 195460

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	102		50 - 200	01/02/26 06:37	01/02/26 20:15	1
13C2-4:2-FTS	116		50 - 200	01/02/26 06:37	01/02/26 20:15	1
13C2-6:2-FTS	112		50 - 200	01/02/26 06:37	01/02/26 20:15	1
13C2-8:2-FTS	106		50 - 200	01/02/26 06:37	01/02/26 20:15	1

Lab Sample ID: LCS 380-195460/24-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	120	101		ng/L		84	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	108		ng/L		90	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	113		ng/L		94	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	110		ng/L		92	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	116		ng/L		96	70 - 130
Perfluorodecanoic acid (PFDA)	120	114		ng/L		95	70 - 130
Perfluorododecanoic acid (PFDoA)	120	119		ng/L		99	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	114		ng/L		95	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	113		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	120	115		ng/L		96	70 - 130
Perfluorononanoic acid (PFNA)	120	117		ng/L		97	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	114		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	120	115		ng/L		96	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	114		ng/L		95	70 - 130
Perfluorobutanoic acid (PFBA)	120	113		ng/L		94	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	118		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	110		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	115		ng/L		96	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	123		ng/L		103	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	120	121		ng/L		101	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	117		ng/L		98	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	97.1		ng/L		81	70 - 130
Perfluoropentanoic acid (PFPeA)	120	118		ng/L		98	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	121		ng/L		100	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-195460/24-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	114		ng/L		95	70 - 130
	LCS	LCS					
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	100		50 - 200				
13C6 PFDA	104		50 - 200				
13C5 PFHxA	102		50 - 200				
13C4 PFHpA	100		50 - 200				
13C8 PFOA	102		50 - 200				
13C9 PFNA	101		50 - 200				
13C7 PFUnA	108		50 - 200				
13C2 PFDoA	101		50 - 200				
13C4 PFBA	104		50 - 200				
13C5 PFPeA	104		50 - 200				
13C3 PFBS	103		50 - 200				
13C3 PFHxS	104		50 - 200				
13C8 PFOS	101		50 - 200				
13C2-4:2-FTS	113		50 - 200				
13C2-6:2-FTS	113		50 - 200				
13C2-8:2-FTS	105		50 - 200				

Lab Sample ID: MRL 380-195460/23-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.00	1.87	J	ng/L		93	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.00	2.12	J	ng/L		106	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.00	1.99	J	ng/L		99	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.00	2.11	J	ng/L		105	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.00	2.28	J	ng/L		114	50 - 150
Perfluorodecanoic acid (PFDA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluorododecanoic acid (PFDoA)	2.00	2.19	J	ng/L		109	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.00	2.13	J	ng/L		106	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.00	2.16	J	ng/L		108	50 - 150
Perfluorohexanoic acid (PFHxA)	2.00	2.27	J	ng/L		113	50 - 150
Perfluorononanoic acid (PFNA)	2.00	2.10	J	ng/L		105	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.00	2.03	J	ng/L		102	50 - 150
Perfluorooctanoic acid (PFOA)	2.00	2.09	J	ng/L		104	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.00	2.17	J	ng/L		108	50 - 150
Perfluorobutanoic acid (PFBA)	2.00	2.25	J	ng/L		112	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-195460/23-A

Matrix: Water

Analysis Batch: 195607

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.00	2.26	J	ng/L		113	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.00	2.23	J	ng/L		111	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.00	2.20	J	ng/L		110	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.00	2.21	J	ng/L		110	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	2.00	2.47	J	ng/L		123	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.00	2.08	J	ng/L		104	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.00	1.76	J	ng/L		88	50 - 150
Perfluoropentanoic acid (PFPeA)	2.00	2.26	J	ng/L		113	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.00	2.08	J	ng/L		104	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.00	1.92	J	ng/L		96	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	Limits
13C3 HFPO-DA	87		50 - 200
13C6 PFDA	99		50 - 200
13C5 PFHxA	94		50 - 200
13C4 PFHpA	99		50 - 200
13C8 PFOA	102		50 - 200
13C9 PFNA	101		50 - 200
13C7 PFUnA	99		50 - 200
13C2 PFDoA	98		50 - 200
13C4 PFBA	101		50 - 200
13C5 PFPeA	105		50 - 200
13C3 PFBS	98		50 - 200
13C3 PFHxS	102		50 - 200
13C8 PFOS	101		50 - 200
13C2-4:2-FTS	117		50 - 200
13C2-6:2-FTS	109		50 - 200
13C2-8:2-FTS	106		50 - 200

Lab Sample ID: 380-190182-1 MS

Matrix: Drinking Water

Analysis Batch: 195607

Client Sample ID: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	97.3		ng/L		81	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	102		ng/L		85	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	111		ng/L		92	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-190182-1 MS

Client Sample ID: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Matrix: Drinking Water

Prep Type: Total/NA

Analysis Batch: 195607

Prep Batch: 195460

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		121	110		ng/L		91	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	118		ng/L		98	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		121	113		ng/L		93	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		121	112		ng/L		93	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		121	112		ng/L		93	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	114		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		121	110		ng/L		91	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		121	109		ng/L		91	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	110		ng/L		91	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		121	114		ng/L		94	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		121	114		ng/L		94	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		121	111		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	112		ng/L		93	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	108		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	121		ng/L		100	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	96.7		ng/L		80	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEA)	<2.0		121	120		ng/L		99	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	109		ng/L		90	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	89.6		ng/L		74	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		121	113		ng/L		93	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	115		ng/L		95	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	113		ng/L		94	70 - 130

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C3 HFPO-DA	97		50 - 200
13C6 PFDA	107		50 - 200
13C5 PFHxA	105		50 - 200
13C4 PFHpA	97		50 - 200
13C8 PFOA	103		50 - 200
13C9 PFNA	107		50 - 200
13C7 PFUnA	108		50 - 200
13C2 PFDoA	103		50 - 200
13C4 PFBA	99		50 - 200
13C5 PFPeA	101		50 - 200
13C3 PFBS	101		50 - 200
13C3 PFHxS	102		50 - 200
13C8 PFOS	105		50 - 200

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-190182-1 MS

Matrix: Drinking Water

Analysis Batch: 195607

Client Sample ID: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Prep Type: Total/NA

Prep Batch: 195460

Isotope Dilution	MS %Recovery	MS Qualifier	Limits
13C2-4:2-FTS	116		50 - 200
13C2-6:2-FTS	107		50 - 200
13C2-8:2-FTS	105		50 - 200

Lab Sample ID: 380-190182-1 MSD

Matrix: Drinking Water

Analysis Batch: 195607

Client Sample ID: AIEA GULCH WELLS PUMP 1 (331-201-TP071)

Prep Type: Total/NA

Prep Batch: 195460

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	99.3		ng/L		82	70 - 130	2	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	106		ng/L		88	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	108		ng/L		90	70 - 130	2	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		121	112		ng/L		93	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	113		ng/L		94	70 - 130	4	30
Perfluorodecanoic acid (PFDA)	<2.0		121	113		ng/L		94	70 - 130	0	30
Perfluorododecanoic acid (PFDoA)	<2.0		121	110		ng/L		92	70 - 130	2	30
Perfluorooheptanoic acid (PFHpA)	<2.0		121	110		ng/L		91	70 - 130	2	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	107		ng/L		88	70 - 130	6	30
Perfluorohexanoic acid (PFHxA)	<2.0		121	112		ng/L		92	70 - 130	1	30
Perfluorononanoic acid (PFNA)	<2.0		121	111		ng/L		92	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	111		ng/L		92	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		121	115		ng/L		95	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		121	113		ng/L		94	70 - 130	1	30
Perfluorobutanoic acid (PFBA)	<2.0		121	109		ng/L		91	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	115		ng/L		95	70 - 130	2	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	107		ng/L		89	70 - 130	1	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	112		ng/L		93	70 - 130	8	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	106		ng/L		88	70 - 130	9	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	120		ng/L		100	70 - 130	0	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	110		ng/L		91	70 - 130	2	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	97.5		ng/L		81	70 - 130	9	30
Perfluoropentanoic acid (PFPeA)	<2.0		121	112		ng/L		93	70 - 130	0	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	114		ng/L		95	70 - 130	0	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	108		ng/L		89	70 - 130	5	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	94		50 - 200
13C6 PFDA	100		50 - 200
13C5 PFHxA	97		50 - 200
13C4 PFHpA	95		50 - 200
13C8 PFOA	96		50 - 200
13C9 PFNA	101		50 - 200
13C7 PFUnA	103		50 - 200
13C2 PFDoA	100		50 - 200
13C4 PFBA	103		50 - 200
13C5 PFPeA	103		50 - 200
13C3 PFBS	103		50 - 200
13C3 PFHxS	105		50 - 200
13C8 PFOS	102		50 - 200
13C2-4:2-FTS	118		50 - 200
13C2-6:2-FTS	111		50 - 200
13C2-8:2-FTS	105		50 - 200

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-195449/19-A
Matrix: Water
Analysis Batch: 195552

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195449

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	<0.30		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		01/02/26 02:20	01/02/26 20:54	1
Surrogate	MBL MBL		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
d5-NEtFOSAA	96		70 - 130			01/02/26 02:20	01/02/26 20:54	1
13C2 PFHxA	103		70 - 130			01/02/26 02:20	01/02/26 20:54	1
13C2 PFDA	109		70 - 130			01/02/26 02:20	01/02/26 20:54	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MBL 380-195449/19-A

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 195449

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	97		70 - 130	01/02/26 02:20	01/02/26 20:54	1

Lab Sample ID: LCS 380-195449/21-A

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195449

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	25.1	20.2		ng/L		80	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	25.1	27.7		ng/L		111	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	25.5		ng/L		102	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	23.0		ng/L		92	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	23.3		ng/L		93	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	22.4		ng/L		89	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	26.3		ng/L		105	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	25.1		ng/L		100	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	25.7		ng/L		103	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	26.0		ng/L		104	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	25.4		ng/L		101	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	24.8		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	25.1	27.4		ng/L		109	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	22.4		ng/L		89	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	24.6		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	25.1	25.6		ng/L		102	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	22.5		ng/L		90	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	22.4		ng/L		89	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	91		70 - 130
13C2 PFHxA	95		70 - 130
13C2 PFDA	107		70 - 130
13C3-GenX	89		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MRL 380-195449/20-A

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195449

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	2.01	1.74	J	ng/L		87	50 - 150
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	2.01	2.06	J	ng/L		102	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	2.04	J	ng/L		102	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	1.76	J	ng/L		87	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.78	J	ng/L		89	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.97	J	ng/L		98	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.98	J	ng/L		98	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.00	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.08	J	ng/L		103	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.01	J	ng/L		100	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.02	J	ng/L		101	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	1.74	J	ng/L		87	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	2.11	J	ng/L		105	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.80	J	ng/L		90	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.64	J	ng/L		82	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.88	J	ng/L		94	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	100		70 - 130
13C2 PFHxA	111		70 - 130
13C2 PFDA	105		70 - 130
13C3-GenX	95		70 - 130

Lab Sample ID: 380-190126-B-1-A MS

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195449

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	<2.0		25.2	25.0		ng/L		99	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.2	28.1		ng/L		112	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.2	25.6		ng/L		102	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.2	24.3		ng/L		96	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 380-190126-B-1-A MS

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195449

	Sample	Sample	Spike	MS	MS				%Rec		
Analyte	Result	Qualifier	Added	Result	Qualifier	Unit	D	%Rec	Limits		
N-ethylperfluorooctanesulfonami doacetic acid (NEtFOSAA)	<2.0		25.2	26.0		ng/L		103	70 - 130		
Perfluorohexanoic acid (PFHxA)	<2.0		25.2	26.2		ng/L		104	70 - 130		
Perfluorododecanoic acid (PFDoA)	<2.0		25.2	25.9		ng/L		103	70 - 130		
Perfluorooctanoic acid (PFOA)	<2.0		25.2	26.7		ng/L		106	70 - 130		
Perfluorodecanoic acid (PFDA)	<2.0		25.2	25.5		ng/L		102	70 - 130		
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.2	28.5		ng/L		113	70 - 130		
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.2	30.3		ng/L		120	70 - 130		
Perfluoroheptanoic acid (PFHpA)	<2.0		25.2	28.2		ng/L		112	70 - 130		
Perfluorononanoic acid (PFNA)	<2.0		25.2	27.8		ng/L		111	70 - 130		
Perfluorotetradecanoic acid (PFTA)	<2.0		25.2	20.2		ng/L		80	70 - 130		
Perfluorotridecanoic acid (PFTrDA)	<2.0		25.2	26.0		ng/L		103	70 - 130		
9-Chlorohexadecafluoro-3-oxan onane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.2	27.1		ng/L		108	70 - 130		
11-Chloroeicosafluoro-3-oxaund ecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.2	24.1		ng/L		96	70 - 130		
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.2	26.0		ng/L		103	70 - 130		

Lab Sample ID: 380-190126-C-1-A MSD

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 195449

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.2	26.7		ng/L		106	70 - 130	7	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		25.2	26.9		ng/L		107	70 - 130	4	30
Perfluoroundecanoic acid (PFUnA)	<2.0		25.2	27.4		ng/L		109	70 - 130	7	30
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.2	26.0		ng/L		103	70 - 130	7	30
N-ethylperfluorooctanesulfonamidoacetic acid (NETFOSAA)	<2.0		25.2	25.5		ng/L		101	70 - 130	2	30
Perfluorohexanoic acid (PFHxA)	<2.0		25.2	27.8		ng/L		111	70 - 130	6	30
Perfluorododecanoic acid (PFDoA)	<2.0		25.2	26.3		ng/L		105	70 - 130	2	30
Perfluorooctanoic acid (PFOA)	<2.0		25.2	26.5		ng/L		105	70 - 130	1	30
Perfluorodecanoic acid (PFDA)	<2.0		25.2	25.5		ng/L		101	70 - 130	0	30

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: 380-190126-C-1-A MSD

Matrix: Water

Analysis Batch: 195552

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 195449

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorohexanesulfonic acid (PFHxS)	<2.0		25.2	27.5		ng/L		109	70 - 130	4	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		25.2	27.7		ng/L		110	70 - 130	9	30
Perfluoroheptanoic acid (PFHpA)	<2.0		25.2	27.5		ng/L		109	70 - 130	2	30
Perfluorononanoic acid (PFNA)	<2.0		25.2	27.4		ng/L		109	70 - 130	1	30
Perfluorotetradecanoic acid (PFTA)	<2.0		25.2	19.6		ng/L		78	70 - 130	3	30
Perfluorotridecanoic acid (PFTrDA)	<2.0		25.2	26.3		ng/L		104	70 - 130	1	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		25.2	25.0		ng/L		99	70 - 130	8	30
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		25.2	24.0		ng/L		95	70 - 130	1	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		25.2	26.6		ng/L		106	70 - 130	2	30
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
d5-NEtFOSAA	106		70 - 130								
13C2 PFHxA	119		70 - 130								
13C2 PFDA	106		70 - 130								
13C3-GenX	112		70 - 130								

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

LCMS

Prep Batch: 195449

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190182-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	537.1 DW	
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	537.1 DW	
MBL 380-195449/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-195449/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-195449/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-190126-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-190126-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Prep Batch: 195460

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190182-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	533	
MBL 380-195460/22-A	Method Blank	Total/NA	Water	533	
LCS 380-195460/24-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-195460/23-A	Lab Control Sample	Total/NA	Water	533	
380-190182-1 MS	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	
380-190182-1 MSD	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	

Analysis Batch: 195552

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190182-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	537.1	195449
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	537.1	195449
MBL 380-195449/19-A	Method Blank	Total/NA	Water	537.1	195449
LCS 380-195449/21-A	Lab Control Sample	Total/NA	Water	537.1	195449
MRL 380-195449/20-A	Lab Control Sample	Total/NA	Water	537.1	195449
380-190126-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	195449
380-190126-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	195449

Analysis Batch: 195607

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190182-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	195460
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	533	195460
MBL 380-195460/22-A	Method Blank	Total/NA	Water	533	195460
LCS 380-195460/24-A	Lab Control Sample	Total/NA	Water	533	195460
MRL 380-195460/23-A	Lab Control Sample	Total/NA	Water	533	195460
380-190182-1 MS	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	195460
380-190182-1 MSD	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	195460

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-1

Date Collected: 12/29/25 10:50

Matrix: Drinking Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195460	E9PK	EA POM	01/02/26 06:37
Total/NA	Analysis	533		1	195607	SZ9R	EA POM	01/02/26 20:43
Total/NA	Prep	537.1 DW			195449	G9MN	EA POM	01/02/26 02:20
Total/NA	Analysis	537.1		1	195552	SZ9R	EA POM	01/03/26 00:31

**Client Sample ID: FB AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-190182-2

Date Collected: 12/29/25 10:50

Matrix: Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195460	E9PK	EA POM	01/02/26 06:37
Total/NA	Analysis	533		1	195607	SZ9R	EA POM	01/03/26 00:04
Total/NA	Prep	537.1 DW			195449	G9MN	EA POM	01/02/26 02:20
Total/NA	Analysis	537.1		1	195552	SZ9R	EA POM	01/03/26 00:40

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190182-1
SDG: PFAS: Aiea Gulch Wells Pump 1

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-190182-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Drinking Water	12/29/25 10:50	12/31/25 09:57	Hawaii
380-190182-2	FB AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Water	12/29/25 10:50	12/31/25 09:57	Hawaii

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

Ver- 01/16/2019

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-190182-1

SDG Number: PFAS: Aiea Gulch Wells Pump 1

Login Number: 190182

List Number: 1

Creator: Ngo, Theodore

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	