

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 1/2/2026 2:38:12 PM

JOB DESCRIPTION

RED-HILL
PFAS: Aiea Gulch Wells P1/P2
RUSH Weekly Red Hill

JOB NUMBER

380-189509-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
1/2/2026 2:38:12 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	5
Detection Summary	6
Client Sample Results	7
Action Limit Summary	15
Surrogate Summary	17
Isotope Dilution Summary	18
QC Sample Results	20
QC Association Summary	31
Lab Chronicle	32
Certification Summary	33
Method Summary	34
Sample Summary	35
Chain of Custody	36
Receipt Checklists	37

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Qualifiers

LCMS

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-189509-1

Job ID: 380-189509-1

Eurofins Pomona

Job Narrative 380-189509-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 12/24/2025 10:37 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.7°C.

PFAS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-1

No Detections.

Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-2

No Detections.

Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-3

No Detections.

Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-4

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Pomona



Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-1

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:28	1
<i>Isotope Dilution</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>			<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
13C3 HFPO-DA	72		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C6 PFDA	85		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C5 PFHxA	82		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C4 PFHpA	81		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C8 PFOA	85		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C9 PFNA	84		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C7 PFUnA	91		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C2 PFDoA	93		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C4 PFBA	84		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C5 PFPeA	89		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C3 PFBS	113		50 - 200			12/31/25 06:49	12/31/25 20:28	1
13C3 PFHxS	103		50 - 200			12/31/25 06:49	12/31/25 20:28	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-1

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C8 PFOS	108		50 - 200	12/31/25 06:49	12/31/25 20:28	1
13C2-4:2-FTS	105		50 - 200	12/31/25 06:49	12/31/25 20:28	1
13C2-6:2-FTS	97		50 - 200	12/31/25 06:49	12/31/25 20:28	1
13C2-8:2-FTS	94		50 - 200	12/31/25 06:49	12/31/25 20:28	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
Perfluorotridecanoic acid (PFTrDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 09:59	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	107		70 - 130	12/29/25 16:25	12/31/25 09:59	1
13C2 PFHxA	106		70 - 130	12/29/25 16:25	12/31/25 09:59	1
13C2 PFDA	109		70 - 130	12/29/25 16:25	12/31/25 09:59	1
13C3-GenX	104		70 - 130	12/29/25 16:25	12/31/25 09:59	1

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-2

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-2

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:37	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	91		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C6 PFDA	106		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C5 PFHxA	108		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C4 PFHpA	103		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C8 PFOA	111		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C9 PFNA	107		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C7 PFUnA	106		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C2 PFDoA	101		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C4 PFBA	100		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C5 PFPeA	107		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C3 PFBS	104		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C3 PFHxS	102		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C8 PFOS	108		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C2-4:2-FTS	105		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C2-6:2-FTS	94		50 - 200	12/31/25 06:49	12/31/25 20:37	1
13C2-8:2-FTS	87		50 - 200	12/31/25 06:49	12/31/25 20:37	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-2

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:09	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	96		70 - 130	12/29/25 16:25	12/31/25 10:09	1
13C2 PFHxA	100		70 - 130	12/29/25 16:25	12/31/25 10:09	1
13C2 PFDA	104		70 - 130	12/29/25 16:25	12/31/25 10:09	1
13C3-GenX	100		70 - 130	12/29/25 16:25	12/31/25 10:09	1

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-3

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-3

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:47	1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	90		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C6 PFDA	102		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C5 PFHxA	107		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C4 PFHpA	101		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C8 PFOA	106		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C9 PFNA	105		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C7 PFUnA	103		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C2 PFDoA	102		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C4 PFBA	102		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C5 PFPeA	105		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C3 PFBS	103		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C3 PFHxS	104		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C8 PFOS	105		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C2-4:2-FTS	105		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C2-6:2-FTS	97		50 - 200			12/31/25 06:49	12/31/25 20:47	1
13C2-8:2-FTS	89		50 - 200			12/31/25 06:49	12/31/25 20:47	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
N-methylperfluorooctanesulfonamidooctic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-3

Date Collected: 12/22/25 09:00
Date Received: 12/24/25 10:37

Matrix: Water
PWSID Number: HI0000331

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
N-ethylperfluorooctanesulfonamidoacetic acid (NETFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Perfluorotridecanoic acid (PFTDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:19	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NETFOSAA	100		70 - 130			12/29/25 16:25	12/31/25 10:19	1
13C2 PFHxA	100		70 - 130			12/29/25 16:25	12/31/25 10:19	1
13C2 PFDA	107		70 - 130			12/29/25 16:25	12/31/25 10:19	1
13C3-GenX	98		70 - 130			12/29/25 16:25	12/31/25 10:19	1

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-4

Date Collected: 12/22/25 09:00
Date Received: 12/24/25 10:37

Matrix: Water
PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11CI-PF3OUdS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9CI-PF3ONS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluorobutanoic acid (PFBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-4

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Nonafluoro-3,6-dioxahexanoic acid (NFDHA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoropentanoic acid (PFPeA)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1
Perfluoropentanesulfonic acid (PFPeS)	<2.0		2.0	ng/L		12/31/25 06:49	12/31/25 20:57	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	89		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C6 PFDA	100		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C5 PFHxA	101		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C4 PFHpA	100		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C8 PFOA	106		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C9 PFNA	103		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C7 PFUnA	101		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C2 PFDoA	100		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C4 PFBA	102		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C5 PFPeA	110		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C3 PFBS	97		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C3 PFHxS	98		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C8 PFOS	103		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C2-4:2-FTS	95		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C2-6:2-FTS	89		50 - 200	12/31/25 06:49	12/31/25 20:57	1
13C2-8:2-FTS	86		50 - 200	12/31/25 06:49	12/31/25 20:57	1

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorooctanesulfonic acid (PFOS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluoroundecanoic acid (PFUnA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
N-methylperfluorooctanesulfonamide acetic acid (NMeFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
N-ethylperfluorooctanesulfonamide acetic acid (NEtFOSAA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorohexanoic acid (PFHxA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorododecanoic acid (PFDoA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorooctanoic acid (PFOA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorodecanoic acid (PFDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-4

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

PWSID Number: HI0000331

Method: EPA 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanesulfonic acid (PFHxS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorobutanesulfonic acid (PFBS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluoroheptanoic acid (PFHpA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorononanoic acid (PFNA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorotetradecanoic acid (PFTA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Perfluorotridecanoic acid (PFTTrDA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
9-Chlorohexadecafluoro-3-oxanonan e-1-sulfonic acid(9Cl-PF3ONS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
11-Chloroeicosafluoro-3-oxaundecan e-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		2.0	ng/L		12/29/25 16:25	12/31/25 10:28	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	104		70 - 130			12/29/25 16:25	12/31/25 10:28	1
13C2 PFHxA	110		70 - 130			12/29/25 16:25	12/31/25 10:28	1
13C2 PFDA	105		70 - 130			12/29/25 16:25	12/31/25 10:28	1
13C3-GenX	104		70 - 130			12/29/25 16:25	12/31/25 10:28	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-2

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-3

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA

Eurofins Pomona

Action Limit Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-189509-1
 SDG: PFAS: Aiea Gulch Wells P1/P2

Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071) (Continued)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-3

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)
PWSID Number: HI0000331

Lab Sample ID: 380-189509-4

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	533	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	533	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	533	Total/NA
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorooctanesulfonic acid (PFOS)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorooctanoic acid (PFOA)	<2.0		ng/L	4	2.0	537.1	Total/NA
Perfluorohexanesulfonic acid (PFHxS)	<2.0		ng/L	10	2.0	537.1	Total/NA
Perfluorononanoic acid (PFNA)	<2.0		ng/L	10	2.0	537.1	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-189509-1	AIEA GULCH WELLS PUMP 1 (331	107	106	109	104
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	96	100	104	100

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA
GenX = 13C3-GenX

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)			
		d5NEFOS (70-130)	PFHxA (70-130)	PFDA (70-130)	GenX (70-130)
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (100	100	107	98
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	104	110	105	104
380-189548-B-1-A MS	Matrix Spike	104	103	109	103
380-189548-C-1-A MSD	Matrix Spike Duplicate	103	106	107	103
LCS 380-194881/21-A	Lab Control Sample	103	113	104	108
MBL 380-194881/19-A	Method Blank	105	101	104	92
MRL 380-194881/20-A	Lab Control Sample	108	104	105	106

Surrogate Legend

d5NEFOS = d5-NEtFOSAA
PFHxA = 13C2 PFHxA
PFDA = 13C2 PFDA
GenX = 13C3-GenX

Isotope Dilution Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFD _o A (50-200)
380-189509-1	AIEA GULCH WELLS PUMP 1 (331	72	85	82	81	85	84	91	93
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	91	106	108	103	111	107	106	101

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-189509-1	AIEA GULCH WELLS PUMP 1 (331	84	89	113	103	108	105	97	94
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	100	107	104	102	108	105	94	87

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFD_oA = 13C2 PFD_oA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Matrix: Water

Prep Type: Total/NA

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	HFPODA (50-200)	C6PFDA (50-200)	13C5PHA (50-200)	C4PFHA (50-200)	C8PFOA (50-200)	C9PFNA (50-200)	13C7PUA (50-200)	PFD _o A (50-200)
380-189501-B-1-A MS	Matrix Spike	105	111	101	109	109	111	110	106
380-189501-C-1-A MSD	Matrix Spike Duplicate	107	109	111	103	108	110	111	111
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	90	102	107	101	106	105	103	102
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	89	100	101	100	106	103	101	100
LCS 380-195218/22-A	Lab Control Sample	104	108	110	102	109	108	108	105
MBL 380-195218/20-A	Method Blank	86	101	100	99	106	105	104	100
MRL 380-195218/21-A	Lab Control Sample	87	106	103	106	110	107	108	105

		Percent Isotope Dilution Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-189501-B-1-A MS	Matrix Spike	108	118	109	102	107	108	100	93
380-189501-C-1-A MSD	Matrix Spike Duplicate	110	118	104	98	106	111	96	90
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	102	105	103	104	105	105	97	89

Isotope Dilution Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-189509-1
 SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)							
		PFBA (50-200)	PFPeA (50-200)	C3PFBS (50-200)	C3PFHS (50-200)	C8PFOS (50-200)	42FTS (50-200)	62FTS (50-200)	82FTS (50-200)
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (102	110	97	98	103	95	89	86
LCS 380-195218/22-A	Lab Control Sample	107	107	104	100	107	99	98	88
MBL 380-195218/20-A	Method Blank	101	105	101	97	105	100	96	89
MRL 380-195218/21-A	Lab Control Sample	101	107	103	102	107	104	94	91

Surrogate Legend

- HFPODA = 13C3 HFPO-DA
- C6PFDA = 13C6 PFDA
- 13C5PHA = 13C5 PFHxA
- C4PFHA = 13C4 PFHpA
- C8PFOA = 13C8 PFOA
- C9PFNA = 13C9 PFNA
- 13C7PUA = 13C7 PFUnA
- PFDoA = 13C2 PFDoA
- PFBA = 13C4 PFBA
- PFPeA = 13C5 PFPeA
- C3PFBS = 13C3 PFBS
- C3PFHS = 13C3 PFHxS
- C8PFOS = 13C8 PFOS
- 42FTS = 13C2-4:2-FTS
- 62FTS = 13C2-6:2-FTS
- 82FTS = 13C2-8:2-FTS

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water

Lab Sample ID: MBL 380-195218/20-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195218

Analyte	MBL Result	MBL Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluorobutanoic acid (PFBA)	<0.69		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<0.38		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<0.37		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<0.48		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<0.47		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoro (2-ethoxyethane) sulfonic acid (PFEESA)	<0.25		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoro-3-methoxypropanoic acid (PFMPA)	<0.46		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoro-4-methoxybutanoic acid (PFMBA)	<0.15		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoropentanoic acid (PFPeA)	<0.38		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoroheptanesulfonic acid (PFHpS)	<0.36		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1
Perfluoropentanesulfonic acid (PFPeS)	<0.39		2.0	ng/L		12/31/25 06:49	12/31/25 17:55	1

Isotope Dilution	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3 HFPO-DA	86		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C6 PFDA	101		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C5 PFHxA	100		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C4 PFHpA	99		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C8 PFOA	106		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C9 PFNA	105		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C7 PFUnA	104		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C2 PFDoA	100		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C4 PFBA	101		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C5 PFPeA	105		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C3 PFBS	101		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C3 PFHxS	97		50 - 200	12/31/25 06:49	12/31/25 17:55	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MBL 380-195218/20-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195218

Isotope Dilution	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
13C8 PFOS	105		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C2-4:2-FTS	100		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C2-6:2-FTS	96		50 - 200	12/31/25 06:49	12/31/25 17:55	1
13C2-8:2-FTS	89		50 - 200	12/31/25 06:49	12/31/25 17:55	1

Lab Sample ID: LCS 380-195218/22-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195218

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	120	108		ng/L		90	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	120	114		ng/L		94	70 - 130
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	120	115		ng/L		96	70 - 130
Perfluorobutanesulfonic acid (PFBS)	120	116		ng/L		96	70 - 130
Perfluorodecanoic acid (PFDA)	120	111		ng/L		92	70 - 130
Perfluorododecanoic acid (PFDoA)	120	108		ng/L		90	70 - 130
Perfluoroheptanoic acid (PFHpA)	120	114		ng/L		94	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	120	113		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	120	103		ng/L		86	70 - 130
Perfluorononanoic acid (PFNA)	120	113		ng/L		94	70 - 130
Perfluorooctanesulfonic acid (PFOS)	120	111		ng/L		92	70 - 130
Perfluorooctanoic acid (PFOA)	120	108		ng/L		89	70 - 130
Perfluoroundecanoic acid (PFUnA)	120	106		ng/L		88	70 - 130
Perfluorobutanoic acid (PFBA)	120	110		ng/L		91	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	120	112		ng/L		93	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	120	119		ng/L		98	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	120	110		ng/L		92	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	120	106		ng/L		88	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	120	113		ng/L		94	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	120	105		ng/L		87	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	120	85.4		ng/L		71	70 - 130
Perfluoropentanoic acid (PFPeA)	120	109		ng/L		91	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	120	111		ng/L		92	70 - 130

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-189509-1
 SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: LCS 380-195218/22-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195218

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoropentanesulfonic acid (PFPeS)	120	118		ng/L		98	70 - 130
LCS LCS							
Isotope Dilution	%Recovery	Qualifier	Limits				
13C3 HFPO-DA	104		50 - 200				
13C6 PFDA	108		50 - 200				
13C5 PFHxA	110		50 - 200				
13C4 PFHpA	102		50 - 200				
13C8 PFOA	109		50 - 200				
13C9 PFNA	108		50 - 200				
13C7 PFUnA	108		50 - 200				
13C2 PFDoA	105		50 - 200				
13C4 PFBA	107		50 - 200				
13C5 PFPeA	107		50 - 200				
13C3 PFBS	104		50 - 200				
13C3 PFHxS	100		50 - 200				
13C8 PFOS	107		50 - 200				
13C2-4:2-FTS	99		50 - 200				
13C2-6:2-FTS	98		50 - 200				
13C2-8:2-FTS	88		50 - 200				

Lab Sample ID: MRL 380-195218/21-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195218

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.83	J	ng/L		91	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	2.04	J	ng/L		102	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	1.81	J	ng/L		90	50 - 150
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	2.16	J	ng/L		107	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.17	J	ng/L		108	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.01	J	ng/L		100	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.96	J	ng/L		98	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.02	J	ng/L		100	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.13	J	ng/L		106	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	1.92	J	ng/L		96	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.06	J	ng/L		103	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.98	J	ng/L		99	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	1.91	J	ng/L		95	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluorobutanoic acid (PFBA)	2.01	2.05	J	ng/L		102	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: MRL 380-195218/21-A
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195218

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	2.01	1.95	J	ng/L		97	50 - 150
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	2.01	2.18	J	ng/L		109	50 - 150
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	2.01	2.21	J	ng/L		110	50 - 150
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	2.01	1.89	J	ng/L		94	50 - 150
Perfluoro-3-methoxypropanoic acid (PFMPA)	2.01	1.79	J	ng/L		89	50 - 150
Perfluoro-4-methoxybutanoic acid (PFMBA)	2.01	1.35	J	ng/L		67	50 - 150
Perfluoropentanoic acid (PFPeA)	2.01	1.99	J	ng/L		99	50 - 150
Perfluoroheptanesulfonic acid (PFHpS)	2.01	1.84	J	ng/L		92	50 - 150
Perfluoropentanesulfonic acid (PFPeS)	2.01	1.92	J	ng/L		96	50 - 150

Isotope Dilution	MRL %Recovery	MRL Qualifier	MRL Limits
13C3 HFPO-DA	87		50 - 200
13C6 PFDA	106		50 - 200
13C5 PFHxA	103		50 - 200
13C4 PFHpA	106		50 - 200
13C8 PFOA	110		50 - 200
13C9 PFNA	107		50 - 200
13C7 PFUnA	108		50 - 200
13C2 PFDoA	105		50 - 200
13C4 PFBA	101		50 - 200
13C5 PFPeA	107		50 - 200
13C3 PFBS	103		50 - 200
13C3 PFHxS	102		50 - 200
13C8 PFOS	107		50 - 200
13C2-4:2-FTS	104		50 - 200
13C2-6:2-FTS	94		50 - 200
13C2-8:2-FTS	91		50 - 200

Lab Sample ID: 380-189501-B-1-A MS
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 195218

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	106		ng/L		88	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	110		ng/L		91	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	105		ng/L		87	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-189501-B-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 195408

Prep Batch: 195218

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
Hexafluoropropylene Oxide	<2.0		121	116		ng/L		96	70 - 130
Dimer Acid (HFPO-DA/GenX)									
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	111		ng/L		92	70 - 130
Perfluorodecanoic acid (PFDA)	<2.0		121	108		ng/L		90	70 - 130
Perfluorododecanoic acid (PFDoA)	<2.0		121	114		ng/L		94	70 - 130
Perfluoroheptanoic acid (PFHpA)	<2.0		121	107		ng/L		88	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	113		ng/L		94	70 - 130
Perfluorohexanoic acid (PFHxA)	<2.0		121	119		ng/L		99	70 - 130
Perfluorononanoic acid (PFNA)	<2.0		121	112		ng/L		93	70 - 130
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	114		ng/L		95	70 - 130
Perfluorooctanoic acid (PFOA)	<2.0		121	109		ng/L		91	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		121	107		ng/L		89	70 - 130
Perfluorobutanoic acid (PFBA)	<2.0		121	111		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	111		ng/L		92	70 - 130
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	107		ng/L		89	70 - 130
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	110		ng/L		91	70 - 130
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	129		ng/L		107	70 - 130
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	110		ng/L		91	70 - 130
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	118		ng/L		98	70 - 130
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	93.3		ng/L		77	70 - 130
Perfluoropentanoic acid (PFPeA)	<2.0		121	108		ng/L		90	70 - 130
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	113		ng/L		94	70 - 130
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	118		ng/L		97	70 - 130

Isotope Dilution	MS MS		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	105		50 - 200
13C6 PFDA	111		50 - 200
13C5 PFHxA	101		50 - 200
13C4 PFHpA	109		50 - 200
13C8 PFOA	109		50 - 200
13C9 PFNA	111		50 - 200
13C7 PFUnA	110		50 - 200
13C2 PFDoA	106		50 - 200
13C4 PFBA	108		50 - 200
13C5 PFPeA	118		50 - 200
13C3 PFBS	109		50 - 200
13C3 PFHxS	102		50 - 200
13C8 PFOS	107		50 - 200

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Lab Sample ID: 380-189501-B-1-A MS
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 195218

<i>Isotope Dilution</i>	<i>MS MS</i>	<i>Limits</i>
<i>%Recovery</i>	<i>Qualifier</i>	
13C2-4:2-FTS	108	50 - 200
13C2-6:2-FTS	100	50 - 200
13C2-8:2-FTS	93	50 - 200

Lab Sample ID: 380-189501-C-1-A MSD
Matrix: Water
Analysis Batch: 195408

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA
Prep Batch: 195218

<i>Analyte</i>	<i>Sample Result</i>	<i>Sample Qualifier</i>	<i>Spike Added</i>	<i>MSD MSD</i>		<i>Unit</i>	<i>D</i>	<i>%Rec</i>	<i>%Rec Limits</i>	<i>RPD</i>	<i>RPD Limit</i>
				<i>Result</i>	<i>Qualifier</i>						
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<2.0		121	106		ng/L		88	70 - 130	0	30
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<2.0		121	114		ng/L		94	70 - 130	4	30
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<2.0		121	109		ng/L		91	70 - 130	4	30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		121	118		ng/L		97	70 - 130	2	30
Perfluorobutanesulfonic acid (PFBS)	<2.0		121	118		ng/L		97	70 - 130	6	30
Perfluorodecanoic acid (PFDA)	<2.0		121	113		ng/L		93	70 - 130	4	30
Perfluorododecanoic acid (PFDoA)	<2.0		121	108		ng/L		89	70 - 130	5	30
Perfluoroheptanoic acid (PFHpA)	<2.0		121	117		ng/L		97	70 - 130	10	30
Perfluorohexanesulfonic acid (PFHxS)	<2.0		121	120		ng/L		99	70 - 130	6	30
Perfluorohexanoic acid (PFHxA)	<2.0		121	108		ng/L		90	70 - 130	10	30
Perfluorononanoic acid (PFNA)	<2.0		121	113		ng/L		93	70 - 130	1	30
Perfluorooctanesulfonic acid (PFOS)	<2.0		121	115		ng/L		95	70 - 130	1	30
Perfluorooctanoic acid (PFOA)	<2.0		121	111		ng/L		91	70 - 130	1	30
Perfluoroundecanoic acid (PFUnA)	<2.0		121	111		ng/L		92	70 - 130	3	30
Perfluorobutanoic acid (PFBA)	<2.0		121	108		ng/L		90	70 - 130	2	30
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2 FTS)	<2.0		121	115		ng/L		95	70 - 130	4	30
1H,1H,2H,2H-Perfluorohexane sulfonic acid (4:2 FTS)	<2.0		121	108		ng/L		89	70 - 130	1	30
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2 FTS)	<2.0		121	115		ng/L		95	70 - 130	4	30
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	<2.0		121	105		ng/L		87	70 - 130	21	30
Perfluoro (2-ethoxyethane) sulfonic acid (PFEEESA)	<2.0		121	117		ng/L		97	70 - 130	6	30
Perfluoro-3-methoxypropanoic acid (PFMPA)	<2.0		121	115		ng/L		95	70 - 130	3	30
Perfluoro-4-methoxybutanoic acid (PFMBA)	<2.0		121	96.7		ng/L		80	70 - 130	4	30
Perfluoropentanoic acid (PFPeA)	<2.0		121	110		ng/L		91	70 - 130	1	30
Perfluoroheptanesulfonic acid (PFHpS)	<2.0		121	115		ng/L		95	70 - 130	2	30
Perfluoropentanesulfonic acid (PFPeS)	<2.0		121	126		ng/L		104	70 - 130	7	30

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 533 - Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water (Continued)

Isotope Dilution	MSD MSD		Limits
	%Recovery	Qualifier	
13C3 HFPO-DA	107		50 - 200
13C6 PFDA	109		50 - 200
13C5 PFHxA	111		50 - 200
13C4 PFHpA	103		50 - 200
13C8 PFOA	108		50 - 200
13C9 PFNA	110		50 - 200
13C7 PFUnA	111		50 - 200
13C2 PFDoA	111		50 - 200
13C4 PFBA	110		50 - 200
13C5 PFPeA	118		50 - 200
13C3 PFBS	104		50 - 200
13C3 PFHxS	98		50 - 200
13C8 PFOS	106		50 - 200
13C2-4:2-FTS	111		50 - 200
13C2-6:2-FTS	96		50 - 200
13C2-8:2-FTS	90		50 - 200

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MBL 380-194881/19-A
Matrix: Water
Analysis Batch: 195216

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 194881

Analyte	MBL MBL		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<1.0		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorooctanesulfonic acid (PFOS)	<0.43		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluoroundecanoic acid (PFUnA)	<0.42		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<0.58		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	<0.42		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorohexanoic acid (PFHxA)	<0.46		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorododecanoic acid (PFDoA)	<0.54		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorooctanoic acid (PFOA)	<0.38		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorodecanoic acid (PFDA)	<0.31		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorohexanesulfonic acid (PFHxS)	<0.32		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorobutanesulfonic acid (PFBS)	<0.37		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluoroheptanoic acid (PFHpA)	<0.39		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorononanoic acid (PFNA)	<0.40		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorotetradecanoic acid (PFTA)	<0.54		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
Perfluorotridecanoic acid (PFTTrDA)	<0.36		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	<0.30		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	<0.30		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	<0.60		2.0	ng/L		12/29/25 16:25	12/31/25 08:31	1

Surrogate	MBL MBL		Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
d5-NEtFOSAA	105		70 - 130	12/29/25 16:25	12/31/25 08:31	1
13C2 PFHxA	101		70 - 130	12/29/25 16:25	12/31/25 08:31	1
13C2 PFDA	104		70 - 130	12/29/25 16:25	12/31/25 08:31	1

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MBL 380-194881/19-A
Matrix: Water
Analysis Batch: 195216

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 194881

Surrogate	MBL %Recovery	MBL Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C3-GenX	92		70 - 130	12/29/25 16:25	12/31/25 08:31	1

Lab Sample ID: LCS 380-194881/21-A
Matrix: Water
Analysis Batch: 195216

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 194881

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide	25.1	27.9		ng/L		112	70 - 130
Dimer Acid (HFPO-DA/GenX)							
Perfluorooctanesulfonic acid (PFOS)	25.1	27.7		ng/L		111	70 - 130
Perfluoroundecanoic acid (PFUnA)	25.1	25.2		ng/L		100	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	25.1	25.0		ng/L		100	70 - 130
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	25.1	25.4		ng/L		101	70 - 130
Perfluorohexanoic acid (PFHxA)	25.1	26.6		ng/L		106	70 - 130
Perfluorododecanoic acid (PFDoA)	25.1	27.6		ng/L		110	70 - 130
Perfluorooctanoic acid (PFOA)	25.1	26.7		ng/L		106	70 - 130
Perfluorodecanoic acid (PFDA)	25.1	26.0		ng/L		104	70 - 130
Perfluorohexanesulfonic acid (PFHxS)	25.1	28.0		ng/L		112	70 - 130
Perfluorobutanesulfonic acid (PFBS)	25.1	26.1		ng/L		104	70 - 130
Perfluoroheptanoic acid (PFHpA)	25.1	28.6		ng/L		114	70 - 130
Perfluorononanoic acid (PFNA)	25.1	26.5		ng/L		106	70 - 130
Perfluorotetradecanoic acid (PFTA)	25.1	19.4		ng/L		78	70 - 130
Perfluorotridecanoic acid (PFTrDA)	25.1	24.5		ng/L		98	70 - 130
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	25.1	27.1		ng/L		108	70 - 130
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	25.1	24.6		ng/L		98	70 - 130
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	25.1	27.5		ng/L		110	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
d5-NEtFOSAA	103		70 - 130
13C2 PFHxA	113		70 - 130
13C2 PFDA	104		70 - 130
13C3-GenX	108		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method: 537.1 - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MRL 380-194881/20-A
Matrix: Water
Analysis Batch: 195216

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 194881

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	2.01	2.10	J	ng/L		104	50 - 150
Perfluorooctanesulfonic acid (PFOS)	2.01	1.79	J	ng/L		89	50 - 150
Perfluoroundecanoic acid (PFUnA)	2.01	1.85	J	ng/L		92	50 - 150
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	2.01	2.10	J	ng/L		104	50 - 150
N-ethylperfluorooctanesulfonamidoacetic acid (NEtFOSAA)	2.01	2.16	J	ng/L		108	50 - 150
Perfluorohexanoic acid (PFHxA)	2.01	2.03	J	ng/L		101	50 - 150
Perfluorododecanoic acid (PFDoA)	2.01	1.78	J	ng/L		88	50 - 150
Perfluorooctanoic acid (PFOA)	2.01	2.11	J	ng/L		105	50 - 150
Perfluorodecanoic acid (PFDA)	2.01	2.09	J	ng/L		104	50 - 150
Perfluorohexanesulfonic acid (PFHxS)	2.01	2.15	J	ng/L		107	50 - 150
Perfluorobutanesulfonic acid (PFBS)	2.01	2.05	J	ng/L		102	50 - 150
Perfluoroheptanoic acid (PFHpA)	2.01	2.04	J	ng/L		101	50 - 150
Perfluorononanoic acid (PFNA)	2.01	2.03	J	ng/L		101	50 - 150
Perfluorotetradecanoic acid (PFTA)	2.01	1.23	J	ng/L		61	50 - 150
Perfluorotridecanoic acid (PFTrDA)	2.01	1.61	J	ng/L		80	50 - 150
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid(9Cl-PF3ONS)	2.01	1.94	J	ng/L		96	50 - 150
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	2.01	1.58	J	ng/L		78	50 - 150
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)	2.01	2.13	J	ng/L		106	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
d5-NEtFOSAA	108		70 - 130
13C2 PFHxA	104		70 - 130
13C2 PFDA	105		70 - 130
13C3-GenX	106		70 - 130

Lab Sample ID: 380-189548-B-1-A MS
Matrix: Water
Analysis Batch: 195216

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 194881

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA/GenX)	<2.0		25.1	25.5		ng/L		102	70 - 130
Perfluorooctanesulfonic acid (PFOS)	2.1		25.1	29.8		ng/L		110	70 - 130
Perfluoroundecanoic acid (PFUnA)	<2.0		25.1	26.1		ng/L		104	70 - 130
N-methylperfluorooctanesulfonamidoacetic acid (NMeFOSAA)	<2.0		25.1	25.8		ng/L		103	70 - 130

Eurofins Pomona

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-189509-1
 SDG: PFAS: Aiea Gulch Wells P1/P2

LCMS

Prep Batch: 194881

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-189509-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	537.1 DW	
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1 DW	
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	537.1 DW	
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	537.1 DW	
MBL 380-194881/19-A	Method Blank	Total/NA	Water	537.1 DW	
LCS 380-194881/21-A	Lab Control Sample	Total/NA	Water	537.1 DW	
MRL 380-194881/20-A	Lab Control Sample	Total/NA	Water	537.1 DW	
380-189548-B-1-A MS	Matrix Spike	Total/NA	Water	537.1 DW	
380-189548-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1 DW	

Analysis Batch: 195216

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-189509-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	537.1	194881
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	537.1	194881
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	537.1	194881
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	537.1	194881
MBL 380-194881/19-A	Method Blank	Total/NA	Water	537.1	194881
LCS 380-194881/21-A	Lab Control Sample	Total/NA	Water	537.1	194881
MRL 380-194881/20-A	Lab Control Sample	Total/NA	Water	537.1	194881
380-189548-B-1-A MS	Matrix Spike	Total/NA	Water	537.1	194881
380-189548-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	537.1	194881

Prep Batch: 195218

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-189509-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	533	
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	533	
MBL 380-195218/20-A	Method Blank	Total/NA	Water	533	
LCS 380-195218/22-A	Lab Control Sample	Total/NA	Water	533	
MRL 380-195218/21-A	Lab Control Sample	Total/NA	Water	533	
380-189501-B-1-A MS	Matrix Spike	Total/NA	Water	533	
380-189501-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	

Analysis Batch: 195408

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-189509-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Drinking Water	533	195218
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Drinking Water	533	195218
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Total/NA	Water	533	195218
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Total/NA	Water	533	195218
MBL 380-195218/20-A	Method Blank	Total/NA	Water	533	195218
LCS 380-195218/22-A	Lab Control Sample	Total/NA	Water	533	195218
MRL 380-195218/21-A	Lab Control Sample	Total/NA	Water	533	195218
380-189501-B-1-A MS	Matrix Spike	Total/NA	Water	533	195218
380-189501-C-1-A MSD	Matrix Spike Duplicate	Total/NA	Water	533	195218

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

**Client Sample ID: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-1

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195218	XTD8	EA POM	12/31/25 06:49
Total/NA	Analysis	533		1	195408	SZ9R	EA POM	12/31/25 20:28
Total/NA	Prep	537.1 DW			194881	E2HD	EA POM	12/29/25 16:25
Total/NA	Analysis	537.1		1	195216	M7ML	EA POM	12/31/25 09:59

**Client Sample ID: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-2

Date Collected: 12/22/25 09:00

Matrix: Drinking Water

Date Received: 12/24/25 10:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195218	XTD8	EA POM	12/31/25 06:49
Total/NA	Analysis	533		1	195408	SZ9R	EA POM	12/31/25 20:37
Total/NA	Prep	537.1 DW			194881	E2HD	EA POM	12/29/25 16:25
Total/NA	Analysis	537.1		1	195216	M7ML	EA POM	12/31/25 10:09

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 1
(331-201-TP071)**

Lab Sample ID: 380-189509-3

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195218	XTD8	EA POM	12/31/25 06:49
Total/NA	Analysis	533		1	195408	SZ9R	EA POM	12/31/25 20:47
Total/NA	Prep	537.1 DW			194881	E2HD	EA POM	12/29/25 16:25
Total/NA	Analysis	537.1		1	195216	M7ML	EA POM	12/31/25 10:19

**Client Sample ID: FB: AIEA GULCH WELLS PUMP 2
(331-202-TP072)**

Lab Sample ID: 380-189509-4

Date Collected: 12/22/25 09:00

Matrix: Water

Date Received: 12/24/25 10:37

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Batch Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	533			195218	XTD8	EA POM	12/31/25 06:49
Total/NA	Analysis	533		1	195408	SZ9R	EA POM	12/31/25 20:57
Total/NA	Prep	537.1 DW			194881	E2HD	EA POM	12/29/25 16:25
Total/NA	Analysis	537.1		1	195216	M7ML	EA POM	12/31/25 10:28

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Laboratory: Eurofins Pomona

The accreditations/certifications listed below are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Method	Method Description	Protocol	Laboratory
533	Perfluorinated and Polyfluorinated Alkyl Substances in Drinking Water	EPA	EA POM
537.1	Perfluorinated Alkyl Acids (LC/MS)	EPA	EA POM
533	Extraction of Perfluorinated and Polyfluorinated Alkyl Acids	EPA	EA POM
537.1 DW	Extraction of Perfluorinated Alkyl Acids	EPA	EA POM

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100



Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-189509-1
SDG: PFAS: Aiea Gulch Wells P1/P2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-189509-1	AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Drinking Water	12/22/25 09:00	12/24/25 10:37	HI0000331
380-189509-2	AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Drinking Water	12/22/25 09:00	12/24/25 10:37	HI0000331
380-189509-3	FB: AIEA GULCH WELLS PUMP 1 (331-201-TP071)	Water	12/22/25 09:00	12/24/25 10:37	HI0000331
380-189509-4	FB: AIEA GULCH WELLS PUMP 2 (331-202-TP072)	Water	12/22/25 09:00	12/24/25 10:37	HI0000331

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17

Chain of Custody Record



Client Information		Lab PM Arada, Rachelle		Carrier Tracking No(e): 380-28005-2757 1	
Sampler: Kai Edison		E-Mail: Rachelle.Arada@et.eurofins.com		Page: Page 1 of 1	
Phone: +1 808 748 5840		PWSID:		Job #:	
City & County of Honolulu		Due Date Requested:		Analysis Requested	
Address: 630 South Beretania Street Chemistry Lab Honolulu		TAT Requested (days):		Total Number of Containers	
State Zip: HI, 96843		Compliance Project: <input type="checkbox"/> Yes <input type="checkbox"/> No		Preservation Codes: R - NaThioSO4 RA - NaThioHCl O - Na2SO3 OA - Na2SO3/HCl Y - Trizma I - NH4 Acetate	
Phone: 808-748-5840 (Tel)		PO #: C20525101 exp 05312023		Other	
Email: kiramoto@hbws.org		WO #:		Special Instructions/Note:	
Project Name: RED-HILL/HBWS Sites Event Desc: RUSH Weekly Red Hill		Project #: 38001111		chlorinated	
Site: Hawaii		SSOW#:		chlorinated	
Sample Identification		Field Filtered Sample (Yes or No)		chlorinated	
Sample Date	Sample Time	Sample Type (C=Comp, G=grab)	Preservation Code:	RA	DA
22-Dec-2025	9:00	9	Water	3	3
22-Dec-2025	9:00	G	Water	3	3
22-Dec-2025	9:00	G	Water	1	1
22-Dec-2025	9:00	G	Water	1	1
Possible Hazard Identification		Perform MS/MSD (Yes or No)		Special Instructions/IOC Requirements	
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		Field Filtered Sample (Yes or No)		Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months	
Deliverable Requested I, II, III, IV, Other (specify)		Date Requested		Date/Time	
Empty Kit Relinquished by		Date: 12/22/25 11:00		Date/Time: 12/24/25 1037	
Relinquished by		Company: HBWS		Company: GSA	
Relinquished by		Company:		Company:	
Relinquished by		Company:		Company:	
Custody Seals Intact <input type="checkbox"/> Yes <input type="checkbox"/> No		Cooler Temperature(s) °C and Other Remarks: 0.7/0.7 631A BWS LG		Cooler Temperature(s) °C and Other Remarks:	

8873 STA 301 FAX

[Signature]



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-189509-1
SDG Number: PFAS: Aiea Gulch Wells P1/P2

Login Number: 189509

List Number: 1

Creator: Segura, Ryan

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	True	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

