

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 6/5/2025 12:19:22 PM

JOB DESCRIPTION

RED-HILL
Weekly
RUSH Weekly Red Hill

JOB NUMBER

380-151743-1

Eurofins Eaton Analytical Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Eaton Analytical, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.
(DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
6/5/2025 12:19:22 PM

Table of Contents

Cover Page 1

Table of Contents 3

Definitions/Glossary 4

Case Narrative 5

Detection Summary 6

Client Sample Results 7

Action Limit Summary 11

Surrogate Summary 12

QC Sample Results 15

QC Association Summary 31

Lab Chronicle 33

Certification Summary 34

Method Summary 36

Sample Summary 37

Chain of Custody 38

Receipt Checklists 40



Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
F5	Duplicate RPD exceeds limit, and one or both sample results are less than 5 times RL, and the absolute difference between results is < the upper reporting limits for both.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

GC Semi VOA

Qualifier	Qualifier Description
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-151743-1

Job ID: 380-151743-1

Eurofins Eaton Analytical Pomona

Job Narrative 380-151743-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 5/21/2025 9:52 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 0.6°C.

GC/MS Semi VOA

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Gasoline Range Organics

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Diesel Range Organics

Method 8015B_DRO_LL_CS: The method reporting limit check (MRL) for preparation batch 570-574445 and analytical batch 570-575632 recovered outside control limits for the following analytes: C10-C28. These analytes were biased high in the MRL and were not detected in the associated samples; therefore, the data have been reported.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)
PWSID Number: HI0000331

Lab Sample ID: 380-151743-1

☐ No Detections.

Client Sample ID: TB: MOANALUA WELLS

Lab Sample ID: 380-151743-2

☐ No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Eaton Analytical Pomona

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-151743-1

Date Collected: 05/19/25 10:08

Matrix: Drinking Water

Date Received: 05/21/25 09:52

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2,4'-DDD	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2,4'-DDE	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2,4'-DDT	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2,4-Dinitrotoluene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2,6-Dinitrotoluene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
2-Methylnaphthalene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
4,4'-DDD	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
4,4'-DDE	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
4,4'-DDT	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Acenaphthene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Acenaphthylene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Acetochlor	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Alachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
alpha-BHC	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
alpha-Chlordane	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Anthracene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 14:09	1
Atrazine	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Benz(a)anthracene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Benzo[a]pyrene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 14:09	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 14:09	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 14:09	1
beta-BHC	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		05/23/25 14:41	05/27/25 14:09	1
Bromacil	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Butachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Butylbenzylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 14:09	1
Chlorobenzilate	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Chloroneb	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Chlorothalonil (Draconil, Bravo)	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Chlorpyrifos	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Chrysene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 14:09	1
delta-BHC	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		05/23/25 14:41	05/27/25 14:09	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Dieldrin	<0.0099		0.0099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Diethylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 14:09	1
Dimethylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 14:09	1
Di-n-butyl phthalate	<0.99		0.99	ug/L		05/23/25 14:41	05/27/25 14:09	1
Di-n-octyl phthalate	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Endosulfan I (Alpha)	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Endosulfan II (Beta)	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Endosulfan sulfate	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Endrin	<0.0099		0.0099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Endrin aldehyde	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
EPTC	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Fluoranthene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-151743-1

Date Collected: 05/19/25 10:08

Matrix: Drinking Water

Date Received: 05/21/25 09:52

PWSID Number: HI0000331

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
gamma-Chlordane	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Heptachlor	<0.0099		0.0099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Heptachlor epoxide (isomer B)	<0.0099		0.0099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Hexachlorobenzene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Isophorone	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Lindane	<0.0099		0.0099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Malathion	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Methoxychlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Metolachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Molinate	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Naphthalene	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Parathion	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Pendimethalin (Penoxaline)	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Phenanthrene	<0.040		0.040	ug/L		05/23/25 14:41	05/27/25 14:09	1
Propachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Pyrene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Simazine	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Terbacil	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Terbutylazine	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Thiobencarb	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		05/23/25 14:41	05/27/25 14:09	1
trans-Nonachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 14:09	1
Trifluralin	<0.099		0.099	ug/L		05/23/25 14:41	05/27/25 14:09	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	05/23/25 14:41	05/27/25 14:09	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	92		70 - 130	05/23/25 14:41	05/27/25 14:09	1
Perylene-d12	91		70 - 130	05/23/25 14:41	05/27/25 14:09	1
Triphenylphosphate	108		70 - 130	05/23/25 14:41	05/27/25 14:09	1

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
2-Methylnaphthalene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Acenaphthene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Acenaphthylene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Anthracene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Benzo[a]anthracene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Benzo[a]pyrene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Benzo[b]fluoranthene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Benzo[g,h,i]perylene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Benzo[k]fluoranthene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Chrysene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Dibenz(a,h)anthracene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Fluoranthene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1

Eurofins Eaton Analytical Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-151743-1

Date Collected: 05/19/25 10:08

Matrix: Drinking Water

Date Received: 05/21/25 09:52

PWSID Number: HI0000331

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Indeno[1,2,3-cd]pyrene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Naphthalene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Phenanthrene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1
Pyrene	<0.19		0.19	ug/L		05/23/25 09:56	06/04/25 14:58	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	69		28 - 127	05/23/25 09:56	06/04/25 14:58	1
2-Fluorobiphenyl (Surr)	65		31 - 120	05/23/25 09:56	06/04/25 14:58	1
2-Fluorophenol (Surr)	41		17 - 120	05/23/25 09:56	06/04/25 14:58	1
Nitrobenzene-d5 (Surr)	67		27 - 120	05/23/25 09:56	06/04/25 14:58	1
Phenol-d6 (Surr)	26		10 - 120	05/23/25 09:56	06/04/25 14:58	1
p-Terphenyl-d14 (Surr)	72		45 - 120	05/23/25 09:56	06/04/25 14:58	1

Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
1H-Indole, 5-methyl-2-phenyl-	6.3	T J N	ug/L		2.56	13228-36-9	05/23/25 09:56	06/04/25 13:28	1
2-Pentenal, (E)-	15	T J N	ug/L		3.15	1576-87-0	05/23/25 09:56	06/04/25 13:28	1
1-Butanol, 2-ethyl-	8.9	T J N	ug/L		3.24	97-95-0	05/23/25 09:56	06/04/25 13:28	1
2-Pentene, 2,4,4-trimethyl-	9.9	T J N	ug/L		3.35	107-40-4	05/23/25 09:56	06/04/25 13:28	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	53		33 - 139	05/23/25 09:56	06/04/25 13:28	1
2-Fluorobiphenyl (Surr)	55		33 - 126	05/23/25 09:56	06/04/25 13:28	1
2-Fluorophenol (Surr)	42		12 - 120	05/23/25 09:56	06/04/25 13:28	1
Nitrobenzene-d5 (Surr)	73		36 - 120	05/23/25 09:56	06/04/25 13:28	1
Phenol-d6 (Surr)	23		10 - 120	05/23/25 09:56	06/04/25 13:28	1
p-Terphenyl-d14 (Surr)	61		47 - 131	05/23/25 09:56	06/04/25 13:28	1

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			05/29/25 14:21	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	80		38 - 134		05/29/25 14:21	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		05/23/25 09:09	05/25/25 08:38	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		05/23/25 09:09	05/25/25 08:38	1
C8-C18	<26		26	ug/L		05/23/25 09:09	05/25/25 08:38	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	83		60 - 130	05/23/25 09:09	05/25/25 08:38	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: TB: MOANALUA WELLS

Lab Sample ID: 380-151743-2

Date Collected: 05/19/25 10:08

Matrix: Water

Date Received: 05/21/25 09:52

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			05/29/25 13:56	1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	80		38 - 134				05/29/25 13:56	1

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)
PWSID Number: HI0000331

Lab Sample ID: 380-151743-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.049		ug/L	2	0.049	525.2	Total/NA
Atrazine	<0.049		ug/L	3	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.59		ug/L	6	0.59	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.59		ug/L	400	0.59	525.2	Total/NA
Endrin	<0.0099		ug/L	2	0.0099	525.2	Total/NA
Heptachlor	<0.0099		ug/L	0.4	0.0099	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Hexachlorobenzene	<0.049		ug/L	1	0.049	525.2	Total/NA
Hexachlorocyclopentadiene	<0.049		ug/L	50	0.049	525.2	Total/NA
Lindane	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Methoxychlor	<0.049		ug/L	40	0.049	525.2	Total/NA
Simazine	<0.049		ug/L	4	0.049	525.2	Total/NA
Benzo[a]pyrene	<0.19		ug/L	0.2	0.19	625.1 SIM	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)						
Lab Sample ID	Client Sample ID	2NMX (70-130)	PRY (70-130)	TPP (70-130)				
380-151743-1	MOANALUA WELLS (331-223-T	92	91	108				
Surrogate Legend								
2NMX = 2-Nitro-m-xylene								
PRY = Perylene-d12								
TPP = Triphenylphosphate								

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)						
Lab Sample ID	Client Sample ID	2NMX (70-130)	PRY (70-130)	TPP (70-130)				
380-151125-W-1-A MS	Matrix Spike	93	96	111				
380-151719-I-1-A DU	Duplicate	90	86	109				
LCS 380-154220/23-A	Lab Control Sample	92	100	110				
MB 380-154220/21-A	Method Blank	92	91	110				
MRL 380-154220/22-A	Lab Control Sample	94	95	106				
Surrogate Legend								
2NMX = 2-Nitro-m-xylene								
PRY = Perylene-d12								
TPP = Triphenylphosphate								

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Drinking Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)		
380-151743-1	MOANALUA WELLS (331-223-T	53	55	42	73	23	61		
Surrogate Legend									
TBP = 2,4,6-Tribromophenol (Surr)									
FBP = 2-Fluorobiphenyl (Surr)									
2FP = 2-Fluorophenol (Surr)									
NBZ = Nitrobenzene-d5 (Surr)									
PHL6 = Phenol-d6 (Surr)									
TPHd14 = p-Terphenyl-d14 (Surr)									

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

		Percent Surrogate Recovery (Acceptance Limits)							
Lab Sample ID	Client Sample ID	TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)		
MB 570-575026/1-A	Method Blank	60	61	48	76	28	67		
Surrogate Legend									
TBP = 2,4,6-Tribromophenol (Surr)									
FBP = 2-Fluorobiphenyl (Surr)									
2FP = 2-Fluorophenol (Surr)									
NBZ = Nitrobenzene-d5 (Surr)									

Eurofins Eaton Analytical Pomona

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-151743-1	MOANALUA WELLS (331-223-T	69	65	41	67	26	72
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	89	81	56	71	38	88
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	82	75	54	67	35	83

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
LCS 570-575026/2-A	Lab Control Sample	95	84	59	71	40	89
LCSD 570-575026/3-A	Lab Control Sample Dup	87	82	59	68	39	85
MB 570-575026/1-A	Method Blank	75	66	43	67	27	76

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Drinking Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		BFB1 (38-134)					
380-151743-1	MOANALUA WELLS (331-223-T	80					
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	85					
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	88					

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)		
Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-151743-2	TB: MOANALUA WELLS	80
LCS 570-576992/4	Lab Control Sample	87
LCSD 570-576992/5	Lab Control Sample Dup	88
MB 570-576992/6	Method Blank	80
MRL 570-576992/3	Lab Control Sample	83
Surrogate Legend		
BFB = 4-Bromofluorobenzene (Surr)		

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Drinking Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)		
Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-151743-1	MOANALUA WELLS (331-223-T	83
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	85
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	99
Surrogate Legend		
OTCSN = n-Octacosane (Surr)		

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)		
Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
LCS 570-574445/2-A	Lab Control Sample	110
LCSD 570-574445/3-A	Lab Control Sample Dup	106
MB 570-574445/1-A	Method Blank	103
MRL 570-574445/4-A	Lab Control Sample	119
Surrogate Legend		
OTCSN = n-Octacosane (Surr)		

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-154220/21-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 154220

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2,4'-DDD	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2,4'-DDE	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2,4'-DDT	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
2-Methylnaphthalene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
4,4'-DDD	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
4,4'-DDE	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
4,4'-DDT	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Acenaphthene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Acenaphthylene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Acetochlor	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Alachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
alpha-BHC	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
alpha-Chlordane	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Anthracene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 12:48	1
Atrazine	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Benz(a)anthracene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Benzo[a]pyrene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 12:48	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 12:48	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 12:48	1
beta-BHC	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		05/23/25 14:41	05/27/25 12:48	1
Bromacil	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Butachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Butylbenzylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 12:48	1
Chlorobenzilate	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Chloroneb	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Chlorpyrifos	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Chrysene	<0.020		0.020	ug/L		05/23/25 14:41	05/27/25 12:48	1
delta-BHC	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		05/23/25 14:41	05/27/25 12:48	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Dieldrin	<0.0098		0.0098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Diethylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 12:48	1
Dimethylphthalate	<0.49		0.49	ug/L		05/23/25 14:41	05/27/25 12:48	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		05/23/25 14:41	05/27/25 12:48	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Endosulfan sulfate	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Endrin	<0.0098		0.0098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Endrin aldehyde	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
EPTC	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-154220/21-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 154220

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluoranthene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Fluorene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
gamma-Chlordane	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Heptachlor	<0.0098		0.0098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Heptachlor epoxide (isomer B)	<0.0098		0.0098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Hexachlorobenzene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Isophorone	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Lindane	<0.0098		0.0098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Malathion	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Methoxychlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Metolachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Molinate	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Naphthalene	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Parathion	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Phenanthrene	<0.039		0.039	ug/L		05/23/25 14:41	05/27/25 12:48	1
Propachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Pyrene	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Simazine	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Terbacil	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Terbutylazine	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Thiobencarb	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		05/23/25 14:41	05/27/25 12:48	1
trans-Nonachlor	<0.049		0.049	ug/L		05/23/25 14:41	05/27/25 12:48	1
Trifluralin	<0.098		0.098	ug/L		05/23/25 14:41	05/27/25 12:48	1

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Unknown	1.09	T J	ug/L		2.53	N/A	05/23/25 14:41	05/27/25 12:48	1
Decane	2.60	T J N	ug/L		2.79	124-18-5	05/23/25 14:41	05/27/25 12:48	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	92		70 - 130	05/23/25 14:41	05/27/25 12:48	1
Perylene-d12	91		70 - 130	05/23/25 14:41	05/27/25 12:48	1
Triphenylphosphate	110		70 - 130	05/23/25 14:41	05/27/25 12:48	1

Lab Sample ID: LCS 380-154220/23-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	1.96	1.98		ug/L		101	70 - 130
2,4'-DDD	1.96	2.06		ug/L		105	70 - 130
2,4'-DDE	1.96	2.19		ug/L		112	70 - 130
2,4'-DDT	1.96	1.97		ug/L		100	70 - 130
2,4-Dinitrotoluene	1.96	1.96		ug/L		100	70 - 130
2,6-Dinitrotoluene	1.96	2.27		ug/L		116	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-154220/23-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
2-Methylnaphthalene	1.96	2.00		ug/L		102	70 - 130
4,4'-DDD	1.96	1.98		ug/L		101	70 - 130
4,4'-DDE	1.96	2.34		ug/L		119	70 - 130
4,4'-DDT	1.96	2.10		ug/L		107	70 - 130
Acenaphthene	1.96	2.06		ug/L		105	70 - 130
Acenaphthylene	1.96	2.06		ug/L		105	70 - 130
Acetochlor	1.96	1.99		ug/L		101	70 - 130
Alachlor	1.96	2.24		ug/L		114	70 - 130
alpha-BHC	1.96	2.07		ug/L		105	70 - 130
alpha-Chlordane	1.96	2.21		ug/L		113	70 - 130
Anthracene	1.96	1.95		ug/L		99	70 - 130
Atrazine	1.96	2.04		ug/L		104	70 - 130
Benz(a)anthracene	1.96	2.12		ug/L		108	70 - 130
Benzo[a]pyrene	1.96	2.26		ug/L		115	70 - 130
Benzo[b]fluoranthene	1.96	2.29		ug/L		116	70 - 130
Benzo[g,h,i]perylene	1.96	2.18		ug/L		111	70 - 130
Benzo[k]fluoranthene	1.96	2.42		ug/L		123	70 - 130
beta-BHC	1.96	2.01		ug/L		102	70 - 130
Bis(2-ethylhexyl) phthalate	1.96	2.10		ug/L		107	70 - 130
Bromacil	1.96	2.03		ug/L		104	70 - 130
Butachlor	1.96	2.05		ug/L		104	70 - 130
Butylbenzylphthalate	1.96	2.16		ug/L		110	70 - 130
Chlorobenzilate	1.96	2.01		ug/L		102	70 - 130
Chloroneb	1.96	2.00		ug/L		102	70 - 130
Chlorothalonil (Draconil, Bravo)	1.96	2.11		ug/L		107	70 - 130
Chlorpyrifos	1.96	2.23		ug/L		114	70 - 130
Chrysene	1.96	2.18		ug/L		111	70 - 130
delta-BHC	1.96	2.07		ug/L		106	70 - 130
Di(2-ethylhexyl)adipate	1.96	2.16		ug/L		110	70 - 130
Dibenz(a,h)anthracene	1.96	2.15		ug/L		109	70 - 130
Diclorvos (DDVP)	1.96	1.99		ug/L		101	70 - 130
Dieldrin	1.96	2.13		ug/L		108	70 - 130
Diethylphthalate	1.96	2.09		ug/L		106	70 - 130
Dimethylphthalate	1.96	2.16		ug/L		110	70 - 130
Di-n-butyl phthalate	3.93	3.94		ug/L		100	70 - 130
Di-n-octyl phthalate	1.96	1.98		ug/L		101	70 - 130
Endosulfan I (Alpha)	1.96	1.97		ug/L		100	70 - 130
Endosulfan II (Beta)	1.96	2.06		ug/L		105	70 - 130
Endosulfan sulfate	1.96	2.12		ug/L		108	70 - 130
Endrin	1.96	2.28		ug/L		116	70 - 130
Endrin aldehyde	1.96	2.34		ug/L		119	60 - 130
EPTC	1.96	2.15		ug/L		110	70 - 130
Fluoranthene	1.96	2.18		ug/L		111	70 - 130
Fluorene	1.96	2.20		ug/L		112	70 - 130
gamma-Chlordane	1.96	2.18		ug/L		111	70 - 130
Heptachlor	1.96	2.25		ug/L		115	70 - 130
Heptachlor epoxide (isomer B)	1.96	2.37		ug/L		120	70 - 130
Hexachlorobenzene	1.96	2.09		ug/L		107	70 - 130
Hexachlorocyclopentadiene	1.96	2.08		ug/L		106	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-154220/23-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Indeno[1,2,3-cd]pyrene	1.96	2.18		ug/L		111	70 - 130
Isophorone	1.96	1.94		ug/L		99	70 - 130
Lindane	1.96	2.14		ug/L		109	70 - 130
Malathion	1.96	2.05		ug/L		104	70 - 130
Methoxychlor	1.96	2.10		ug/L		107	70 - 130
Metolachlor	1.96	1.98		ug/L		101	70 - 130
Molinate	1.96	2.12		ug/L		108	70 - 130
Naphthalene	1.96	1.93		ug/L		98	70 - 130
Parathion	1.96	2.07		ug/L		106	70 - 130
Pendimethalin (Penoxaline)	1.96	2.05		ug/L		104	70 - 130
Phenanthrene	1.96	2.02		ug/L		103	70 - 130
Propachlor	1.96	2.11		ug/L		107	70 - 130
Pyrene	1.96	2.08		ug/L		106	70 - 130
Simazine	1.96	1.95		ug/L		99	70 - 130
Terbacil	1.96	2.31		ug/L		118	70 - 130
Terbutylazine	1.96	2.02		ug/L		103	70 - 130
Thiobencarb	1.96	1.95		ug/L		100	70 - 130
trans-Nonachlor	1.96	2.20		ug/L		112	70 - 130
Trifluralin	1.96	1.88		ug/L		96	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2-Nitro-m-xylene	92		70 - 130
Perylene-d12	100		70 - 130
Triphenylphosphate	110		70 - 130

Lab Sample ID: MRL 380-154220/22-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0986	0.108		ug/L		109	50 - 150
2,4'-DDD	0.0986	0.0887	J	ug/L		90	50 - 150
2,4'-DDE	0.0986	0.104		ug/L		105	50 - 150
2,4'-DDT	0.0986	0.115		ug/L		116	50 - 150
2,4-Dinitrotoluene	0.0986	0.117		ug/L		118	50 - 150
2,6-Dinitrotoluene	0.0986	0.127		ug/L		129	50 - 150
2-Methylnaphthalene	0.0986	0.104		ug/L		105	50 - 150
4,4'-DDD	0.0986	0.110		ug/L		111	50 - 150
4,4'-DDE	0.0986	0.103		ug/L		105	50 - 150
4,4'-DDT	0.0986	0.117		ug/L		119	50 - 150
Acenaphthene	0.0986	0.0963	J	ug/L		98	50 - 150
Acenaphthylene	0.0986	0.0934	J	ug/L		95	50 - 150
Acetochlor	0.0986	0.120		ug/L		122	50 - 150
Alachlor	0.0493	0.0568		ug/L		115	50 - 150
alpha-BHC	0.0986	0.100		ug/L		102	50 - 150
alpha-Chlordane	0.0247	0.0318	J	ug/L		129	50 - 150
Anthracene	0.0197	0.0219		ug/L		111	50 - 150
Atrazine	0.0493	0.0646		ug/L		131	50 - 150

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-154220/22-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Benz(a)anthracene	0.0493	0.0555		ug/L		113	50 - 150
Benzo[a]pyrene	0.0197	0.0216		ug/L		110	50 - 150
Benzo[b]fluoranthene	0.0197	0.0215		ug/L		109	50 - 150
Benzo[g,h,i]perylene	0.0493	0.0540		ug/L		109	50 - 150
Benzo[k]fluoranthene	0.0197	0.0224		ug/L		114	50 - 150
beta-BHC	0.0986	0.112		ug/L		114	50 - 150
Bis(2-ethylhexyl) phthalate	0.592	0.580	J	ug/L		98	50 - 150
Bromacil	0.0986	0.113		ug/L		115	50 - 150
Butachlor	0.0493	0.0692		ug/L		140	50 - 150
Butylbenzylphthalate	0.493	0.555		ug/L		112	50 - 150
Chlorobenzilate	0.0986	0.109		ug/L		110	50 - 150
Chloroneb	0.0986	0.105		ug/L		106	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0986	0.102		ug/L		103	50 - 150
Chlorpyrifos	0.0493	0.0505		ug/L		102	50 - 150
Chrysene	0.0197	0.0233		ug/L		118	50 - 150
delta-BHC	0.0986	0.103		ug/L		104	50 - 150
Di(2-ethylhexyl)adipate	0.592	0.606		ug/L		102	50 - 150
Dibenz(a,h)anthracene	0.0493	0.0546		ug/L		111	50 - 150
Diclorvos (DDVP)	0.0493	0.0553		ug/L		112	50 - 150
Dieldrin	0.00986	0.0131		ug/L		133	50 - 150
Diethylphthalate	0.493	0.530		ug/L		108	50 - 150
Dimethylphthalate	0.493	0.545		ug/L		111	50 - 150
Di-n-butyl phthalate	0.493	0.443	J	ug/L		90	49 - 243
Di-n-octyl phthalate	0.0986	0.0957	J	ug/L		97	50 - 150
Endosulfan I (Alpha)	0.0986	0.0928	J	ug/L		94	50 - 150
Endosulfan II (Beta)	0.0986	0.103		ug/L		104	50 - 150
Endosulfan sulfate	0.0986	0.113		ug/L		115	50 - 150
Endrin	0.00986	0.0120		ug/L		122	50 - 150
Endrin aldehyde	0.0986	0.126		ug/L		128	50 - 150
EPTC	0.0986	0.100		ug/L		101	50 - 150
Fluoranthene	0.0986	0.104		ug/L		105	50 - 150
Fluorene	0.0493	0.0545		ug/L		111	50 - 150
gamma-Chlordane	0.0247	0.0332	J	ug/L		135	50 - 150
Heptachlor	0.00986	0.0115		ug/L		117	50 - 150
Heptachlor epoxide (isomer B)	0.00986	0.0129		ug/L		131	50 - 150
Hexachlorobenzene	0.0493	0.0509		ug/L		103	50 - 150
Hexachlorocyclopentadiene	0.0493	0.0396	J	ug/L		80	50 - 150
Indeno[1,2,3-cd]pyrene	0.0493	0.0608		ug/L		123	50 - 150
Isophorone	0.0986	0.116		ug/L		117	50 - 150
Lindane	0.00986	0.0123		ug/L		125	50 - 150
Malathion	0.0986	0.114		ug/L		116	50 - 150
Methoxychlor	0.0493	0.0735		ug/L		149	50 - 150
Metolachlor	0.0493	0.0623		ug/L		126	50 - 150
Molinate	0.0986	0.104		ug/L		106	50 - 150
Naphthalene	0.0986	0.102		ug/L		104	50 - 150
Parathion	0.0986	0.0974	J	ug/L		99	50 - 150
Pendimethalin (Penoxaline)	0.0986	0.107		ug/L		108	50 - 150
Phenanthrene	0.0394	0.0458		ug/L		116	50 - 150
Propachlor	0.0493	0.0546		ug/L		111	50 - 150

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-154220/22-A

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
Pyrene	0.0493	0.0595		ug/L		121	50 - 150
Simazine	0.0493	0.0637		ug/L		129	50 - 150
Terbacil	0.0986	0.120		ug/L		122	50 - 150
Terbutylazine	0.0986	0.109		ug/L		111	50 - 150
Thiobencarb	0.0986	0.106		ug/L		107	50 - 150
trans-Nonachlor	0.0247	0.0309	J	ug/L		125	50 - 150
Trifluralin	0.0986	0.109		ug/L		111	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
2-Nitro-m-xylene	94		70 - 130
Perylene-d12	95		70 - 130
Triphenylphosphate	106		70 - 130

Lab Sample ID: 380-151125-W-1-A MS

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.10		1.98	2.01		ug/L		102	70 - 130
2,4'-DDD	<0.10		1.98	2.10		ug/L		106	70 - 130
2,4'-DDE	<0.10		1.98	2.18		ug/L		110	70 - 130
2,4'-DDT	<0.10		1.98	2.02		ug/L		102	70 - 130
2,4-Dinitrotoluene	<0.10		1.98	2.01		ug/L		101	70 - 130
2,6-Dinitrotoluene	<0.10		1.98	2.29		ug/L		116	70 - 130
2-Methylnaphthalene	<0.10		1.98	2.04		ug/L		103	70 - 130
4,4'-DDD	<0.10		1.98	2.00		ug/L		101	70 - 130
4,4'-DDE	<0.10		1.98	2.36		ug/L		119	70 - 130
4,4'-DDT	<0.10		1.98	2.11		ug/L		107	70 - 130
Acenaphthene	<0.10		1.98	2.08		ug/L		105	70 - 130
Acenaphthylene	<0.10		1.98	2.04		ug/L		103	70 - 130
Acetochlor	<0.10		1.98	2.01		ug/L		102	70 - 130
Alachlor	<0.050		1.98	2.30		ug/L		116	70 - 130
alpha-BHC	<0.10		1.98	2.08		ug/L		105	70 - 130
alpha-Chlordane	<0.050		1.98	2.21		ug/L		111	70 - 130
Anthracene	<0.020		1.98	1.43		ug/L		72	70 - 130
Atrazine	<0.050		1.98	2.11		ug/L		107	70 - 130
Benz(a)anthracene	<0.050		1.98	2.03		ug/L		103	70 - 130
Benzo[a]pyrene	<0.020		1.98	2.03		ug/L		103	70 - 130
Benzo[b]fluoranthene	<0.020		1.98	2.35		ug/L		119	70 - 130
Benzo[g,h,i]perylene	<0.050		1.98	2.18		ug/L		110	70 - 130
Benzo[k]fluoranthene	<0.020		1.98	2.31		ug/L		117	70 - 130
beta-BHC	<0.10		1.98	2.10		ug/L		106	70 - 130
Bis(2-ethylhexyl) phthalate	<0.60		1.98	2.05		ug/L		103	70 - 130
Bromacil	<0.10		1.98	2.04		ug/L		103	70 - 130
Butachlor	<0.050		1.98	2.10		ug/L		106	70 - 130
Butylbenzylphthalate	<0.50		1.98	2.25		ug/L		114	70 - 130
Chlorobenzilate	<0.10		1.98	2.11		ug/L		106	70 - 130
Chloroneb	<0.10		1.98	2.01		ug/L		101	70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-151125-W-1-A MS

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
Chlorothalonil (Draconil, Bravo)	<0.10		1.98	2.14		ug/L		108	70 - 130
Chlorpyrifos	<0.050		1.98	2.26		ug/L		114	70 - 130
Chrysene	<0.020		1.98	2.15		ug/L		108	70 - 130
delta-BHC	<0.10		1.98	2.09		ug/L		106	70 - 130
Di(2-ethylhexyl)adipate	<0.60		1.98	2.15		ug/L		108	70 - 130
Dibenz(a,h)anthracene	<0.050		1.98	2.16		ug/L		109	70 - 130
Diclorvos (DDVP)	<0.050		1.98	2.00		ug/L		101	70 - 130
Dieldrin	<0.010		1.98	2.21		ug/L		112	70 - 130
Diethylphthalate	<0.50		1.98	2.16		ug/L		109	70 - 130
Dimethylphthalate	<0.50		1.98	2.20		ug/L		111	70 - 130
Di-n-butyl phthalate	<1.0		3.96	3.91		ug/L		99	70 - 130
Di-n-octyl phthalate	<0.10		1.98	1.95		ug/L		99	70 - 130
Endosulfan I (Alpha)	<0.10		1.98	2.06		ug/L		104	70 - 130
Endosulfan II (Beta)	<0.10		1.98	2.13		ug/L		108	70 - 130
Endosulfan sulfate	<0.10		1.98	2.20		ug/L		111	70 - 130
Endrin	<0.010		1.98	2.31		ug/L		116	70 - 130
Endrin aldehyde	<0.10		1.98	1.74		ug/L		88	60 - 130
EPTC	<0.10		1.98	2.25		ug/L		113	70 - 130
Fluoranthene	<0.10		1.98	2.16		ug/L		109	70 - 130
Fluorene	<0.050		1.98	2.19		ug/L		111	70 - 130
gamma-Chlordane	<0.050		1.98	2.22		ug/L		112	70 - 130
Heptachlor	<0.010		1.98	2.23		ug/L		113	70 - 130
Heptachlor epoxide (isomer B)	<0.010		1.98	2.40		ug/L		121	70 - 130
Hexachlorobenzene	<0.050		1.98	2.10		ug/L		106	70 - 130
Hexachlorocyclopentadiene	<0.050		1.98	2.05		ug/L		103	70 - 130
Indeno[1,2,3-cd]pyrene	<0.050		1.98	2.22		ug/L		112	70 - 130
Isophorone	<0.10		1.98	1.98		ug/L		100	70 - 130
Lindane	<0.010		1.98	2.19		ug/L		111	70 - 130
Malathion	<0.10		1.98	2.13		ug/L		108	70 - 130
Methoxychlor	<0.050		1.98	2.09		ug/L		106	70 - 130
Metolachlor	<0.050		1.98	2.04		ug/L		103	70 - 130
Molinate	<0.10		1.98	2.19		ug/L		110	70 - 130
Naphthalene	<0.10		1.98	1.97		ug/L		100	70 - 130
Parathion	<0.10		1.98	2.13		ug/L		107	70 - 130
Pendimethalin (Penoxaline)	<0.10		1.98	2.07		ug/L		105	70 - 130
Phenanthrene	<0.040		1.98	2.04		ug/L		103	70 - 130
Propachlor	<0.050		1.98	2.19		ug/L		111	70 - 130
Pyrene	<0.050		1.98	2.06		ug/L		104	70 - 130
Simazine	<0.050		1.98	2.00		ug/L		101	70 - 130
Terbacil	<0.10		1.98	2.31		ug/L		117	70 - 130
Terbutylazine	<0.10		1.98	2.13		ug/L		107	70 - 130
Thiobencarb	<0.10		1.98	1.98		ug/L		100	70 - 130
trans-Nonachlor	<0.050		1.98	2.29		ug/L		116	70 - 130
Trifluralin	<0.10		1.98	1.99		ug/L		100	70 - 130

Surrogate	MS %Recovery	MS Qualifier	Limits
2-Nitro-m-xylene	93		70 - 130
Perylene-d12	96		70 - 130

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-151125-W-1-A MS

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 154220

Surrogate	MS %Recovery	MS Qualifier	Limits
Triphenylphosphate	111		70 - 130

Lab Sample ID: 380-151719-I-1-A DU

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	RPD Limit
1-Methylnaphthalene	<0.098		<0.097		ug/L		NC	20
2,4'-DDD	<0.098		<0.097		ug/L		NC	20
2,4'-DDE	<0.098		<0.097		ug/L		NC	20
2,4'-DDT	<0.098		<0.097		ug/L		NC	20
2,4-Dinitrotoluene	<0.098		<0.097		ug/L		NC	20
2,6-Dinitrotoluene	<0.098		<0.097		ug/L		NC	20
2-Methylnaphthalene	<0.098		<0.097		ug/L		NC	20
4,4'-DDD	<0.098		<0.097		ug/L		NC	20
4,4'-DDE	<0.098		<0.097		ug/L		NC	20
4,4'-DDT	<0.098		<0.097		ug/L		NC	20
Acenaphthene	<0.098		<0.097		ug/L		NC	20
Acenaphthylene	<0.098		<0.097		ug/L		NC	20
Acetochlor	<0.098		<0.097		ug/L		NC	20
Alachlor	<0.049		<0.049		ug/L		NC	20
alpha-BHC	<0.098		<0.097		ug/L		NC	20
alpha-Chlordane	<0.049		<0.049		ug/L		NC	20
Anthracene	<0.020		<0.019		ug/L		NC	20
Atrazine	<0.049		<0.049		ug/L		NC	20
Benz(a)anthracene	<0.049		<0.049		ug/L		NC	20
Benzo[a]pyrene	<0.020		<0.019		ug/L		NC	20
Benzo[b]fluoranthene	<0.020		<0.019		ug/L		NC	20
Benzo[g,h,i]perylene	<0.049		<0.049		ug/L		NC	20
Benzo[k]fluoranthene	<0.020		<0.019		ug/L		NC	20
beta-BHC	<0.098		<0.097		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	<0.59		<0.58		ug/L		NC	20
Bromacil	<0.098		<0.097		ug/L		NC	20
Butachlor	<0.049		<0.049		ug/L		NC	20
Butylbenzylphthalate	<0.49		<0.49		ug/L		NC	20
Chlorobenzilate	<0.098		<0.097		ug/L		NC	20
Chloroneb	<0.098		<0.097		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	<0.098		<0.097		ug/L		NC	20
Chlorpyrifos	<0.049		<0.049		ug/L		NC	20
Chrysene	<0.020		<0.019		ug/L		NC	20
delta-BHC	<0.098		<0.097		ug/L		NC	20
Di(2-ethylhexyl)adipate	<0.59		<0.58		ug/L		NC	20
Dibenz(a,h)anthracene	<0.049		<0.049		ug/L		NC	20
Diclorvos (DDVP)	<0.049		<0.049		ug/L		NC	20
Dieldrin	0.031		0.0397	F5	ug/L		24	20
Diethylphthalate	<0.49		<0.49		ug/L		NC	20
Dimethylphthalate	<0.49		<0.49		ug/L		NC	20
Di-n-butyl phthalate	<0.98		<0.97		ug/L		NC	20

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-151719-I-1-A DU

Matrix: Water

Analysis Batch: 154425

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 154220

Analyte	Sample Result	Sample Qualifier	DU Result	DU Qualifier	Unit	D	RPD	Limit
Di-n-octyl phthalate	<0.098		<0.097		ug/L		NC	20
Endosulfan I (Alpha)	<0.098		<0.097		ug/L		NC	20
Endosulfan II (Beta)	<0.098		<0.097		ug/L		NC	20
Endosulfan sulfate	<0.098		<0.097		ug/L		NC	20
Endrin	<0.0098		<0.0097		ug/L		NC	20
Endrin aldehyde	<0.098		<0.097		ug/L		NC	20
EPTC	<0.098		<0.097		ug/L		NC	20
Fluoranthene	<0.098		<0.097		ug/L		NC	20
Fluorene	<0.049		<0.049		ug/L		NC	20
gamma-Chlordane	<0.049		<0.049		ug/L		NC	20
Heptachlor	<0.0098		<0.0097		ug/L		NC	20
Heptachlor epoxide (isomer B)	0.014		0.0141		ug/L		2	20
Hexachlorobenzene	<0.049		<0.049		ug/L		NC	20
Hexachlorocyclopentadiene	<0.049		<0.049		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	<0.049		<0.049		ug/L		NC	20
Isophorone	<0.098		<0.097		ug/L		NC	20
Lindane	<0.0098		<0.0097		ug/L		NC	20
Malathion	<0.098		<0.097		ug/L		NC	20
Methoxychlor	<0.049		<0.049		ug/L		NC	20
Metolachlor	<0.049		<0.049		ug/L		NC	20
Molinate	<0.098		<0.097		ug/L		NC	20
Naphthalene	<0.098		<0.097		ug/L		NC	20
Parathion	<0.098		<0.097		ug/L		NC	20
Pendimethalin (Penoxaline)	<0.098		<0.097		ug/L		NC	20
Phenanthrene	<0.039		<0.039		ug/L		NC	20
Propachlor	<0.049		<0.049		ug/L		NC	20
Pyrene	<0.049		<0.049		ug/L		NC	20
Simazine	<0.049		<0.049		ug/L		NC	20
Terbacil	<0.098		<0.097		ug/L		NC	20
Terbutylazine	<0.098		<0.097		ug/L		NC	20
Thiobencarb	<0.098		<0.097		ug/L		NC	20
Total Permethrin (mixed isomers)	<0.20		<0.19		ug/L		NC	20
trans-Nonachlor	<0.049		<0.049		ug/L		NC	20
Trifluralin	<0.098		<0.097		ug/L		NC	20

Surrogate	DU %Recovery	DU Qualifier	Limits
2-Nitro-m-xylene	90		70 - 130
Perylene-d12	86		70 - 130
Triphenylphosphate	109		70 - 130

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-575026/1-A

Matrix: Water

Analysis Batch: 579411

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 575026

Tentatively Identified Compound	MB Est. Result	MB Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	05/23/25 09:56	06/04/25 13:05	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 625.1 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 570-575026/1-A

Matrix: Water

Analysis Batch: 579411

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 575026

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	60		33 - 139	05/23/25 09:56	06/04/25 13:05	1
2-Fluorobiphenyl (Surr)	61		33 - 126	05/23/25 09:56	06/04/25 13:05	1
2-Fluorophenol (Surr)	48		12 - 120	05/23/25 09:56	06/04/25 13:05	1
Nitrobenzene-d5 (Surr)	76		36 - 120	05/23/25 09:56	06/04/25 13:05	1
Phenol-d6 (Surr)	28		10 - 120	05/23/25 09:56	06/04/25 13:05	1
p-Terphenyl-d14 (Surr)	67		47 - 131	05/23/25 09:56	06/04/25 13:05	1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-575026/1-A

Matrix: Water

Analysis Batch: 579367

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 575026

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
2-Methylnaphthalene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Acenaphthene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Acenaphthylene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Anthracene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Benzo[a]anthracene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Benzo[a]pyrene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Chrysene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Fluoranthene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Fluorene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Naphthalene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Phenanthrene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1
Pyrene	<0.20		0.20	ug/L		05/23/25 09:56	06/04/25 13:06	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	75		28 - 127	05/23/25 09:56	06/04/25 13:06	1
2-Fluorobiphenyl (Surr)	66		31 - 120	05/23/25 09:56	06/04/25 13:06	1
2-Fluorophenol (Surr)	43		17 - 120	05/23/25 09:56	06/04/25 13:06	1
Nitrobenzene-d5 (Surr)	67		27 - 120	05/23/25 09:56	06/04/25 13:06	1
Phenol-d6 (Surr)	27		10 - 120	05/23/25 09:56	06/04/25 13:06	1
p-Terphenyl-d14 (Surr)	76		45 - 120	05/23/25 09:56	06/04/25 13:06	1

Lab Sample ID: LCS 570-575026/2-A

Matrix: Water

Analysis Batch: 579367

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	20.0	18.8		ug/L		94	47 - 120
2-Methylnaphthalene	20.0	20.7		ug/L		103	43 - 120

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCS 570-575026/2-A

Matrix: Water

Analysis Batch: 579367

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Acenaphthene	20.0	21.4		ug/L		107	60 - 132
Acenaphthylene	20.0	20.3		ug/L		101	54 - 126
Anthracene	20.0	22.0		ug/L		110	43 - 120
Benzo[a]anthracene	20.0	22.7		ug/L		114	42 - 133
Benzo[a]pyrene	20.0	23.5		ug/L		118	32 - 148
Benzo[b]fluoranthene	20.0	23.6		ug/L		118	42 - 140
Benzo[g,h,i]perylene	20.0	23.5		ug/L		118	1 - 195
Benzo[k]fluoranthene	20.0	23.6		ug/L		118	25 - 146
Chrysene	20.0	21.9		ug/L		109	44 - 140
Dibenz(a,h)anthracene	20.0	24.1		ug/L		121	1 - 200
Fluoranthene	20.0	21.9		ug/L		109	43 - 121
Fluorene	20.0	21.5		ug/L		108	70 - 120
Indeno[1,2,3-cd]pyrene	20.0	24.0		ug/L		120	1 - 151
Naphthalene	20.0	18.2		ug/L		91	36 - 120
Phenanthrene	20.0	20.9		ug/L		105	65 - 120
Pyrene	20.0	22.1		ug/L		110	70 - 120

Surrogate	LCS %Recovery	LCS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	95		28 - 127
2-Fluorobiphenyl (Surr)	84		31 - 120
2-Fluorophenol (Surr)	59		17 - 120
Nitrobenzene-d5 (Surr)	71		27 - 120
Phenol-d6 (Surr)	40		10 - 120
p-Terphenyl-d14 (Surr)	89		45 - 120

Lab Sample ID: LCSD 570-575026/3-A

Matrix: Water

Analysis Batch: 579367

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
1-Methylnaphthalene	20.0	18.7		ug/L		93	47 - 120	1	20
2-Methylnaphthalene	20.0	20.3		ug/L		102	43 - 120	2	20
Acenaphthene	20.0	20.9		ug/L		104	60 - 132	3	29
Acenaphthylene	20.0	19.4		ug/L		97	54 - 126	4	45
Anthracene	20.0	21.7		ug/L		108	43 - 120	1	40
Benzo[a]anthracene	20.0	22.1		ug/L		111	42 - 133	3	32
Benzo[a]pyrene	20.0	22.3		ug/L		112	32 - 148	5	43
Benzo[b]fluoranthene	20.0	22.7		ug/L		114	42 - 140	4	43
Benzo[g,h,i]perylene	20.0	21.6		ug/L		108	1 - 195	8	61
Benzo[k]fluoranthene	20.0	22.0		ug/L		110	25 - 146	7	38
Chrysene	20.0	21.4		ug/L		107	44 - 140	2	53
Dibenz(a,h)anthracene	20.0	22.3		ug/L		112	1 - 200	8	75
Fluoranthene	20.0	22.2		ug/L		111	43 - 121	2	40
Fluorene	20.0	20.9		ug/L		105	70 - 120	3	23
Indeno[1,2,3-cd]pyrene	20.0	22.6		ug/L		113	1 - 151	6	60
Naphthalene	20.0	18.1		ug/L		90	36 - 120	1	39
Phenanthrene	20.0	20.8		ug/L		104	65 - 120	1	24
Pyrene	20.0	21.3		ug/L		107	70 - 120	3	30

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCSD 570-575026/3-A

Matrix: Water

Analysis Batch: 579367

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 575026

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	87		28 - 127
2-Fluorobiphenyl (Surr)	82		31 - 120
2-Fluorophenol (Surr)	59		17 - 120
Nitrobenzene-d5 (Surr)	68		27 - 120
Phenol-d6 (Surr)	39		10 - 120
p-Terphenyl-d14 (Surr)	85		45 - 120

Lab Sample ID: 380-151743-1 MS

Matrix: Drinking Water

Analysis Batch: 579367

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.19		19.5	19.0		ug/L		98	36 - 120
2-Methylnaphthalene	<0.19		19.5	20.5		ug/L		105	32 - 124
Acenaphthene	<0.19		19.5	20.3		ug/L		104	47 - 145
Acenaphthylene	<0.19		19.5	18.8		ug/L		96	33 - 145
Anthracene	<0.19		19.5	21.4		ug/L		110	27 - 133
Benzo[a]anthracene	<0.19		19.5	21.4		ug/L		110	33 - 143
Benzo[a]pyrene	<0.19		19.5	22.3		ug/L		115	17 - 163
Benzo[b]fluoranthene	<0.19		19.5	22.5		ug/L		115	24 - 159
Benzo[g,h,i]perylene	<0.19		19.5	21.7		ug/L		111	1 - 219
Benzo[k]fluoranthene	<0.19		19.5	22.3		ug/L		114	11 - 162
Chrysene	<0.19		19.5	20.6		ug/L		106	17 - 168
Dibenz(a,h)anthracene	<0.19		19.5	22.2		ug/L		114	1 - 227
Fluoranthene	<0.19		19.5	21.7		ug/L		111	26 - 137
Fluorene	<0.19		19.5	20.3		ug/L		104	59 - 121
Indeno[1,2,3-cd]pyrene	<0.19		19.5	21.9		ug/L		112	1 - 171
Naphthalene	<0.19		19.5	18.3		ug/L		94	21 - 133
Phenanthrene	<0.19		19.5	20.5		ug/L		105	54 - 120
Pyrene	<0.19		19.5	21.2		ug/L		109	52 - 120

Surrogate	MS %Recovery	MS Qualifier	Limits
2,4,6-Tribromophenol (Surr)	89		28 - 127
2-Fluorobiphenyl (Surr)	81		31 - 120
2-Fluorophenol (Surr)	56		17 - 120
Nitrobenzene-d5 (Surr)	71		27 - 120
Phenol-d6 (Surr)	38		10 - 120
p-Terphenyl-d14 (Surr)	88		45 - 120

Lab Sample ID: 380-151743-1 MSD

Matrix: Drinking Water

Analysis Batch: 579367

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
1-Methylnaphthalene	<0.19		19.7	18.1		ug/L		92	36 - 120	5	30
2-Methylnaphthalene	<0.19		19.7	19.5		ug/L		99	32 - 124	5	30
Acenaphthene	<0.19		19.7	19.0		ug/L		97	47 - 145	6	48
Acenaphthylene	<0.19		19.7	18.0		ug/L		91	33 - 145	5	74

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: 380-151743-1 MSD

Matrix: Drinking Water

Analysis Batch: 579367

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Prep Batch: 575026

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Anthracene	<0.19		19.7	19.5		ug/L		99	27 - 133	9	66
Benzo[a]anthracene	<0.19		19.7	20.7		ug/L		105	33 - 143	3	53
Benzo[a]pyrene	<0.19		19.7	20.4		ug/L		104	17 - 163	9	72
Benzo[b]fluoranthene	<0.19		19.7	20.6		ug/L		105	24 - 159	9	71
Benzo[g,h,i]perylene	<0.19		19.7	21.0		ug/L		107	1 - 219	3	97
Benzo[k]fluoranthene	<0.19		19.7	20.5		ug/L		104	11 - 162	8	63
Chrysene	<0.19		19.7	19.3		ug/L		98	17 - 168	7	87
Dibenz(a,h)anthracene	<0.19		19.7	21.0		ug/L		107	1 - 227	5	126
Fluoranthene	<0.19		19.7	21.4		ug/L		109	26 - 137	1	66
Fluorene	<0.19		19.7	19.6		ug/L		100	59 - 121	3	38
Indeno[1,2,3-cd]pyrene	<0.19		19.7	21.6		ug/L		110	1 - 171	1	99
Naphthalene	<0.19		19.7	17.3		ug/L		88	21 - 133	5	65
Phenanthrene	<0.19		19.7	19.2		ug/L		98	54 - 120	6	39
Pyrene	<0.19		19.7	20.3		ug/L		103	52 - 120	4	49

Surrogate	MSD %Recovery	MSD Qualifier	Limits
2,4,6-Tribromophenol (Surr)	82		28 - 127
2-Fluorobiphenyl (Surr)	75		31 - 120
2-Fluorophenol (Surr)	54		17 - 120
Nitrobenzene-d5 (Surr)	67		27 - 120
Phenol-d6 (Surr)	35		10 - 120
p-Terphenyl-d14 (Surr)	83		45 - 120

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-576992/6

Matrix: Water

Analysis Batch: 576992

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			05/29/25 12:24	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	80		38 - 134		05/29/25 12:24	1

Lab Sample ID: LCS 570-576992/4

Matrix: Water

Analysis Batch: 576992

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Gasoline Range Organics (C4-C13)	400	346		ug/L		87	78 - 120

Surrogate	LCS %Recovery	LCS Qualifier	Limits
4-Bromofluorobenzene (Surr)	87		38 - 134

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Lab Sample ID: LCSD 570-576992/5

Matrix: Water

Analysis Batch: 576992

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Analyte			Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)			400	355		ug/L		89	78 - 120	2	10
Surrogate	LCSD %Recovery	LCSD Qualifier	Limits								
4-Bromofluorobenzene (Surr)	88		38 - 134								

Lab Sample ID: MRL 570-576992/3

Matrix: Water

Analysis Batch: 576992

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte			Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits		
Gasoline Range Organics (C4-C13)			10.0	10.6		ug/L		106	50 - 150		
Surrogate	MRL %Recovery	MRL Qualifier	Limits								
4-Bromofluorobenzene (Surr)	83		38 - 134								

Lab Sample ID: 380-151743-1 MS

Matrix: Drinking Water

Analysis Batch: 576992

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits		
Gasoline Range Organics (C4-C13)	<10		400	392		ug/L		98	68 - 122		
Surrogate	MS %Recovery	MS Qualifier	Limits								
4-Bromofluorobenzene (Surr)	85		38 - 134								

Lab Sample ID: 380-151743-1 MSD

Matrix: Drinking Water

Analysis Batch: 576992

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Gasoline Range Organics (C4-C13)	<10		400	383		ug/L		96	68 - 122	2	18
Surrogate	MSD %Recovery	MSD Qualifier	Limits								
4-Bromofluorobenzene (Surr)	88		38 - 134								

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-574445/1-A

Matrix: Water

Analysis Batch: 575632

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 574445

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<25		25	ug/L		05/22/25 09:19	05/25/25 04:46	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		05/22/25 09:19	05/25/25 04:46	1
C8-C18	<25		25	ug/L		05/22/25 09:19	05/25/25 04:46	1

Eurofins Eaton Analytical Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	103		60 - 130	05/22/25 09:19	05/25/25 04:46	1

Lab Sample ID: LCS 570-574445/2-A
Matrix: Water
Analysis Batch: 575632

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 574445

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	1600	1360		ug/L		85	56 - 127

Surrogate	LCS %Recovery	LCS Qualifier	Limits
n-Octacosane (Surr)	110		60 - 130

Lab Sample ID: LCSD 570-574445/3-A
Matrix: Water
Analysis Batch: 575632

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 574445

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	1600	1600		ug/L		100	56 - 127	16	23

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
n-Octacosane (Surr)	106		60 - 130

Lab Sample ID: MRL 570-574445/4-A
Matrix: Water
Analysis Batch: 575632

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 574445

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	0.0200	0.0307	^3+	mg/L		154	50 - 150

Surrogate	MRL %Recovery	MRL Qualifier	Limits
n-Octacosane (Surr)	119		60 - 130

Lab Sample ID: 380-151743-1 MS
Matrix: Drinking Water
Analysis Batch: 575632

Client Sample ID: MOANALUA WELLS (331-223-TP202)
Prep Type: Total/NA
Prep Batch: 574445

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
C10-C28	<26	^3+	1660	1570		ug/L		95	70 - 130

Surrogate	MS %Recovery	MS Qualifier	Limits
n-Octacosane (Surr)	85		60 - 130

Lab Sample ID: 380-151743-1 MSD
Matrix: Drinking Water
Analysis Batch: 575632

Client Sample ID: MOANALUA WELLS (331-223-TP202)
Prep Type: Total/NA
Prep Batch: 574445

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26	^3+	1650	1860		ug/L		112	70 - 130	17	20

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: 380-151743-1 MSD
Matrix: Drinking Water
Analysis Batch: 575632

Client Sample ID: MOANALUA WELLS (331-223-TP202)
Prep Type: Total/NA
Prep Batch: 574445

	MSD	MSD	
Surrogate	%Recovery	Qualifier	Limits
<i>n</i> -Octacosane (Surr)	99		60 - 130

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

GC/MS Semi VOA

Prep Batch: 154220

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	525.2	
MB 380-154220/21-A	Method Blank	Total/NA	Water	525.2	
LCS 380-154220/23-A	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-154220/22-A	Lab Control Sample	Total/NA	Water	525.2	
380-151125-W-1-A MS	Matrix Spike	Total/NA	Water	525.2	
380-151719-I-1-A DU	Duplicate	Total/NA	Water	525.2	

Analysis Batch: 154425

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	525.2	154220
MB 380-154220/21-A	Method Blank	Total/NA	Water	525.2	154220
LCS 380-154220/23-A	Lab Control Sample	Total/NA	Water	525.2	154220
MRL 380-154220/22-A	Lab Control Sample	Total/NA	Water	525.2	154220
380-151125-W-1-A MS	Matrix Spike	Total/NA	Water	525.2	154220
380-151719-I-1-A DU	Duplicate	Total/NA	Water	525.2	154220

Prep Batch: 575026

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1	
MB 570-575026/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-575026/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-575026/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1	
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1	

Analysis Batch: 579367

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1 SIM	575026
MB 570-575026/1-A	Method Blank	Total/NA	Water	625.1 SIM	575026
LCS 570-575026/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	575026
LCSD 570-575026/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	575026
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1 SIM	575026
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1 SIM	575026

Analysis Batch: 579411

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	625.1	575026
MB 570-575026/1-A	Method Blank	Total/NA	Water	625.1	575026

GC VOA

Analysis Batch: 576992

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B GRO LL	
380-151743-2	TB: MOANALUA WELLS	Total/NA	Water	8015B GRO LL	
MB 570-576992/6	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-576992/4	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-576992/5	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-576992/3	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B GRO LL	
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B GRO LL	

Eurofins Eaton Analytical Pomona

QC Association Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

GC Semi VOA

Prep Batch: 574445

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	3510C	
MB 570-574445/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-574445/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-574445/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-574445/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	3510C	
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	3510C	

Analysis Batch: 575632

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-151743-1	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B	574445
MB 570-574445/1-A	Method Blank	Total/NA	Water	8015B	574445
LCS 570-574445/2-A	Lab Control Sample	Total/NA	Water	8015B	574445
LCSD 570-574445/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	574445
MRL 570-574445/4-A	Lab Control Sample	Total/NA	Water	8015B	574445
380-151743-1 MS	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B	574445
380-151743-1 MSD	MOANALUA WELLS (331-223-TP202)	Total/NA	Drinking Water	8015B	574445

Lab Chronicle

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Client Sample ID: MOANALUA WELLS (331-223-TP202)

Lab Sample ID: 380-151743-1

Date Collected: 05/19/25 10:08

Matrix: Drinking Water

Date Received: 05/21/25 09:52

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			154220	IQ42	EA POM	05/23/25 14:41
Total/NA	Analysis	525.2		1	154425	UPAC	EA POM	05/27/25 14:09
Total/NA	Prep	625.1			575026	OAJ3	EET CAL 4	05/23/25 09:56
Total/NA	Analysis	625.1		1	579411	CG	EET CAL 4	06/04/25 13:28
Total/NA	Prep	625.1			575026	OAJ3	EET CAL 4	05/23/25 09:56
Total/NA	Analysis	625.1 SIM		1	579367	J7WE	EET CAL 4	06/04/25 14:58
Total/NA	Analysis	8015B GRO LL		1	576992	A9VE	EET CAL 4	05/29/25 14:21
Total/NA	Prep	3510C			574445	TVD6	EET CAL 4	05/23/25 09:09
Total/NA	Analysis	8015B		1	575632	H6FE	EET CAL 4	05/25/25 08:38

Client Sample ID: TB: MOANALUA WELLS

Lab Sample ID: 380-151743-2

Date Collected: 05/19/25 10:08

Matrix: Water

Date Received: 05/21/25 09:52

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	576992	A9VE	EET CAL 4	05/29/25 13:56

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Laboratory: Eurofins Eaton Analytical Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26
The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.			
Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Drinking Water	1-Methylnaphthalene
525.2	525.2	Drinking Water	2,4'-DDD
525.2	525.2	Drinking Water	2,4'-DDE
525.2	525.2	Drinking Water	2,4'-DDT
525.2	525.2	Drinking Water	2,4-Dinitrotoluene
525.2	525.2	Drinking Water	2,6-Dinitrotoluene
525.2	525.2	Drinking Water	2-Methylnaphthalene
525.2	525.2	Drinking Water	4,4'-DDD
525.2	525.2	Drinking Water	4,4'-DDE
525.2	525.2	Drinking Water	4,4' DDT
525.2	525.2	Drinking Water	Acetochlor
525.2	525.2	Drinking Water	alpha-BHC
525.2	525.2	Drinking Water	alpha-Chlordane
525.2	525.2	Drinking Water	beta-BHC
525.2	525.2	Drinking Water	Chlorobenzilate
525.2	525.2	Drinking Water	Chloroneb
525.2	525.2	Drinking Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Drinking Water	Chlorpyrifos
525.2	525.2	Drinking Water	delta-BHC
525.2	525.2	Drinking Water	Diclorvos (DDVP)
525.2	525.2	Drinking Water	Endosulfan I (Alpha)
525.2	525.2	Drinking Water	Endosulfan II (Beta)
525.2	525.2	Drinking Water	Endosulfan sulfate
525.2	525.2	Drinking Water	Endrin aldehyde
525.2	525.2	Drinking Water	EPTC
525.2	525.2	Drinking Water	gamma-Chlordane
525.2	525.2	Drinking Water	Isophorone
525.2	525.2	Drinking Water	Malathion
525.2	525.2	Drinking Water	Parathion
525.2	525.2	Drinking Water	Pendimethalin (Penoxaline)
525.2	525.2	Drinking Water	Terbacil
525.2	525.2	Drinking Water	Terbutylazine
525.2	525.2	Drinking Water	Total Permethrin (mixed isomers)
525.2	525.2	Drinking Water	trans-Nonachlor

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
Arizona	State	AZ0830	11-16-25
Arkansas DEQ	State	88-01672	07-02-25
California	Los Angeles County Sanitation Districts	9257304	07-31-26
California	SCAQMD LAP	17LA0919	11-30-25
California	State	3082	07-31-25

Eurofins Eaton Analytical Pomona

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
Kansas	NELAP	E-10420	07-31-25
Nevada	State	CA00111	07-31-25
Oregon	NELAP	4175	02-02-26
USDA	US Federal Programs	525-23-159-97150	06-08-26
Utah	NELAP	CA00111	02-28-26
Washington	State	C916	10-11-25

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EA POM = Eurofins Eaton Analytical Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-151743-1
SDG: Weekly

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	PWSID Number
380-151743-1	MOANALUA WELLS (331-223-TP202)	Drinking Water	05/19/25 10:08	05/21/25 09:52	HI0000331
380-151743-2	TB: MOANALUA WELLS	Water	05/19/25 10:08	05/21/25 09:52	

[illegible]

941 Corporate Center Drive
Pomona, CA 91768-2642
Phone: 626-386-1100

Chain of Custody Record



eurofins

Loc: 380
151743

[illegible]

Note: Since laboratory accreditations are subject to change, Eurofins Eaton Analytical, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/test/matrix being analyzed, the samples must be shipped back to the Eurofins Eaton Analytical, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Eaton Analytical, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Eaton Analytical, LLC.

Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)	
Unconfirmed		☐ Return To Client ☐ Disposal By Lab ☐ Archive For _____ Months	
Deliverable Requested: I, II, III, IV, Other (specify)		Special Instructions/QC Requirements:	
Empty Kit Relinquished by:		Method of Shipment:	
Relinquished by: <i>Xm</i>	Date/Time: <i>5/22/25 1300</i>	Company: <i>EGA</i>	Received by: <i>bl</i>
Relinquished by:	Date/Time:	Company:	Received by:
Relinquished by:	Date/Time:	Company:	Received by:
Custody Seals Intact: Δ Yes Δ No	Custody Seal No.:	Cooler Temperature(s) °C and Other Remarks: <i>1.1/2.1 In92</i>	

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-151743-1

SDG Number: Weekly

Login Number: 151743

List Source: Eurofins Eaton Analytical Pomona

List Number: 1

Creator: Hernandez, Orlando

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-151743-1

SDG Number: Weekly

Login Number: 151743

List Number: 2

Creator: Khana, Piyush

List Source: Eurofins Calscience

List Creation: 05/22/25 02:00 PM

Question	Answer	Comment
Radioactivity wasn't checked or is $\leq$ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	2.1
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is $<6\text{mm}$ (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	