

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Erwin Kawata
City & County of Honolulu
630 South Beretania Street
Public Service Bldg. Room 310
Honolulu, Hawaii 96843

Generated 1/14/2026 9:27:49 PM

JOB DESCRIPTION

RED-HILL
Weekly: Halawa Shaft Viewing Pool

JOB NUMBER

380-190084-1

Eurofins Pomona

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

The test results in this report relate only to the samples as received by the laboratory and meet all requirements of the methodology, with any exceptions noted. This report shall not be reproduced except in full, without the express written approval of the laboratory. All questions should be directed to the Eurofins Drinking Water and Wastewater West, LLC Project Manager.

Compliance Statement

1. Laboratory is accredited in accordance with TNI 2016 Standards and ISO/IEC 17025:2017.
2. Laboratory certifies that the test results meet all TNI 2016 and ISO/IEC 17025:2017 requirements unless noted under the individual analysis
3. Test results relate only to the sample(s) tested.
4. This report shall not be reproduced except in full, without the written approval of the laboratory.
5. Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below. (DW, Water matrices)

Authorization



Authorized for release by
Maria Lopez, Project Manager
Maria.Lopez@et.eurofinsus.com
(626)386-1100

Generated
1/14/2026 9:27:49 PM



Table of Contents

Cover Page	1
Table of Contents	3
Definitions/Glossary	4
Case Narrative	6
Detection Summary	7
Client Sample Results	8
Action Limit Summary	12
Surrogate Summary	13
QC Sample Results	15
QC Association Summary	31
Lab Chronicle	33
Certification Summary	34
Method Summary	36
Sample Summary	37
Chain of Custody	38
Receipt Checklists	40

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Qualifiers

GC/MS Semi VOA

Qualifier	Qualifier Description
*-	LCS and/or LCSD is outside acceptance limits, low biased.
*1	LCS/LCSD RPD exceeds control limits.
^3+	Reporting Limit Check Standard is outside acceptance limits, high biased
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
S1-	Surrogate recovery exceeds control limits, low biased.
S1+	Surrogate recovery exceeds control limits, high biased.

GC/MS Semi VOA TICs

Qualifier	Qualifier Description
J	Indicates an Estimated Value for TICs
N	Presumptive evidence of material.
T	Result is a tentatively identified compound (TIC) and an estimated value.

GC VOA

Qualifier	Qualifier Description
F1	MS and/or MSD recovery exceeds control limits.
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.

GC Semi VOA

Qualifier	Qualifier Description
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
S1+	Surrogate recovery exceeds control limits, high biased.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)

Definitions/Glossary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Glossary (Continued)

Abbreviation	These commonly used abbreviations may or may not be present in this report.
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Case Narrative

Client: City & County of Honolulu
Project: RED-HILL

Job ID: 380-190084-1

Job ID: 380-190084-1

Eurofins Pomona

Job Narrative 380-190084-1

The analytical test results presented in this report meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page, unless otherwise noted. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable. Regulated compliance samples (e.g. SDWA, NPDES) must comply with associated agency requirements/permits.

- Matrix-specific batch QC (e.g., MS, MSD, SD) may not be reported when insufficient sample volume is available or when site-specific QC samples are not submitted. In such cases, a Laboratory Control Sample Duplicate (LCSD) may be analyzed to provide precision data for the batch.
- For samples analyzed using surrogate and/or isotope dilution analytes, any recoveries falling outside of established acceptance criteria are re-prepared and/or re-analyzed to confirm results, unless the deviation is due to sample dilution or otherwise explained in the case narrative.

Receipt

The samples were received on 12/31/2025 9:57 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 1.3°C.

GC/MS Semi VOA

Method 625.1: The surrogate recovery for the blank associated with preparation batch 570-678015 and analytical batch 570-681700 was outside the upper control limits.

Method 625.1_SIM: The spiking solution was inadvertently mis-spiked during the extraction process for the laboratory control sample duplicate (LCSD) associated with 570-678015; therefore, the relative percent difference between LCS/LCSD is unavailable. The batch matrix spike/matrix spike duplicate (MS/MSD) was within acceptance limits and may be used to evaluate precision.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Gasoline Range Organics

Method 8015B_GRO_LL: The matrix spike duplicate (MSD) recovery for analytical batch 570-678421 was outside control limits for Gasoline Range Organics (C4-C13). See QC Sample Results for detail. Sample matrix interference and/or non-homogeneity are suspected because the associated laboratory control sample (LCS) and laboratory control sample duplicate (LCSD) recoveries were within acceptance limits.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Diesel Range Organics

Method 8015B_DRO_LL_CS: An incorrect volume of surrogate spiking solution was inadvertently added the following samples: Halawa Shaft Viewing Pool (380-190084-1). Percent recoveries are based on the amount spiked.
Method: 8015B_DRO_LL_CS

Method 8015B_DRO_LL_CS: Surrogate recovery for the following sample was outside the upper control limit: Halawa Shaft Viewing Pool (380-190084-1). This sample did not contain any target analytes; therefore, re-extraction and/or re-analysis was not performed.
Method: 8015B_DRO_LL_CS

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Eurofins Pomona

Detection Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

No Detections.

Client Sample ID: TB: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-2

No Detections.

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

This Detection Summary does not include radiochemical test results.

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2,4'-DDD	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2,4'-DDE	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2,4'-DDT	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2,4-Dinitrotoluene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2,6-Dinitrotoluene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
2-Methylnaphthalene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
4,4'-DDD	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
4,4'-DDE	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
4,4'-DDT	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Acenaphthene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Acenaphthylene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Acetochlor	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Alachlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
alpha-BHC	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
alpha-Chlordane	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Anthracene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 16:32	1
Atrazine	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Benz(a)anthracene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Benzo[a]pyrene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 16:32	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 16:32	1
Benzo[g,h,i]perylene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 16:32	1
beta-BHC	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Bis(2-ethylhexyl) phthalate	<0.60		0.60	ug/L		01/02/26 14:48	01/04/26 16:32	1
Bromacil	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Butachlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Butylbenzylphthalate	<0.50		0.50	ug/L		01/02/26 14:48	01/04/26 16:32	1
Chlorobenzilate	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Chloroneb	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Chlorothalonil (Draconil, Bravo)	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Chlorpyrifos	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Chrysene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 16:32	1
delta-BHC	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Di(2-ethylhexyl)adipate	<0.60		0.60	ug/L		01/02/26 14:48	01/04/26 16:32	1
Dibenz(a,h)anthracene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Diclorvos (DDVP)	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Dieldrin	<0.0099		0.0099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Diethylphthalate	<0.50		0.50	ug/L		01/02/26 14:48	01/04/26 16:32	1
Dimethylphthalate	<0.50		0.50	ug/L		01/02/26 14:48	01/04/26 16:32	1
Di-n-butyl phthalate	<0.99		0.99	ug/L		01/02/26 14:48	01/04/26 16:32	1
Di-n-octyl phthalate	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Endosulfan I (Alpha)	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Endosulfan II (Beta)	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Endosulfan sulfate	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Endrin	<0.0099		0.0099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Endrin aldehyde	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
EPTC	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Fluoranthene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
gamma-Chlordane	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Heptachlor	<0.0099		0.0099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Heptachlor epoxide (isomer B)	<0.0099	^3+	0.0099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Hexachlorobenzene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Hexachlorocyclopentadiene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Indeno[1,2,3-cd]pyrene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Isophorone	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Lindane	<0.0099		0.0099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Malathion	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Methoxychlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Metolachlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Molinate	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Naphthalene	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Parathion	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Pendimethalin (Penoxaline)	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Phenanthrene	<0.040		0.040	ug/L		01/02/26 14:48	01/04/26 16:32	1
Propachlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Pyrene	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Simazine	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Terbacil	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Terbutylazine	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Thiobencarb	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		01/02/26 14:48	01/04/26 16:32	1
trans-Nonachlor	<0.050		0.050	ug/L		01/02/26 14:48	01/04/26 16:32	1
Trifluralin	<0.099		0.099	ug/L		01/02/26 14:48	01/04/26 16:32	1

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
9-Octadecenamamide, (Z)-	5.5	T J N	ug/L		7.96	301-02-0	01/02/26 14:48	01/04/26 16:32	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2-Nitro-m-xylene	98		70 - 130	01/02/26 14:48	01/04/26 16:32	1
Perylene-d12	91		70 - 130	01/02/26 14:48	01/04/26 16:32	1
Triphenylphosphate	105		70 - 130	01/02/26 14:48	01/04/26 16:32	1

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
1-Methylnaphthalene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
2-Methylnaphthalene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Acenaphthene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Acenaphthylene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Anthracene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Benzo[a]anthracene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Benzo[a]pyrene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Benzo[b]fluoranthene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Benzo[g,h,i]perylene	<0.19	*1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Benzo[k]fluoranthene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Chrysene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Dibenz(a,h)anthracene	<0.19	*1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Fluoranthene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1

Eurofins Pomona

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Method: EPA 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Fluorene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Indeno[1,2,3-cd]pyrene	<0.19	*1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Naphthalene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Phenanthrene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1
Pyrene	<0.19	*- *1	0.19	ug/L		01/02/26 12:55	01/13/26 15:39	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	95		28 - 127	01/02/26 12:55	01/13/26 15:39	1
2-Fluorobiphenyl (Surr)	96		31 - 120	01/02/26 12:55	01/13/26 15:39	1
2-Fluorophenol (Surr)	62		17 - 120	01/02/26 12:55	01/13/26 15:39	1
Nitrobenzene-d5 (Surr)	101		27 - 120	01/02/26 12:55	01/13/26 15:39	1
Phenol-d6 (Surr)	37		10 - 120	01/02/26 12:55	01/13/26 15:39	1
p-Terphenyl-d14 (Surr)	97		45 - 120	01/02/26 12:55	01/13/26 15:39	1

Method: EPA 625.1 - Semivolatile Organic Compounds (GC/MS)

Tentatively Identified Compound	Est. Result	Qualifier	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
Tentatively Identified Compound	None		ug/L			N/A	01/02/26 12:55	01/13/26 19:29	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
2,4,6-Tribromophenol (Surr)	69		33 - 139	01/02/26 12:55	01/13/26 19:29	1
2-Fluorobiphenyl (Surr)	93		33 - 126	01/02/26 12:55	01/13/26 19:29	1
2-Fluorophenol (Surr)	61		12 - 120	01/02/26 12:55	01/13/26 19:29	1
Nitrobenzene-d5 (Surr)	108		36 - 120	01/02/26 12:55	01/13/26 19:29	1
Phenol-d6 (Surr)	37		10 - 120	01/02/26 12:55	01/13/26 19:29	1
p-Terphenyl-d14 (Surr)	98		47 - 131	01/02/26 12:55	01/13/26 19:29	1

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			01/04/26 21:56	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
4-Bromofluorobenzene (Surr)	105		38 - 134		01/04/26 21:56	1

Method: SW846 8015B - Diesel Range Organics (DRO) (GC) Low Level

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Diesel Range Organics (C10-C24)	<26		26	ug/L		01/04/26 09:33	01/07/26 23:58	1
Motor Oil Range Organics [C24-C36]	<26		26	ug/L		01/04/26 09:33	01/07/26 23:58	1
C8-C18	<26		26	ug/L		01/04/26 09:33	01/07/26 23:58	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
n-Octacosane (Surr)	139	S1+	60 - 130	01/04/26 09:33	01/07/26 23:58	1

Client Sample ID: TB: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-2

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Method: SW846 8015B GRO LL - Gasoline Range Organics - (GC)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
GRO (C6-C10)	<10		10	ug/L			01/05/26 01:37	1

Client Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: TB: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-2

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

<u>Surrogate</u>	<u>%Recovery</u>	<u>Qualifier</u>	<u>Limits</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Dil Fac</u>
4-Bromofluorobenzene (Surr)	108		38 - 134		01/05/26 01:37	1

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Action Limit Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

Compliance Check

The results obtained from the analytical testing of this data set were checked against compliance limits received from the client. Any results at or above the compliance limits have been highlighted for your convenience.

Analyte	Result	Qualifier	Unit	EPAMCL	RL	Method	Prep Type
				Limit			
Alachlor	<0.050		ug/L	2	0.050	525.2	Total/NA
Atrazine	<0.050		ug/L	3	0.050	525.2	Total/NA
Benzo[a]pyrene	<0.020		ug/L	0.2	0.020	525.2	Total/NA
Bis(2-ethylhexyl) phthalate	<0.60		ug/L	6	0.60	525.2	Total/NA
Di(2-ethylhexyl)adipate	<0.60		ug/L	400	0.60	525.2	Total/NA
Endrin	<0.0099		ug/L	2	0.0099	525.2	Total/NA
Heptachlor	<0.0099		ug/L	0.4	0.0099	525.2	Total/NA
Heptachlor epoxide (isomer B)	<0.0099	^3+	ug/L	0.2	0.0099	525.2	Total/NA
Hexachlorobenzene	<0.050		ug/L	1	0.050	525.2	Total/NA
Hexachlorocyclopentadiene	<0.050		ug/L	50	0.050	525.2	Total/NA
Lindane	<0.0099		ug/L	0.2	0.0099	525.2	Total/NA
Methoxychlor	<0.050		ug/L	40	0.050	525.2	Total/NA
Simazine	<0.050		ug/L	4	0.050	525.2	Total/NA
Benzo[a]pyrene	<0.19	*- *1	ug/L	0.2	0.19	625.1 SIM	Total/NA

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		2NMX (70-130)	PRY (70-130)	TPP (70-130)
380-190084-1	Halawa Shaft Viewing Pool	98	91	105
380-190097-I-1-A MS	Matrix Spike	96	97	107
380-190160-B-1-A DU	Duplicate	96	91	107
LCS 380-195565/23-A	Lab Control Sample	95	88	107
MB 380-195565/21-A	Method Blank	97	92	106
MRL 380-195565/22-A	Lab Control Sample	96	95	105

Surrogate Legend

2NMX = 2-Nitro-m-xylene
PRY = Perylene-d12
TPP = Triphenylphosphate

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (33-139)	FBP (33-126)	2FP (12-120)	NBZ (36-120)	PHL6 (10-120)	TPHd14 (47-131)
380-190084-1	Halawa Shaft Viewing Pool	69	93	61	108	37	98
MB 570-678015/1-A	Method Blank	90	102	71	125 S1+	46	112

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)					
		TBP (28-127)	FBP (31-120)	2FP (17-120)	NBZ (27-120)	PHL6 (10-120)	TPHd14 (45-120)
380-190084-1	Halawa Shaft Viewing Pool	95	96	62	101	37	97
380-190087-A-1-A MS	Matrix Spike	92	92	70	80	44	101
380-190087-A-1-B MSD	Matrix Spike Duplicate	89	85	66	78	42	96
LCS 570-678015/2-A	Lab Control Sample	99	94	75	85	49	107
LCSD 570-678015/3-A	Lab Control Sample Dup	16 S1-	16 S1-	11 S1-	16 S1-	7 S1-	19 S1-
MB 570-678015/1-A	Method Blank	102	95	72	112	45	113

Surrogate Legend

TBP = 2,4,6-Tribromophenol (Surr)
FBP = 2-Fluorobiphenyl (Surr)
2FP = 2-Fluorophenol (Surr)
NBZ = Nitrobenzene-d5 (Surr)
PHL6 = Phenol-d6 (Surr)
TPHd14 = p-Terphenyl-d14 (Surr)

Surrogate Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	BFB1 (38-134)
380-190084-1	Halawa Shaft Viewing Pool	105
380-190084-2	TB: Halawa Shaft Viewing Pool	108
380-190087-C-1 MS	Matrix Spike	102
380-190087-C-1 MSD	Matrix Spike Duplicate	107
LCS 570-678421/1011	Lab Control Sample	91
LCSD 570-678421/12	Lab Control Sample Dup	94
MB 570-678421/13	Method Blank	98
MRL 570-678421/14	Lab Control Sample	94

Surrogate Legend

BFB = 4-Bromofluorobenzene (Surr)

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Matrix: Water

Prep Type: Total/NA

Percent Surrogate Recovery (Acceptance Limits)

Lab Sample ID	Client Sample ID	OTCSN1 (60-130)
380-190084-1	Halawa Shaft Viewing Pool	139 S1+
380-190087-B-1-B MS	Matrix Spike	107
380-190087-B-1-C MSD	Matrix Spike Duplicate	114
LCS 570-678355/2-A	Lab Control Sample	97
LCSD 570-678355/3-A	Lab Control Sample Dup	97
MB 570-678355/1-A	Method Blank	104
MRL 570-678355/4-A	Lab Control Sample	107

Surrogate Legend

OTCSN = n-Octacosane (Surr)

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 380-195565/21-A
Matrix: Water
Analysis Batch: 195644

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195565

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
1-Methylnaphthalene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2,4'-DDD	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2,4'-DDE	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2,4'-DDT	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2,4-Dinitrotoluene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2,6-Dinitrotoluene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
2-Methylnaphthalene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
4,4'-DDD	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
4,4'-DDE	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
4,4'-DDT	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Acenaphthene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Acenaphthylene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Acetochlor	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Alachlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
alpha-BHC	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
alpha-Chlordane	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Anthracene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 13:11	1
Atrazine	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Benz(a)anthracene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Benzo[a]pyrene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 13:11	1
Benzo[b]fluoranthene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 13:11	1
Benzo[g,h,i]perylene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Benzo[k]fluoranthene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 13:11	1
beta-BHC	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Bis(2-ethylhexyl) phthalate	<0.59		0.59	ug/L		01/02/26 14:48	01/04/26 13:11	1
Bromacil	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Butachlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Butylbenzylphthalate	<0.49		0.49	ug/L		01/02/26 14:48	01/04/26 13:11	1
Chlorobenzilate	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Chloroneb	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Chlorothalonil (Draconil, Bravo)	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Chlorpyrifos	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Chrysene	<0.020		0.020	ug/L		01/02/26 14:48	01/04/26 13:11	1
delta-BHC	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Di(2-ethylhexyl)adipate	<0.59		0.59	ug/L		01/02/26 14:48	01/04/26 13:11	1
Dibenz(a,h)anthracene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Diclorvos (DDVP)	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Dieldrin	<0.0098		0.0098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Diethylphthalate	<0.49		0.49	ug/L		01/02/26 14:48	01/04/26 13:11	1
Dimethylphthalate	<0.49		0.49	ug/L		01/02/26 14:48	01/04/26 13:11	1
Di-n-butyl phthalate	<0.98		0.98	ug/L		01/02/26 14:48	01/04/26 13:11	1
Di-n-octyl phthalate	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Endosulfan I (Alpha)	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Endosulfan II (Beta)	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Endosulfan sulfate	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Endrin	<0.0098		0.0098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Endrin aldehyde	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
EPTC	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MB 380-195565/21-A
Matrix: Water
Analysis Batch: 195644

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 195565

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Fluoranthene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Fluorene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
gamma-Chlordane	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Heptachlor	<0.0098		0.0098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Heptachlor epoxide (isomer B)	<0.0098	^3+	0.0098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Hexachlorobenzene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Hexachlorocyclopentadiene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Indeno[1,2,3-cd]pyrene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Isophorone	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Lindane	<0.0098		0.0098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Malathion	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Methoxychlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Metolachlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Molinate	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Naphthalene	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Parathion	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Pendimethalin (Penoxaline)	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Phenanthrene	<0.039		0.039	ug/L		01/02/26 14:48	01/04/26 13:11	1
Propachlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Pyrene	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Simazine	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Terbacil	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Terbutylazine	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Thiobencarb	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1
Total Permethrin (mixed isomers)	<0.20		0.20	ug/L		01/02/26 14:48	01/04/26 13:11	1
trans-Nonachlor	<0.049		0.049	ug/L		01/02/26 14:48	01/04/26 13:11	1
Trifluralin	<0.098		0.098	ug/L		01/02/26 14:48	01/04/26 13:11	1

Tentatively Identified Compound	MB	MB	Unit	D	RT	CAS No.	Prepared	Analyzed	Dil Fac
	Est. Result	Qualifier							
Cyclopentene, 1,2,3,4,5-pentamethyl-	1.12	T J N	ug/L		2.53	1000154-28-6	01/02/26 14:48	01/04/26 13:11	1
Decane, 5-methyl-	0.507	T J N	ug/L		2.99	13151-35-4	01/02/26 14:48	01/04/26 13:11	1
Undecane	2.77	T J N	ug/L		3.14	1120-21-4	01/02/26 14:48	01/04/26 13:11	1
Plumbane, diethyldimethyl-	0.628	T J N	ug/L		3.27	1762-27-2	01/02/26 14:48	01/04/26 13:11	1
Hexadecanamide	0.575	T J N	ug/L		7.09	629-54-9	01/02/26 14:48	01/04/26 13:11	1
9-Octadecenamide, (Z)-	6.47	T J N	ug/L		7.97	301-02-0	01/02/26 14:48	01/04/26 13:11	1
Unknown	0.497	T J	ug/L		15.14	N/A	01/02/26 14:48	01/04/26 13:11	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
2-Nitro-m-xylene	97		70 - 130	01/02/26 14:48	01/04/26 13:11	1
Perylene-d12	92		70 - 130	01/02/26 14:48	01/04/26 13:11	1
Triphenylphosphate	106		70 - 130	01/02/26 14:48	01/04/26 13:11	1

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-195565/23-A
Matrix: Water
Analysis Batch: 195644

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 195565

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	1.97	2.04		ug/L		104	70 - 130
2,4'-DDD	1.97	1.92		ug/L		98	70 - 130
2,4'-DDE	1.97	2.10		ug/L		107	70 - 130
2,4'-DDT	1.97	1.86		ug/L		95	70 - 130
2,4-Dinitrotoluene	1.97	1.82		ug/L		93	70 - 130
2,6-Dinitrotoluene	1.97	1.78		ug/L		91	70 - 130
2-Methylnaphthalene	1.97	2.02		ug/L		103	70 - 130
4,4'-DDD	1.97	1.99		ug/L		101	70 - 130
4,4'-DDE	1.97	2.10		ug/L		107	70 - 130
4,4'-DDT	1.97	1.88		ug/L		96	70 - 130
Acenaphthene	1.97	2.05		ug/L		104	70 - 130
Acenaphthylene	1.97	2.04		ug/L		104	70 - 130
Acetochlor	1.97	2.07		ug/L		105	70 - 130
Alachlor	1.97	2.12		ug/L		108	70 - 130
alpha-BHC	1.97	2.22		ug/L		113	70 - 130
alpha-Chlordane	1.97	1.83		ug/L		93	70 - 130
Anthracene	1.97	2.12		ug/L		108	70 - 130
Atrazine	1.97	2.04		ug/L		104	70 - 130
Benz(a)anthracene	1.97	1.95		ug/L		99	70 - 130
Benzo[a]pyrene	1.97	1.88		ug/L		95	70 - 130
Benzo[b]fluoranthene	1.97	1.99		ug/L		101	70 - 130
Benzo[g,h,i]perylene	1.97	2.11		ug/L		107	70 - 130
Benzo[k]fluoranthene	1.97	1.78		ug/L		91	70 - 130
beta-BHC	1.97	2.33		ug/L		118	70 - 130
Bis(2-ethylhexyl) phthalate	1.97	2.17		ug/L		111	70 - 130
Bromacil	1.97	1.74		ug/L		89	70 - 130
Butachlor	1.97	2.32		ug/L		118	70 - 130
Butylbenzylphthalate	1.97	2.21		ug/L		112	70 - 130
Chlorobenzilate	1.97	2.31		ug/L		117	70 - 130
Chloroneb	1.97	2.16		ug/L		110	70 - 130
Chlorothalonil (Draconil, Bravo)	1.97	2.05		ug/L		104	70 - 130
Chlorpyrifos	1.97	1.95		ug/L		99	70 - 130
Chrysene	1.97	1.71		ug/L		87	70 - 130
delta-BHC	1.97	2.26		ug/L		115	70 - 130
Di(2-ethylhexyl)adipate	1.97	2.39		ug/L		122	70 - 130
Dibenz(a,h)anthracene	1.97	2.17		ug/L		110	70 - 130
Diclorvos (DDVP)	1.97	2.00		ug/L		102	70 - 130
Dieldrin	1.97	2.02		ug/L		103	70 - 130
Diethylphthalate	1.97	2.24		ug/L		114	70 - 130
Dimethylphthalate	1.97	2.24		ug/L		114	70 - 130
Di-n-butyl phthalate	3.93	4.14		ug/L		105	70 - 130
Di-n-octyl phthalate	1.97	2.04		ug/L		104	70 - 130
Endosulfan I (Alpha)	1.97	1.95		ug/L		99	70 - 130
Endosulfan II (Beta)	1.97	2.19		ug/L		112	70 - 130
Endosulfan sulfate	1.97	2.03		ug/L		103	70 - 130
Endrin	1.97	1.93		ug/L		98	70 - 130
Endrin aldehyde	1.97	2.06		ug/L		105	60 - 130
EPTC	1.97	2.35		ug/L		120	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: LCS 380-195565/23-A

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Fluoranthene	1.97	2.21		ug/L		112	70 - 130
Fluorene	1.97	2.22		ug/L		113	70 - 130
gamma-Chlordane	1.97	1.84		ug/L		93	70 - 130
Heptachlor	1.97	2.13		ug/L		108	70 - 130
Heptachlor epoxide (isomer B)	1.97	2.04		ug/L		104	70 - 130
Hexachlorobenzene	1.97	2.27		ug/L		116	70 - 130
Hexachlorocyclopentadiene	1.97	1.83		ug/L		93	70 - 130
Indeno[1,2,3-cd]pyrene	1.97	2.19		ug/L		112	70 - 130
Isophorone	1.97	2.00		ug/L		102	70 - 130
Lindane	1.97	2.15		ug/L		109	70 - 130
Malathion	1.97	2.14		ug/L		109	70 - 130
Methoxychlor	1.97	1.93		ug/L		98	70 - 130
Metolachlor	1.97	2.20		ug/L		112	70 - 130
Molinate	1.97	2.23		ug/L		113	70 - 130
Naphthalene	1.97	2.09		ug/L		106	70 - 130
Parathion	1.97	2.05		ug/L		104	70 - 130
Pendimethalin (Penoxaline)	1.97	1.79		ug/L		91	70 - 130
Phenanthrene	1.97	2.09		ug/L		106	70 - 130
Propachlor	1.97	2.19		ug/L		111	70 - 130
Pyrene	1.97	2.00		ug/L		102	70 - 130
Simazine	1.97	1.98		ug/L		101	70 - 130
Terbacil	1.97	1.90		ug/L		97	70 - 130
Terbutylazine	1.97	2.25		ug/L		115	70 - 130
Thiobencarb	1.97	2.04		ug/L		104	70 - 130
trans-Nonachlor	1.97	1.80		ug/L		92	70 - 130
Trifluralin	1.97	1.84		ug/L		94	70 - 130

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	95		70 - 130
Perylene-d12	88		70 - 130
Triphenylphosphate	107		70 - 130

Lab Sample ID: MRL 380-195565/22-A

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	0.0984	0.114		ug/L		116	50 - 150
2,4'-DDD	0.0984	0.0829	J	ug/L		84	50 - 150
2,4'-DDE	0.0984	0.0883	J	ug/L		90	50 - 150
2,4'-DDT	0.0984	0.116		ug/L		118	50 - 150
2,4-Dinitrotoluene	0.0984	0.130		ug/L		132	50 - 150
2,6-Dinitrotoluene	0.0984	0.129		ug/L		131	50 - 150
2-Methylnaphthalene	0.0984	0.109		ug/L		111	50 - 150
4,4'-DDD	0.0984	0.113		ug/L		115	50 - 150
4,4'-DDE	0.0984	0.106		ug/L		107	50 - 150
4,4'-DDT	0.0984	0.120		ug/L		122	50 - 150
Acenaphthene	0.0984	0.0911	J	ug/L		93	50 - 150

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-195565/22-A

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Spike	MRL	MRL	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
Acenaphthylene	0.0984	0.0927	J	ug/L		94	50 - 150
Acetochlor	0.0984	0.115		ug/L		117	50 - 150
Alachlor	0.0492	0.0530		ug/L		108	50 - 150
alpha-BHC	0.0984	0.0950	J	ug/L		97	50 - 150
alpha-Chlordane	0.0246	<0.029		ug/L		111	50 - 150
Anthracene	0.0197	0.0226		ug/L		115	50 - 150
Atrazine	0.0492	0.0619		ug/L		126	50 - 150
Benz(a)anthracene	0.0492	0.0594		ug/L		121	50 - 150
Benzo[a]pyrene	0.0197	0.0268		ug/L		136	50 - 150
Benzo[b]fluoranthene	0.0197	0.0260		ug/L		132	50 - 150
Benzo[g,h,i]perylene	0.0492	0.0594		ug/L		121	50 - 150
Benzo[k]fluoranthene	0.0197	0.0237		ug/L		120	50 - 150
beta-BHC	0.0984	0.107		ug/L		109	50 - 150
Bis(2-ethylhexyl) phthalate	0.590	0.665		ug/L		113	50 - 150
Bromacil	0.0984	0.123		ug/L		125	50 - 150
Butachlor	0.0492	0.0631		ug/L		128	50 - 150
Butylbenzylphthalate	0.492	0.540		ug/L		110	50 - 150
Chlorobenzilate	0.0984	0.125		ug/L		127	50 - 150
Chloroneb	0.0984	0.102		ug/L		104	50 - 150
Chlorothalonil (Draconil, Bravo)	0.0984	0.141		ug/L		144	50 - 150
Chlorpyrifos	0.0492	0.0574		ug/L		117	50 - 150
Chrysene	0.0197	0.0217		ug/L		110	50 - 150
delta-BHC	0.0984	0.101		ug/L		102	50 - 150
Di(2-ethylhexyl)adipate	0.590	0.757		ug/L		128	50 - 150
Dibenz(a,h)anthracene	0.0492	0.0629		ug/L		128	50 - 150
Diclorvos (DDVP)	0.0492	0.0524		ug/L		106	50 - 150
Dieldrin	0.00984	0.00972	J	ug/L		99	50 - 150
Diethylphthalate	0.492	0.510		ug/L		104	50 - 150
Dimethylphthalate	0.492	0.493		ug/L		100	50 - 150
Di-n-butyl phthalate	0.492	0.472	J	ug/L		96	49 - 243
Di-n-octyl phthalate	0.0984	0.131		ug/L		133	50 - 150
Endosulfan I (Alpha)	0.0984	0.0926	J	ug/L		94	50 - 150
Endosulfan II (Beta)	0.0984	0.110		ug/L		112	50 - 150
Endosulfan sulfate	0.0984	0.111		ug/L		113	50 - 150
Endrin	0.00984	0.00797	J	ug/L		81	50 - 150
Endrin aldehyde	0.0984	0.130		ug/L		132	50 - 150
EPTC	0.0984	0.105		ug/L		107	50 - 150
Fluoranthene	0.0984	0.102		ug/L		104	50 - 150
Fluorene	0.0492	0.0499		ug/L		101	50 - 150
gamma-Chlordane	0.0246	0.0307	J	ug/L		125	50 - 150
Heptachlor	0.00984	0.0140		ug/L		142	50 - 150
Heptachlor epoxide (isomer B)	0.00984	0.0153	^3+	ug/L		156	50 - 150
Hexachlorobenzene	0.0492	0.0438	J	ug/L		89	50 - 150
Hexachlorocyclopentadiene	0.0492	0.0503		ug/L		102	50 - 150
Indeno[1,2,3-cd]pyrene	0.0492	0.0645		ug/L		131	50 - 150
Isophorone	0.0984	0.115		ug/L		117	50 - 150
Lindane	0.00984	0.0106		ug/L		107	50 - 150
Malathion	0.0984	0.117		ug/L		119	50 - 150
Methoxychlor	0.0492	0.0648		ug/L		132	50 - 150

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: MRL 380-195565/22-A

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Spike	MRL	MRL	Unit	D	%Rec	%Rec Limits
	Added	Result	Qualifier				
Metolachlor	0.0492	0.0541		ug/L		110	50 - 150
Molinate	0.0984	0.127		ug/L		129	50 - 150
Naphthalene	0.0984	0.0884	J	ug/L		90	50 - 150
Parathion	0.0984	0.0922	J	ug/L		94	50 - 150
Pendimethalin (Penoxaline)	0.0984	0.117		ug/L		119	50 - 150
Phenanthrene	0.0394	0.0431		ug/L		110	50 - 150
Propachlor	0.0492	0.0553		ug/L		113	50 - 150
Pyrene	0.0492	0.0541		ug/L		110	50 - 150
Simazine	0.0492	0.0523		ug/L		106	50 - 150
Terbacil	0.0984	0.107		ug/L		109	50 - 150
Terbuthylazine	0.0984	0.102		ug/L		104	50 - 150
Thiobencarb	0.0984	0.111		ug/L		113	50 - 150
trans-Nonachlor	0.0246	0.0273	J	ug/L		111	50 - 150
Trifluralin	0.0984	0.108		ug/L		110	50 - 150

Surrogate	MRL	MRL	Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	96		70 - 130
Perylene-d 12	95		70 - 130
Triphenylphosphate	105		70 - 130

Lab Sample ID: 380-190097-I-1-A MS

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Matrix Spike

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
1-Methylnaphthalene	<0.099		1.97	2.03		ug/L		102	70 - 130
2,4'-DDD	<0.099		1.97	1.99		ug/L		101	70 - 130
2,4'-DDE	<0.099		1.97	2.18		ug/L		111	70 - 130
2,4'-DDT	<0.099		1.97	2.00		ug/L		101	70 - 130
2,4-Dinitrotoluene	<0.099		1.97	1.98		ug/L		101	70 - 130
2,6-Dinitrotoluene	<0.099		1.97	1.81		ug/L		92	70 - 130
2-Methylnaphthalene	<0.099		1.97	2.01		ug/L		102	70 - 130
4,4'-DDD	<0.099		1.97	2.05		ug/L		104	70 - 130
4,4'-DDE	<0.099		1.97	2.15		ug/L		109	70 - 130
4,4'-DDT	<0.099		1.97	1.99		ug/L		101	70 - 130
Acenaphthene	<0.099		1.97	2.04		ug/L		103	70 - 130
Acenaphthylene	<0.099		1.97	2.08		ug/L		106	70 - 130
Acetochlor	<0.099		1.97	2.09		ug/L		106	70 - 130
Alachlor	<0.050		1.97	2.19		ug/L		111	70 - 130
alpha-BHC	<0.099		1.97	2.18		ug/L		111	70 - 130
alpha-Chlordane	<0.050		1.97	1.88		ug/L		95	70 - 130
Anthracene	<0.020		1.97	1.47		ug/L		75	70 - 130
Atrazine	<0.050		1.97	2.00		ug/L		102	70 - 130
Benz(a)anthracene	<0.050		1.97	1.92		ug/L		97	70 - 130
Benzo[a]pyrene	<0.020		1.97	1.73		ug/L		88	70 - 130
Benzo[b]fluoranthene	<0.020		1.97	2.00		ug/L		102	70 - 130
Benzo[g,h,i]perylene	<0.050		1.97	2.13		ug/L		108	70 - 130
Benzo[k]fluoranthene	<0.020		1.97	1.84		ug/L		93	70 - 130

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-190097-I-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 195644

Prep Batch: 195565

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
beta-BHC	<0.099		1.97	2.25		ug/L		114	70 - 130
Bis(2-ethylhexyl) phthalate	<0.59		1.97	2.06		ug/L		105	70 - 130
Bromacil	<0.099		1.97	1.76		ug/L		89	70 - 130
Butachlor	<0.050		1.97	2.32		ug/L		118	70 - 130
Butylbenzylphthalate	<0.50		1.97	2.30		ug/L		117	70 - 130
Chlorobenzilate	<0.099		1.97	2.32		ug/L		118	70 - 130
Chloroneb	<0.099		1.97	2.14		ug/L		109	70 - 130
Chlorothalonil (Draconil, Bravo)	<0.099		1.97	2.13		ug/L		108	70 - 130
Chlorpyrifos	<0.050		1.97	1.98		ug/L		101	70 - 130
Chrysene	<0.020		1.97	1.77		ug/L		90	70 - 130
delta-BHC	<0.099		1.97	2.28		ug/L		116	70 - 130
Di(2-ethylhexyl)adipate	<0.59		1.97	2.35		ug/L		119	70 - 130
Dibenz(a,h)anthracene	<0.050		1.97	2.18		ug/L		111	70 - 130
Diclorvos (DDVP)	<0.050		1.97	1.97		ug/L		100	70 - 130
Dieldrin	<0.0099		1.97	2.11		ug/L		107	70 - 130
Diethylphthalate	<0.50		1.97	2.15		ug/L		109	70 - 130
Dimethylphthalate	<0.50		1.97	2.18		ug/L		111	70 - 130
Di-n-butyl phthalate	<0.99		3.94	4.19		ug/L		106	70 - 130
Di-n-octyl phthalate	<0.099		1.97	1.95		ug/L		99	70 - 130
Endosulfan I (Alpha)	<0.099		1.97	2.00		ug/L		101	70 - 130
Endosulfan II (Beta)	<0.099		1.97	2.23		ug/L		113	70 - 130
Endosulfan sulfate	<0.099		1.97	2.11		ug/L		107	70 - 130
Endrin	<0.0099		1.97	2.12		ug/L		108	70 - 130
Endrin aldehyde	<0.099		1.97	1.86		ug/L		95	60 - 130
EPTC	<0.099		1.97	2.32		ug/L		118	70 - 130
Fluoranthene	<0.099		1.97	2.24		ug/L		114	70 - 130
Fluorene	<0.050		1.97	2.21		ug/L		112	70 - 130
gamma-Chlordane	<0.050		1.97	1.92		ug/L		97	70 - 130
Heptachlor	<0.0099		1.97	2.25		ug/L		114	70 - 130
Heptachlor epoxide (isomer B)	<0.0099	^3+	1.97	2.09		ug/L		106	70 - 130
Hexachlorobenzene	<0.050		1.97	2.26		ug/L		115	70 - 130
Hexachlorocyclopentadiene	<0.050		1.97	1.86		ug/L		94	70 - 130
Indeno[1,2,3-cd]pyrene	<0.050		1.97	2.32		ug/L		118	70 - 130
Isophorone	<0.099		1.97	1.92		ug/L		98	70 - 130
Lindane	<0.0099		1.97	2.12		ug/L		108	70 - 130
Malathion	<0.099		1.97	2.18		ug/L		110	70 - 130
Methoxychlor	<0.050		1.97	2.04		ug/L		104	70 - 130
Metolachlor	<0.050		1.97	2.22		ug/L		113	70 - 130
Molinate	<0.099		1.97	2.17		ug/L		110	70 - 130
Naphthalene	<0.099		1.97	2.11		ug/L		107	70 - 130
Parathion	<0.099		1.97	2.23		ug/L		113	70 - 130
Pendimethalin (Penoxaline)	<0.099		1.97	2.05		ug/L		104	70 - 130
Phenanthrene	<0.040		1.97	2.14		ug/L		108	70 - 130
Propachlor	<0.050		1.97	2.10		ug/L		107	70 - 130
Pyrene	<0.050		1.97	2.04		ug/L		104	70 - 130
Simazine	<0.050		1.97	1.91		ug/L		97	70 - 130
Terbacil	<0.099		1.97	1.85		ug/L		94	70 - 130
Terbuthylazine	<0.099		1.97	2.18		ug/L		111	70 - 130
Thiobencarb	<0.099		1.97	2.06		ug/L		105	70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-190097-I-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 195644

Prep Batch: 195565

Analyte	Sample	Sample	Spike	MS	MS	Unit	D	%Rec	%Rec Limits
	Result	Qualifier	Added	Result	Qualifier				
trans-Nonachlor	<0.050		1.97	1.91		ug/L		97	70 - 130
Trifluralin	<0.099		1.97	1.98		ug/L		101	70 - 130

Surrogate	MS	MS	Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	96		70 - 130
Perylene-d12	97		70 - 130
Triphenylphosphate	107		70 - 130

Lab Sample ID: 380-190160-B-1-A DU

Client Sample ID: Duplicate

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 195644

Prep Batch: 195565

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	Limit
	Result	Qualifier	Result	Qualifier				
1-Methylnaphthalene	<0.098		<0.098		ug/L		NC	20
2,4'-DDD	<0.098		<0.098		ug/L		NC	20
2,4'-DDE	<0.098		<0.098		ug/L		NC	20
2,4'-DDT	<0.098		<0.098		ug/L		NC	20
2,4-Dinitrotoluene	<0.098		<0.098		ug/L		NC	20
2,6-Dinitrotoluene	<0.098		<0.098		ug/L		NC	20
2-Methylnaphthalene	<0.098		<0.098		ug/L		NC	20
4,4'-DDD	<0.098		<0.098		ug/L		NC	20
4,4'-DDE	<0.098		<0.098		ug/L		NC	20
4,4'-DDT	<0.098		<0.098		ug/L		NC	20
Acenaphthene	<0.098		<0.098		ug/L		NC	20
Acenaphthylene	<0.098		<0.098		ug/L		NC	20
Acetochlor	<0.098		<0.098		ug/L		NC	20
Alachlor	<0.049		<0.049		ug/L		NC	20
alpha-BHC	<0.098		<0.098		ug/L		NC	20
alpha-Chlordane	<0.049		<0.049		ug/L		NC	20
Anthracene	<0.020		<0.020		ug/L		NC	20
Atrazine	<0.049		<0.049		ug/L		NC	20
Benz(a)anthracene	<0.049		<0.049		ug/L		NC	20
Benzo[a]pyrene	<0.020		<0.020		ug/L		NC	20
Benzo[b]fluoranthene	<0.020		<0.020		ug/L		NC	20
Benzo[g,h,i]perylene	<0.049		<0.049		ug/L		NC	20
Benzo[k]fluoranthene	<0.020		<0.020		ug/L		NC	20
beta-BHC	<0.098		<0.098		ug/L		NC	20
Bis(2-ethylhexyl) phthalate	<0.59		<0.59		ug/L		NC	20
Bromacil	<0.098		<0.098		ug/L		NC	20
Butachlor	<0.049		<0.049		ug/L		NC	20
Butylbenzylphthalate	<0.49		<0.49		ug/L		NC	20
Chlorobenzilate	<0.098		<0.098		ug/L		NC	20
Chloroneb	<0.098		<0.098		ug/L		NC	20
Chlorothalonil (Draconil, Bravo)	<0.098		<0.098		ug/L		NC	20
Chlorpyrifos	<0.049		<0.049		ug/L		NC	20
Chrysene	<0.020		<0.020		ug/L		NC	20
delta-BHC	<0.098		<0.098		ug/L		NC	20
Di(2-ethylhexyl)adipate	<0.59		<0.59		ug/L		NC	20

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 525.2 - Semivolatile Organic Compounds (GC/MS) (Continued)

Lab Sample ID: 380-190160-B-1-A DU

Matrix: Water

Analysis Batch: 195644

Client Sample ID: Duplicate

Prep Type: Total/NA

Prep Batch: 195565

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	Limit
	Result	Qualifier	Result	Qualifier				
Dibenz(a,h)anthracene	<0.049		<0.049		ug/L		NC	20
Diclorvos (DDVP)	<0.049		<0.049		ug/L		NC	20
Dieldrin	<0.0098		<0.0098		ug/L		NC	20
Diethylphthalate	<0.49		<0.49		ug/L		NC	20
Dimethylphthalate	<0.49		<0.49		ug/L		NC	20
Di-n-butyl phthalate	<0.98		<0.98		ug/L		NC	20
Di-n-octyl phthalate	<0.098		<0.098		ug/L		NC	20
Endosulfan I (Alpha)	<0.098		<0.098		ug/L		NC	20
Endosulfan II (Beta)	<0.098		<0.098		ug/L		NC	20
Endosulfan sulfate	<0.098		<0.098		ug/L		NC	20
Endrin	<0.0098		<0.0098		ug/L		NC	20
Endrin aldehyde	<0.098		<0.098		ug/L		NC	20
EPTC	<0.098		<0.098		ug/L		NC	20
Fluoranthene	<0.098		<0.098		ug/L		NC	20
Fluorene	<0.049		<0.049		ug/L		NC	20
gamma-Chlordane	<0.049		<0.049		ug/L		NC	20
Heptachlor	<0.0098		<0.0098		ug/L		NC	20
Heptachlor epoxide (isomer B)	<0.0098	^3+	<0.0098		ug/L		NC	20
Hexachlorobenzene	<0.049		<0.049		ug/L		NC	20
Hexachlorocyclopentadiene	<0.049		<0.049		ug/L		NC	20
Indeno[1,2,3-cd]pyrene	<0.049		<0.049		ug/L		NC	20
Isophorone	<0.098		<0.098		ug/L		NC	20
Lindane	<0.0098		<0.0098		ug/L		NC	20
Malathion	<0.098		<0.098		ug/L		NC	20
Methoxychlor	<0.049		<0.049		ug/L		NC	20
Metolachlor	<0.049		<0.049		ug/L		NC	20
Molinate	<0.098		<0.098		ug/L		NC	20
Naphthalene	<0.098		<0.098		ug/L		NC	20
Parathion	<0.098		<0.098		ug/L		NC	20
Pendimethalin (Penoxaline)	<0.098		<0.098		ug/L		NC	20
Phenanthrene	<0.039		<0.039		ug/L		NC	20
Propachlor	<0.049		<0.049		ug/L		NC	20
Pyrene	<0.049		<0.049		ug/L		NC	20
Simazine	<0.049		<0.049		ug/L		NC	20
Terbacil	<0.098		<0.098		ug/L		NC	20
Terbutylazine	<0.098		<0.098		ug/L		NC	20
Thiobencarb	<0.098		<0.098		ug/L		NC	20
Total Permethrin (mixed isomers)	<0.20		<0.20		ug/L		NC	20
trans-Nonachlor	<0.049		<0.049		ug/L		NC	20
Trifluralin	<0.098		<0.098		ug/L		NC	20

Surrogate	DU DU		Limits
	%Recovery	Qualifier	
2-Nitro-m-xylene	96		70 - 130
Perylene-d12	91		70 - 130
Triphenylphosphate	107		70 - 130

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 625.1 - Semivolatile Organic Compounds (GC/MS)

Lab Sample ID: MB 570-678015/1-A
Matrix: Water
Analysis Batch: 681700

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 678015

<i>Tentatively Identified Compound</i>	<i>Est. Result</i>	<i>Qualifier</i>	<i>Unit</i>	<i>D</i>	<i>RT</i>	<i>CAS No.</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>Tentatively Identified Compound</i>	<i>None</i>		<i>ug/L</i>			<i>N/A</i>	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	<i>1</i>

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>2,4,6-Tribromophenol (Surr)</i>	90		33 - 139	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1
<i>2-Fluorobiphenyl (Surr)</i>	102		33 - 126	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1
<i>2-Fluorophenol (Surr)</i>	71		12 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1
<i>Nitrobenzene-d5 (Surr)</i>	125	S1+	36 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1
<i>Phenol-d6 (Surr)</i>	46		10 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1
<i>p-Terphenyl-d14 (Surr)</i>	112		47 - 131	<i>01/02/26 12:55</i>	<i>01/13/26 15:02</i>	1

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM)

Lab Sample ID: MB 570-678015/1-A
Matrix: Water
Analysis Batch: 681620

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 678015

<i>Analyte</i>	<i>Result</i>	<i>Qualifier</i>	<i>RL</i>	<i>Unit</i>	<i>D</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
1-Methylnaphthalene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
2-Methylnaphthalene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Acenaphthene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Acenaphthylene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Anthracene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Benzo[a]anthracene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Benzo[a]pyrene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Benzo[b]fluoranthene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Benzo[g,h,i]perylene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Benzo[k]fluoranthene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Chrysene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Dibenz(a,h)anthracene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Fluoranthene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Fluorene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Indeno[1,2,3-cd]pyrene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Naphthalene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Phenanthrene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1
Pyrene	<0.20		0.20	ug/L		01/02/26 12:55	01/13/26 08:10	1

<i>Surrogate</i>	<i>%Recovery</i>	<i>Qualifier</i>	<i>Limits</i>	<i>Prepared</i>	<i>Analyzed</i>	<i>Dil Fac</i>
<i>2,4,6-Tribromophenol (Surr)</i>	102		28 - 127	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1
<i>2-Fluorobiphenyl (Surr)</i>	95		31 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1
<i>2-Fluorophenol (Surr)</i>	72		17 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1
<i>Nitrobenzene-d5 (Surr)</i>	112		27 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1
<i>Phenol-d6 (Surr)</i>	45		10 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1
<i>p-Terphenyl-d14 (Surr)</i>	113		45 - 120	<i>01/02/26 12:55</i>	<i>01/13/26 08:10</i>	1

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCS 570-678015/2-A

Matrix: Water

Analysis Batch: 681620

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 678015

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec	
							Limits	
1-Methylnaphthalene	20.0	16.6		ug/L		83	47 - 120	
2-Methylnaphthalene	20.0	16.2		ug/L		81	43 - 120	
Acenaphthene	20.0	19.2		ug/L		96	60 - 132	
Acenaphthylene	20.0	19.4		ug/L		97	54 - 126	
Anthracene	20.0	20.2		ug/L		101	43 - 120	
Benzo[a]anthracene	20.0	22.0		ug/L		110	42 - 133	
Benzo[a]pyrene	20.0	20.7		ug/L		104	32 - 148	
Benzo[b]fluoranthene	20.0	21.1		ug/L		106	42 - 140	
Benzo[g,h,i]perylene	20.0	20.6		ug/L		103	1 - 195	
Benzo[k]fluoranthene	20.0	20.1		ug/L		101	25 - 146	
Chrysene	20.0	20.3		ug/L		101	44 - 140	
Dibenz(a,h)anthracene	20.0	22.3		ug/L		112	1 - 200	
Fluoranthene	20.0	19.8		ug/L		99	43 - 121	
Fluorene	20.0	19.2		ug/L		96	70 - 120	
Indeno[1,2,3-cd]pyrene	20.0	23.1		ug/L		116	1 - 151	
Naphthalene	20.0	15.3		ug/L		77	36 - 120	
Phenanthrene	20.0	19.3		ug/L		96	65 - 120	
Pyrene	20.0	21.8		ug/L		109	70 - 120	

Surrogate	LCS LCS		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	99		28 - 127
2-Fluorobiphenyl (Surr)	94		31 - 120
2-Fluorophenol (Surr)	75		17 - 120
Nitrobenzene-d5 (Surr)	85		27 - 120
Phenol-d6 (Surr)	49		10 - 120
p-Terphenyl-d14 (Surr)	107		45 - 120

Lab Sample ID: LCSD 570-678015/3-A

Matrix: Water

Analysis Batch: 681620

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 678015

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec		RPD	
							Limits		RPD	Limit
1-Methylnaphthalene	20.0	3.45	*- *1	ug/L		17	47 - 120	131	20	
2-Methylnaphthalene	20.0	3.38	*- *1	ug/L		17	43 - 120	131	20	
Acenaphthene	20.0	3.51	*- *1	ug/L		18	60 - 132	138	29	
Acenaphthylene	20.0	3.70	*- *1	ug/L		18	54 - 126	136	45	
Anthracene	20.0	3.93	*- *1	ug/L		20	43 - 120	135	40	
Benzo[a]anthracene	20.0	3.88	*- *1	ug/L		19	42 - 133	140	32	
Benzo[a]pyrene	20.0	3.50	*- *1	ug/L		18	32 - 148	142	43	
Benzo[b]fluoranthene	20.0	3.71	*- *1	ug/L		19	42 - 140	140	43	
Benzo[g,h,i]perylene	20.0	3.69	*1	ug/L		18	1 - 195	139	61	
Benzo[k]fluoranthene	20.0	3.47	*- *1	ug/L		17	25 - 146	141	38	
Chrysene	20.0	3.70	*- *1	ug/L		18	44 - 140	138	53	
Dibenz(a,h)anthracene	20.0	4.17	*1	ug/L		21	1 - 200	137	75	
Fluoranthene	20.0	3.63	*- *1	ug/L		18	43 - 121	138	40	
Fluorene	20.0	3.63	*- *1	ug/L		18	70 - 120	136	23	
Indeno[1,2,3-cd]pyrene	20.0	4.23	*1	ug/L		21	1 - 151	138	60	
Naphthalene	20.0	3.20	*- *1	ug/L		16	36 - 120	131	39	

Eurofins Pomona

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 625.1 SIM - Semivolatle Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: LCSD 570-678015/3-A

Client Sample ID: Lab Control Sample Dup

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 681620

Prep Batch: 678015

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Phenanthrene	20.0	3.67	*- *1	ug/L		18	65 - 120	136	24
Pyrene	20.0	4.19	*- *1	ug/L		21	70 - 120	136	30

Surrogate	LCSD LCSD		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	16	S1-	28 - 127
2-Fluorobiphenyl (Surr)	16	S1-	31 - 120
2-Fluorophenol (Surr)	11	S1-	17 - 120
Nitrobenzene-d5 (Surr)	16	S1-	27 - 120
Phenol-d6 (Surr)	7	S1-	10 - 120
p-Terphenyl-d14 (Surr)	19	S1-	45 - 120

Lab Sample ID: 380-190087-A-1-A MS

Client Sample ID: Matrix Spike

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 681620

Prep Batch: 678015

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec Limits
1-Methylnaphthalene	<0.19	*- *1	19.2	15.3		ug/L		80	36 - 120
2-Methylnaphthalene	<0.19	*- *1	19.2	15.4		ug/L		80	32 - 124
Acenaphthene	<0.19	*- *1	19.2	18.1		ug/L		94	47 - 145
Acenaphthylene	<0.19	*- *1	19.2	18.6		ug/L		97	33 - 145
Anthracene	<0.19	*- *1	19.2	18.8		ug/L		98	27 - 133
Benzo[a]anthracene	<0.19	*- *1	19.2	20.3		ug/L		106	33 - 143
Benzo[a]pyrene	<0.19	*- *1	19.2	19.2		ug/L		100	17 - 163
Benzo[b]fluoranthene	<0.19	*- *1	19.2	20.0		ug/L		104	24 - 159
Benzo[g,h,i]perylene	<0.19	*1	19.2	18.7		ug/L		97	1 - 219
Benzo[k]fluoranthene	<0.19	*- *1	19.2	18.7		ug/L		97	11 - 162
Chrysene	<0.19	*- *1	19.2	19.0		ug/L		99	17 - 168
Dibenz(a,h)anthracene	<0.19	*1	19.2	20.6		ug/L		107	1 - 227
Fluoranthene	<0.19	*- *1	19.2	18.4		ug/L		96	26 - 137
Fluorene	<0.19	*- *1	19.2	18.1		ug/L		94	59 - 121
Indeno[1,2,3-cd]pyrene	<0.19	*1	19.2	21.1		ug/L		110	1 - 171
Naphthalene	<0.19	*- *1	19.2	14.3		ug/L		75	21 - 133
Phenanthrene	<0.19	*- *1	19.2	18.0		ug/L		94	54 - 120
Pyrene	<0.19	*- *1	19.2	20.3		ug/L		106	52 - 120

Surrogate	MS MS		Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	92		28 - 127
2-Fluorobiphenyl (Surr)	92		31 - 120
2-Fluorophenol (Surr)	70		17 - 120
Nitrobenzene-d5 (Surr)	80		27 - 120
Phenol-d6 (Surr)	44		10 - 120
p-Terphenyl-d14 (Surr)	101		45 - 120

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 625.1 SIM - Semivolatile Organic Compounds GC/MS (SIM) (Continued)

Lab Sample ID: 380-190087-A-1-B MSD

Matrix: Water

Analysis Batch: 681620

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 678015

Analyte	Sample	Sample	Spike	MSD	MSD	Unit	D	%Rec	%Rec	RPD	RPD
	Result	Qualifier	Added	Result	Qualifier				Limits		Limit
1-Methylnaphthalene	<0.19	*- *1	19.2	15.2		ug/L		79	36 - 120	0	30
2-Methylnaphthalene	<0.19	*- *1	19.2	15.4		ug/L		80	32 - 124	0	30
Acenaphthene	<0.19	*- *1	19.2	17.7		ug/L		92	47 - 145	3	48
Acenaphthylene	<0.19	*- *1	19.2	18.2		ug/L		95	33 - 145	2	74
Anthracene	<0.19	*- *1	19.2	18.7		ug/L		97	27 - 133	1	66
Benzo[a]anthracene	<0.19	*- *1	19.2	20.1		ug/L		105	33 - 143	1	53
Benzo[a]pyrene	<0.19	*- *1	19.2	18.9		ug/L		98	17 - 163	1	72
Benzo[b]fluoranthene	<0.19	*- *1	19.2	19.5		ug/L		102	24 - 159	3	71
Benzo[g,h,i]perylene	<0.19	*1	19.2	18.1		ug/L		94	1 - 219	3	97
Benzo[k]fluoranthene	<0.19	*- *1	19.2	17.8		ug/L		93	11 - 162	4	63
Chrysene	<0.19	*- *1	19.2	18.4		ug/L		96	17 - 168	3	87
Dibenz(a,h)anthracene	<0.19	*1	19.2	19.8		ug/L		103	1 - 227	4	126
Fluoranthene	<0.19	*- *1	19.2	18.4		ug/L		96	26 - 137	0	66
Fluorene	<0.19	*- *1	19.2	17.8		ug/L		93	59 - 121	2	38
Indeno[1,2,3-cd]pyrene	<0.19	*1	19.2	20.4		ug/L		106	1 - 171	3	99
Naphthalene	<0.19	*- *1	19.2	14.0		ug/L		73	21 - 133	3	65
Phenanthrene	<0.19	*- *1	19.2	17.9		ug/L		93	54 - 120	1	39
Pyrene	<0.19	*- *1	19.2	20.1		ug/L		105	52 - 120	1	49

Surrogate	MSD	MSD	Limits
	%Recovery	Qualifier	
2,4,6-Tribromophenol (Surr)	89		28 - 127
2-Fluorobiphenyl (Surr)	85		31 - 120
2-Fluorophenol (Surr)	66		17 - 120
Nitrobenzene-d5 (Surr)	78		27 - 120
Phenol-d6 (Surr)	42		10 - 120
p-Terphenyl-d14 (Surr)	96		45 - 120

Method: 8015B GRO LL - Gasoline Range Organics - (GC)

Lab Sample ID: MB 570-678421/13

Matrix: Water

Analysis Batch: 678421

Client Sample ID: Method Blank

Prep Type: Total/NA

Analyte	MB	MB	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
GRO (C6-C10)	<10		10	ug/L			01/04/26 18:26	1

Surrogate	MB	MB	Limits	Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier				
4-Bromofluorobenzene (Surr)	98		38 - 134		01/04/26 18:26	1

Lab Sample ID: LCS 570-678421/1011

Matrix: Water

Analysis Batch: 678421

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Analyte	Spike Added	LCS	LCS	Unit	D	%Rec	%Rec
		Result	Qualifier				Limits
Gasoline Range Organics (C4-C13)	400	444		ug/L		111	78 - 120

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 8015B GRO LL - Gasoline Range Organics - (GC) (Continued)

Lab Sample ID: LCS 570-678421/1011
Matrix: Water
Analysis Batch: 678421

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

	LCS	LCS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	91		38 - 134

Lab Sample ID: LCSD 570-678421/12
Matrix: Water
Analysis Batch: 678421

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA

Analyte		Spike Added	LCSD	LCSD	Unit	D	%Rec	%Rec	RPD	Limit
			Result	Qualifier				Limits		
Gasoline Range Organics (C4-C13)		400	426		ug/L		107	78 - 120	4	10

	LCSD	LCSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	94		38 - 134

Lab Sample ID: MRL 570-678421/14
Matrix: Water
Analysis Batch: 678421

Client Sample ID: Lab Control Sample
Prep Type: Total/NA

Analyte		Spike Added	MRL	MRL	Unit	D	%Rec	%Rec	
			Result	Qualifier				Limits	
Gasoline Range Organics (C4-C13)		10.0	9.44	J	ug/L		94	50 - 150	

	MRL	MRL	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	94		38 - 134

Lab Sample ID: 380-190087-C-1 MS
Matrix: Water
Analysis Batch: 678421

Client Sample ID: Matrix Spike
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MS	MS	Unit	D	%Rec	%Rec	
				Result	Qualifier				Limits	
Gasoline Range Organics (C4-C13)	<10	F1	400	490		ug/L		122	68 - 122	

	MS	MS	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	102		38 - 134

Lab Sample ID: 380-190087-C-1 MSD
Matrix: Water
Analysis Batch: 678421

Client Sample ID: Matrix Spike Duplicate
Prep Type: Total/NA

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD	MSD	Unit	D	%Rec	%Rec	RPD	Limit
				Result	Qualifier				Limits		
Gasoline Range Organics (C4-C13)	<10	F1	400	498	F1	ug/L		125	68 - 122	2	18

	MSD	MSD	
Surrogate	%Recovery	Qualifier	Limits
4-Bromofluorobenzene (Surr)	107		38 - 134

QC Sample Results

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level

Lab Sample ID: MB 570-678355/1-A
Matrix: Water
Analysis Batch: 679714

Client Sample ID: Method Blank
Prep Type: Total/NA
Prep Batch: 678355

Analyte	MB MB		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Diesel Range Organics (C10-C24)	<25		25	ug/L		01/04/26 09:32	01/07/26 21:05	1
Motor Oil Range Organics [C24-C36]	<25		25	ug/L		01/04/26 09:32	01/07/26 21:05	1
C8-C18	<25		25	ug/L		01/04/26 09:32	01/07/26 21:05	1
Surrogate	MB MB		Limits			Prepared	Analyzed	Dil Fac
	%Recovery	Qualifier						
n-Octacosane (Surr)	104		60 - 130			01/04/26 09:32	01/07/26 21:05	1

Lab Sample ID: LCS 570-678355/2-A
Matrix: Water
Analysis Batch: 679714

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 678355

Analyte	Spike Added	LCS LCS		Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
C10-C28	1600	1500		ug/L		94	56 - 127
Surrogate	LCS LCS		Limits				
	%Recovery	Qualifier					
n-Octacosane (Surr)	97		60 - 130				

Lab Sample ID: LCSD 570-678355/3-A
Matrix: Water
Analysis Batch: 679714

Client Sample ID: Lab Control Sample Dup
Prep Type: Total/NA
Prep Batch: 678355

Analyte	Spike Added	LCSD LCSD		Unit	D	%Rec	%Rec Limits	RPD	
		Result	Qualifier					RPD	Limit
C10-C28	1600	1520		ug/L		95	56 - 127	1	23
Surrogate	LCSD LCSD		Limits						
	%Recovery	Qualifier							
n-Octacosane (Surr)	97		60 - 130						

Lab Sample ID: MRL 570-678355/4-A
Matrix: Water
Analysis Batch: 679714

Client Sample ID: Lab Control Sample
Prep Type: Total/NA
Prep Batch: 678355

Analyte	Spike Added	MRL MRL		Unit	D	%Rec	%Rec Limits
		Result	Qualifier				
C10-C28	0.0200	0.0214	J	mg/L		107	50 - 150
Surrogate	MRL MRL		Limits				
	%Recovery	Qualifier					
n-Octacosane (Surr)	107		60 - 130				

Lab Sample ID: 380-190087-B-1-B MS
Matrix: Water
Analysis Batch: 679714

Client Sample ID: Matrix Spike
Prep Type: Total/NA
Prep Batch: 678355

Analyte	Sample Result	Sample Qualifier	Spike Added	MS MS		Unit	D	%Rec	%Rec Limits
				Result	Qualifier				
C10-C28	<26		1660	1690		ug/L		102	70 - 130
Surrogate	MS MS		Limits						
	%Recovery	Qualifier							
n-Octacosane (Surr)	107		60 - 130						

QC Sample Results

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-190084-1
 SDG: Weekly: Halawa Shaft Viewing Pool

Method: 8015B - Diesel Range Organics (DRO) (GC) Low Level (Continued)

Lab Sample ID: 380-190087-B-1-C MSD

Matrix: Water

Analysis Batch: 679714

Client Sample ID: Matrix Spike Duplicate

Prep Type: Total/NA

Prep Batch: 678355

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
C10-C28	<26		1640	1800		ug/L		110	70 - 130	6	20
<i>MSD MSD</i>											
Surrogate	%Recovery	MSD Qualifier	Limits								
<i>n-Octacosane (Surr)</i>	114		60 - 130								

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-190084-1
 SDG: Weekly: Halawa Shaft Viewing Pool

GC/MS Semi VOA

Prep Batch: 195565

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	525.2	
MB 380-195565/21-A	Method Blank	Total/NA	Water	525.2	
LCS 380-195565/23-A	Lab Control Sample	Total/NA	Water	525.2	
MRL 380-195565/22-A	Lab Control Sample	Total/NA	Water	525.2	
380-190097-I-1-A MS	Matrix Spike	Total/NA	Water	525.2	
380-190160-B-1-A DU	Duplicate	Total/NA	Water	525.2	

Analysis Batch: 195644

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	525.2	195565
MB 380-195565/21-A	Method Blank	Total/NA	Water	525.2	195565
LCS 380-195565/23-A	Lab Control Sample	Total/NA	Water	525.2	195565
MRL 380-195565/22-A	Lab Control Sample	Total/NA	Water	525.2	195565
380-190097-I-1-A MS	Matrix Spike	Total/NA	Water	525.2	195565
380-190160-B-1-A DU	Duplicate	Total/NA	Water	525.2	195565

Prep Batch: 678015

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	625.1	
MB 570-678015/1-A	Method Blank	Total/NA	Water	625.1	
LCS 570-678015/2-A	Lab Control Sample	Total/NA	Water	625.1	
LCSD 570-678015/3-A	Lab Control Sample Dup	Total/NA	Water	625.1	
380-190087-A-1-A MS	Matrix Spike	Total/NA	Water	625.1	
380-190087-A-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1	

Analysis Batch: 681620

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	625.1 SIM	678015
MB 570-678015/1-A	Method Blank	Total/NA	Water	625.1 SIM	678015
LCS 570-678015/2-A	Lab Control Sample	Total/NA	Water	625.1 SIM	678015
LCSD 570-678015/3-A	Lab Control Sample Dup	Total/NA	Water	625.1 SIM	678015
380-190087-A-1-A MS	Matrix Spike	Total/NA	Water	625.1 SIM	678015
380-190087-A-1-B MSD	Matrix Spike Duplicate	Total/NA	Water	625.1 SIM	678015

Analysis Batch: 681700

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	625.1	678015
MB 570-678015/1-A	Method Blank	Total/NA	Water	625.1	678015

GC VOA

Analysis Batch: 678421

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	8015B GRO LL	
380-190084-2	TB: Halawa Shaft Viewing Pool	Total/NA	Water	8015B GRO LL	
MB 570-678421/13	Method Blank	Total/NA	Water	8015B GRO LL	
LCS 570-678421/1011	Lab Control Sample	Total/NA	Water	8015B GRO LL	
LCSD 570-678421/12	Lab Control Sample Dup	Total/NA	Water	8015B GRO LL	
MRL 570-678421/14	Lab Control Sample	Total/NA	Water	8015B GRO LL	
380-190087-C-1 MS	Matrix Spike	Total/NA	Water	8015B GRO LL	
380-190087-C-1 MSD	Matrix Spike Duplicate	Total/NA	Water	8015B GRO LL	

QC Association Summary

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-190084-1
 SDG: Weekly: Halawa Shaft Viewing Pool

GC Semi VOA

Prep Batch: 678355

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	3510C	
MB 570-678355/1-A	Method Blank	Total/NA	Water	3510C	
LCS 570-678355/2-A	Lab Control Sample	Total/NA	Water	3510C	
LCSD 570-678355/3-A	Lab Control Sample Dup	Total/NA	Water	3510C	
MRL 570-678355/4-A	Lab Control Sample	Total/NA	Water	3510C	
380-190087-B-1-B MS	Matrix Spike	Total/NA	Water	3510C	
380-190087-B-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	3510C	

Analysis Batch: 679714

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
380-190084-1	Halawa Shaft Viewing Pool	Total/NA	Water	8015B	678355
MB 570-678355/1-A	Method Blank	Total/NA	Water	8015B	678355
LCS 570-678355/2-A	Lab Control Sample	Total/NA	Water	8015B	678355
LCSD 570-678355/3-A	Lab Control Sample Dup	Total/NA	Water	8015B	678355
MRL 570-678355/4-A	Lab Control Sample	Total/NA	Water	8015B	678355
380-190087-B-1-B MS	Matrix Spike	Total/NA	Water	8015B	678355
380-190087-B-1-C MSD	Matrix Spike Duplicate	Total/NA	Water	8015B	678355



Lab Chronicle

Client: City & County of Honolulu
 Project/Site: RED-HILL

Job ID: 380-190084-1
 SDG: Weekly: Halawa Shaft Viewing Pool

Client Sample ID: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-1

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	525.2			195565	IQ42	EA POM	01/02/26 14:48
Total/NA	Analysis	525.2		1	195644	UPAC	EA POM	01/04/26 16:32
Total/NA	Prep	625.1			678015	VAW2	EET CAL 4	01/02/26 12:55
Total/NA	Analysis	625.1		1	681700	PQS1	EET CAL 4	01/13/26 19:29
Total/NA	Prep	625.1			678015	VAW2	EET CAL 4	01/02/26 12:55
Total/NA	Analysis	625.1 SIM		1	681620	PQS1	EET CAL 4	01/13/26 15:39
Total/NA	Analysis	8015B GRO LL		1	678421	YD9V	EET CAL 4	01/04/26 21:56
Total/NA	Prep	3510C			678355	TVD6	EET CAL 4	01/04/26 09:33
Total/NA	Analysis	8015B		1	679714	NR	EET CAL 4	01/07/26 23:58

Client Sample ID: TB: Halawa Shaft Viewing Pool

Lab Sample ID: 380-190084-2

Date Collected: 12/30/25 10:00

Matrix: Water

Date Received: 12/31/25 09:57

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Analysis	8015B GRO LL		1	678421	YD9V	EET CAL 4	01/05/26 01:37

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Laboratory: Eurofins Pomona

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
Hawaii	State	CA00006	01-31-26

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
525.2	525.2	Water	1-Methylnaphthalene
525.2	525.2	Water	2,4'-DDD
525.2	525.2	Water	2,4'-DDE
525.2	525.2	Water	2,4'-DDT
525.2	525.2	Water	2,4-Dinitrotoluene
525.2	525.2	Water	2,6-Dinitrotoluene
525.2	525.2	Water	2-Methylnaphthalene
525.2	525.2	Water	4,4'-DDD
525.2	525.2	Water	4,4'-DDE
525.2	525.2	Water	4,4' DDT
525.2	525.2	Water	Acetochlor
525.2	525.2	Water	alpha-BHC
525.2	525.2	Water	alpha-Chlordane
525.2	525.2	Water	beta-BHC
525.2	525.2	Water	Chlorobenzilate
525.2	525.2	Water	Chloroneb
525.2	525.2	Water	Chlorothalonil (Draconil, Bravo)
525.2	525.2	Water	Chlorpyrifos
525.2	525.2	Water	delta-BHC
525.2	525.2	Water	Diclorvos (DDVP)
525.2	525.2	Water	Endosulfan I (Alpha)
525.2	525.2	Water	Endosulfan II (Beta)
525.2	525.2	Water	Endosulfan sulfate
525.2	525.2	Water	Endrin aldehyde
525.2	525.2	Water	EPTC
525.2	525.2	Water	gamma-Chlordane
525.2	525.2	Water	Isophorone
525.2	525.2	Water	Malathion
525.2	525.2	Water	Parathion
525.2	525.2	Water	Pendimethalin (Penoxaline)
525.2	525.2	Water	Terbacil
525.2	525.2	Water	Terbutylazine
525.2	525.2	Water	Total Permethrin (mixed isomers)
525.2	525.2	Water	trans-Nonachlor

Laboratory: Eurofins Calscience

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
A2LA	Dept. of Defense ELAP	7296.01	11-30-26
A2LA	ISO/IEC 17025	7296.01	11-30-26
Alaska (UST)	State	25-005	03-02-26
Arizona	State	AZ0830	11-17-26
California	Los Angeles County Sanitation Districts	9257304	07-31-26
California	SCAQMD LAP	17LA0919	11-30-26

Accreditation/Certification Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Laboratory: Eurofins Calscience (Continued)

All accreditations/certifications held by this laboratory are listed. Not all accreditations/certifications are applicable to this report.

Authority	Program	Identification Number	Expiration Date
California	State	3082	07-31-26
Kansas	NELAP	E-10420	07-31-26
Nevada	State	CA00111	07-31-26
Oregon	NELAP	4175	02-02-26
USDA	US Federal Programs	525-23-159-97150	06-08-26
Utah	NELAP	CA00111	02-28-26
Washington	State	C916	10-11-26

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Method Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Method	Method Description	Protocol	Laboratory
525.2	Semivolatile Organic Compounds (GC/MS)	EPA	EA POM
625.1	Semivolatile Organic Compounds (GC/MS)	EPA	EET CAL 4
625.1 SIM	Semivolatile Organic Compounds GC/MS (SIM)	EPA	EET CAL 4
8015B GRO LL	Gasoline Range Organics - (GC)	SW846	EET CAL 4
8015B	Diesel Range Organics (DRO) (GC) Low Level	SW846	EET CAL 4
3510C	Liquid-Liquid Extraction (Separatory Funnel)	SW846	EET CAL 4
5030C	Purge and Trap	SW846	EET CAL 4
525.2	Extraction of Semivolatile Compounds	EPA	EA POM
625.1	Liquid-Liquid Extraction	40CFR136A	EET CAL 4

Protocol References:

40CFR136A = "Methods for Organic Chemical Analysis of Municipal Industrial Wastewater", 40CFR, Part 136, Appendix A, October 26, 1984 and subsequent revisions.

EPA = US Environmental Protection Agency

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

Laboratory References:

EA POM = Eurofins Pomona, 941 Corporate Center Drive, Pomona, CA 91768-2642, TEL (626)386-1100

EET CAL 4 = Eurofins Calscience Tustin, 2841 Dow Avenue, Tustin, CA 92780, TEL (714)895-5494

Sample Summary

Client: City & County of Honolulu
Project/Site: RED-HILL

Job ID: 380-190084-1
SDG: Weekly: Halawa Shaft Viewing Pool

Lab Sample ID	Client Sample ID	Matrix	Collected	Received	Sample Origin
380-190084-1	Halawa Shaft Viewing Pool	Water	12/30/25 10:00	12/31/25 09:57	Hawaii
380-190084-2	TB: Halawa Shaft Viewing Pool	Water	12/30/25 10:00	12/31/25 09:57	Hawaii

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16

Eurofins Eaton Analytical Pomona

941 Corporate Center Drive
 Pomona, CA 91768-2642
 Phone: 626-386-1100

Chain of Custody Record



eurofins | Environ

Loc: 380
190084

Client Information (Sub Contract Lab)		Sampler: N/A	Lab PM: Lopez, Maria	Carrier Tracking No(s): N/A	COC No: 380-291397.1							
Client Contact: Shipping/Receiving		Phone: N/A	E-Mail: Maria.Lopez@et.eurofinsus.com	State of Origin: Hawaii	Page: Page 1 of 1							
Company: Eurofins Environment Testing Southwest L		Accreditations Required (See note): State - Hawaii			Job #: 380-190084-1							
Address: 2841 Dow Avenue, Suite 100, City: Tustin State, Zip: CA, 92780 Phone: 714-895-5494(Tel) Email: N/A Project Name: RED-HILL Site: Honolulu BWS Sites		Due Date Requested: 1/14/2026 TAT Requested (days): N/A	Analysis Requested			Preservation Codes: -						
		PO #: N/A	WO #: N/A	 380-190084 Chain of Custody								
		Project #: 38001111	SSOW#: N/A									
Sample Identification - Client ID (Lab ID)		Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=wastewater, BT=Tissue, A=Air)	Field Filtered Sample (Yes or No)	Perform MS/MSD (Yes or No)	8018B_DRO_LL_CS/AS10C_LL.HPL Ranges: C10-C24/C24-C36/C8-C18	8018B_GRO_LL/5030C(IMOD) GRO	625.1_SIMAC25_Prep(IMOD) Extended PAH List	Total Number of	Special Instructions/Note:
		Preservation Code:										
Halawa Shaft Viewing Pool (380-190084-1)		12/30/25	10:00 Hawaiian	G	Water		X	X	X		7	MRLs are needed. Confirm any hits >RL.
TB: Halawa Shaft Viewing Pool (380-190084-2)		12/30/25	10:00 Hawaiian	G	Water				X		8	MRLs are needed.
<p>Note: Since laboratory accreditations are subject to change, Eurofins Eaton Analytical, LLC places the ownership of method, analyte & accreditation compliance upon our subcontract laboratories. This sample shipment is forwarded under chain-of-custody. If the laboratory does not currently maintain accreditation in the State of Origin listed above for analysis/tests/matrix being analyzed, the samples must be shipped back to the Eurofins Eaton Analytical, LLC laboratory or other instructions will be provided. Any changes to accreditation status should be brought to Eurofins Eaton Analytical, LLC attention immediately. If all requested accreditations are current to date, return the signed Chain of Custody attesting to said compliance to Eurofins Eaton Analytical, LLC.</p>												
Possible Hazard Identification						Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)						
Unconfirmed						<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months						
Deliverable Requested: I, II, III, IV, Other (specify)			Primary Deliverable Rank: 2			Special Instructions/QC Requirements:						
Empty Kit Relinquished by:		Date:	Time:	Method of Shipment:								
Relinquished by: <i>Mark Urcia</i>		Date/Time: 12/31/25 19:00	Company: <i>FEAP</i>	Received by: <i>LT</i>		Date/Time: 12-31-25 16:00	Company: <i>WY</i>					
Relinquished by: <i>LT</i>		Date/Time: 12-31-25 15:52	Company: <i>WY</i>	Received by: <i>FC</i>		Date/Time: 12-31-25 15:52	Company: <i>FC</i>					
Relinquished by:		Date/Time:	Company:	Received by:		Date/Time:	Company:					
Custody Seals Intact: Δ Yes Δ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks: 2-1 / 2-2 IR4								

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16

Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-190084-1

SDG Number: Weekly: Halawa Shaft Viewing Pool

Login Number: 190084

List Number: 1

Creator: Tran, Kristine

List Source: Eurofins Pomona

Question	Answer	Comment
The coolers custody seal, if present, is intact.	N/A	
Sample custody seals, if present, are intact.	N/A	
Samples were received on ice.	True	
Cooler(s) Temperature is acceptable.	True	
Cooler(s) Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and is legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	N/A	
CIO4 headspace requirement met (>50% for CA, >30% for other states).	N/A	
Samples do not require splitting or compositing.	True	
Container provided by EEA	True	



Login Sample Receipt Checklist

Client: City & County of Honolulu

Job Number: 380-190084-1
SDG Number: Weekly: Halawa Shaft Viewing Pool

Login Number: 190084

List Number: 2

Creator: Szymborski, Jessica

List Source: Eurofins Calscience

List Creation: 01/02/26 10:42 AM

Question	Answer	Comment
Radioactivity wasn't checked or is <=/ background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	2.1/2.2
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	N/A	Received project as a subcontract.
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	